

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX020922\
 Data File : VX026897.D
 Acq On : 10 Feb 2022 03:20
 Operator : JC/MD
 Sample : N1446-09
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 39 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-136C

Quant Time: Feb 10 03:54:37 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020822W.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 08 07:07:46 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	127765	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	216743	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	192096	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	80360	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	83457	54.416	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 108.840%	
35) Dibromofluoromethane	5.391	113	67577	52.332	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 104.660%	
50) Toluene-d8	8.653	98	252487	53.267	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 106.540%	
62) 4-Bromofluorobenzene	11.079	95	89856	53.844	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 107.680%	
Target Compounds						
					Qvalue	
11) Tert butyl alcohol	2.971	59	114974	414.969	ug/l	96
16) Acetone	2.386	43	1746	2.592	ug/l	95
19) Methyl tert-butyl Ether	3.117	73	1607664	347.359	ug/l	99
22) Diisopropyl ether	3.763	45	4157	0.891	ug/l #	78

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX020922\
Data File : VX026897.D
Acq On : 10 Feb 2022 03:20
Operator : JC/MD
Sample : N1446-09
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 39 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-136C

Quant Time: Feb 10 03:54:37 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020822W.M
Quant Title : SW846 8260
QLast Update : Tue Feb 08 07:07:46 2022
Response via : Initial Calibration

