

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX020922\
 Data File : VX026901.D
 Acq On : 10 Feb 2022 04:52
 Operator : JC/MD
 Sample : N1447-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 43 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-138B

Quant Time: Feb 10 07:30:03 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020822W.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 08 07:07:46 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	126308	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	216110	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	193626	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	80060	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	83369	54.985	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 109.980%	
35) Dibromofluoromethane	5.391	113	67476	52.407	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 104.820%	
50) Toluene-d8	8.653	98	251798	53.277	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 106.560%	
62) 4-Bromofluorobenzene	11.079	95	92280	55.459	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 110.920%	
Target Compounds						
19) Methyl tert-butyl Ether	3.117	73	523376	114.387	ug/l	Qvalue 100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX020922\
Data File : VX026901.D
Acq On : 10 Feb 2022 04:52
Operator : JC/MD
Sample : N1447-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 43 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-138B

Quant Time: Feb 10 07:30:03 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020822W.M
Quant Title : SW846 8260
QLast Update : Tue Feb 08 07:07:46 2022
Response via : Initial Calibration

