

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX020922\
 Data File : VX026904.D
 Acq On : 10 Feb 2022 06:01
 Operator : JC/MD
 Sample : N1447-06
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 46 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-140A

Quant Time: Feb 10 07:31:02 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020822W.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 08 07:07:46 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	125296	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	213347	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	193852	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	85265	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	84104	55.918	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 111.840%	
35) Dibromofluoromethane	5.391	113	67366	52.999	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 106.000%	
50) Toluene-d8	8.653	98	252277	54.070	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 108.140%	
62) 4-Bromofluorobenzene	11.079	95	95182	57.944	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 115.880%	
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	1456	2.204	ug/l #	83
19) Methyl tert-butyl Ether	3.117	73	6533	1.439	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX020922\
Data File : VX026904.D
Acq On : 10 Feb 2022 06:01
Operator : JC/MD
Sample : N1447-06
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 46 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-140A

Quant Time: Feb 10 07:31:02 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020822W.M
Quant Title : SW846 8260
QLast Update : Tue Feb 08 07:07:46 2022
Response via : Initial Calibration

