

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX020922\
 Data File : VX026905.D
 Acq On : 10 Feb 2022 06:24
 Operator : JC/MD
 Sample : N1447-07
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 47 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-140B

Quant Time: Feb 10 07:31:25 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020822W.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 08 07:07:46 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	130364	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	220699	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	197334	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	82058	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	85259	54.482	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 108.960%	
35) Dibromofluoromethane	5.391	113	69733	53.034	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 106.060%	
50) Toluene-d8	8.653	98	256431	53.129	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 106.260%	
62) 4-Bromofluorobenzene	11.079	95	93043	54.755	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 109.500%	
Target Compounds						Qvalue
16) Acetone	2.386	43	1325	1.928	ug/l #	79
19) Methyl tert-butyl Ether	3.111	73	7502	1.589	ug/l #	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX020922\
Data File : VX026905.D
Acq On : 10 Feb 2022 06:24
Operator : JC/MD
Sample : N1447-07
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 47 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-140B

Quant Time: Feb 10 07:31:25 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020822W.M
Quant Title : SW846 8260
QLast Update : Tue Feb 08 07:07:46 2022
Response via : Initial Calibration

