

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX020922\
 Data File : VX026884.D
 Acq On : 09 Feb 2022 21:11
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Feb 10 01:11:28 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020822W.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 08 07:07:46 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	99	0.00
2 T	Dichlorodifluoromethane	0.544	0.533	2.0	98	0.00
3 P	Chloromethane	0.506	0.507	-0.2	103	0.00
4 C	Vinyl Chloride	0.529	0.525	0.8#	100	0.00
5 T	Bromomethane	0.225	0.209	7.1	101	0.00
6 T	Chloroethane	0.281	0.303	-7.8	99	0.00
7 T	Trichlorofluoromethane	0.779	0.774	0.6	99	0.00
8 T	Diethyl Ether	0.289	0.275	4.8	98	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.511	0.493	3.5	97	0.00
10 T	Methyl Iodide	0.656	0.659	-0.5	97	0.00
11 T	Tert butyl alcohol	0.108	0.090	16.7	86	0.00
12 CM	1,1-Dichloroethene	0.512	0.490	4.3#	97	0.00
13 T	Acrolein	0.100	0.097	3.0	99	0.00
14 T	Allyl chloride	0.906	0.883	2.5	98	0.00
15 T	Acrylonitrile	0.311	0.301	3.2	98	0.00
16 T	Acetone	0.264	0.244	7.6	98	0.00
17 T	Carbon Disulfide	1.449	1.335	7.9	94	0.00
18 T	Methyl Acetate	0.758	0.697	8.0	97	0.00
19 T	Methyl tert-butyl Ether	1.811	1.737	4.1	98	0.00
20 T	Methylene Chloride	0.571	0.548	4.0	101	0.00
21 T	trans-1,2-Dichloroethene	0.538	0.522	3.0	97	0.00
22 T	Diisopropyl ether	1.825	1.771	3.0	99	0.00
23 T	Vinyl Acetate	1.534	1.526	0.5	98	0.00
24 P	1,1-Dichloroethane	0.987	0.970	1.7	100	0.00
25 T	2-Butanone	0.425	0.409	3.8	97	0.00
26 T	2,2-Dichloropropane	0.667	0.660	1.0	97	0.00
27 T	cis-1,2-Dichloroethene	0.618	0.615	0.5	99	0.00
28 T	Bromochloromethane	0.409	0.409	0.0	101	0.00
29 T	Tetrahydrofuran	0.289	0.277	4.2	98	0.00
30 C	Chloroform	1.026	0.996	2.9#	101	0.00
31 T	Cyclohexane	0.954	0.919	3.7	98	0.00
32 T	1,1,1-Trichloroethane	0.905	0.878	3.0	97	0.00
33 S	1,2-Dichloroethane-d4	0.600	0.613	-2.2	109	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	100	0.00
35 S	Dibromofluoromethane	0.298	0.292	2.0	107	0.00
36 T	1,1-Dichloropropene	0.470	0.456	3.0	101	0.00
37 T	Ethyl Acetate	0.534	0.499	6.6	97	0.00
38 T	Carbon Tetrachloride	0.493	0.467	5.3	98	0.00
39 T	Methylcyclohexane	0.576	0.561	2.6	99	0.00
40 TM	Benzene	1.370	1.319	3.7	100	0.00
41 T	Methacrylonitrile	0.289	0.282	2.4	99	0.00
42 TM	1,2-Dichloroethane	0.483	0.480	0.6	101	0.00
43 T	Isopropyl Acetate	0.822	0.809	1.6	98	0.00
44 TM	Trichloroethene	0.411	0.396	3.6	99	0.00
45 C	1,2-Dichloropropane	0.347	0.339	2.3#	101	0.00
46 T	Dibromomethane	0.240	0.234	2.5	101	0.00
47 T	Bromodichloromethane	0.459	0.464	-1.1	101	0.00
48 T	Methyl methacrylate	0.407	0.395	2.9	99	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX020922\
 Data File : VX026884.D
 Acq On : 09 Feb 2022 21:11
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Feb 10 01:11:28 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020822W.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 08 07:07:46 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.007	0.006	14.3	87	0.00
50 S	Toluene-d8	1.093	1.119	-2.4	111	0.00
51 T	4-Methyl-2-Pentanone	0.498	0.481	3.4	100	0.00
52 CM	Toluene	0.867	0.842	2.9#	102	0.00
53 T	t-1,3-Dichloropropene	0.446	0.479	-7.4	103	0.00
54 T	cis-1,3-Dichloropropene	0.507	0.530	-4.5	101	0.00
55 T	1,1,2-Trichloroethane	0.328	0.329	-0.3	102	0.00
56 T	Ethyl methacrylate	0.520	0.519	0.2	100	0.00
57 T	1,3-Dichloropropane	0.564	0.566	-0.4	102	0.00
58 T	2-Chloroethyl Vinyl ether	0.267	0.266	0.4	95	0.00
59 T	2-Hexanone	0.364	0.347	4.7	98	0.00
60 T	Dibromochloromethane	0.338	0.353	-4.4	103	0.00
61 T	1,2-Dibromoethane	0.347	0.349	-0.6	102	0.00
62 S	4-Bromofluorobenzene	0.385	0.387	-0.5	110	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	103	0.00
64 T	Tetrachloroethene	0.502	0.472	6.0	96	0.00
65 PM	Chlorobenzene	1.034	0.994	3.9	102	0.00
66 T	1,1,1,2-Tetrachloroethane	0.380	0.372	2.1	102	0.00
67 C	Ethyl Benzene	1.830	1.805	1.4#	103	0.00
68 T	m/p-Xylenes	0.704	0.682	3.1	102	0.00
69 T	o-Xylene	0.690	0.668	3.2	102	0.00
70 T	Styrene	1.119	1.094	2.2	102	0.00
71 P	Bromoform	0.263	0.256	2.7	99	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	100	0.00
73 T	Isopropylbenzene	3.938	3.981	-1.1	102	0.00
74 T	N-amyl acetate	1.574	1.558	1.0	99	0.00
75 P	1,1,2,2-Tetrachloroethane	1.069	1.012	5.3	102	0.00
76 T	1,2,3-Trichloropropane	1.095	1.051	4.0	99	0.00
77 T	Bromobenzene	0.936	0.912	2.6	101	0.00
78 T	n-propylbenzene	4.366	4.494	-2.9	103	0.00
79 T	2-Chlorotoluene	2.698	2.636	2.3	102	0.00
80 T	1,3,5-Trimethylbenzene	3.241	3.240	0.0	102	0.00
81 T	trans-1,4-Dichloro-2-butene	0.317	0.306	3.5	94	0.00
82 T	4-Chlorotoluene	3.026	2.992	1.1	101	0.00
83 T	tert-Butylbenzene	3.089	3.033	1.8	102	0.00
84 T	1,2,4-Trimethylbenzene	3.269	3.231	1.2	102	0.00
85 T	sec-Butylbenzene	3.916	3.937	-0.5	102	0.00
86 T	p-Isopropyltoluene	3.318	3.334	-0.5	102	0.00
87 T	1,3-Dichlorobenzene	1.715	1.665	2.9	101	0.00
88 T	1,4-Dichlorobenzene	1.730	1.666	3.7	101	0.00
89 T	n-Butylbenzene	2.714	2.751	-1.4	102	0.00
90 T	Hexachloroethane	0.479	0.486	-1.5	101	0.00
91 T	1,2-Dichlorobenzene	1.655	1.605	3.0	99	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.263	0.258	1.9	96	0.00
93 T	1,2,4-Trichlorobenzene	1.045	1.031	1.3	99	0.00
94 T	Hexachlorobutadiene	0.442	0.411	7.0	95	0.00
95 T	Naphthalene	3.659	3.812	-4.2	102	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX020922\
Data File : VX026884.D
Acq On : 09 Feb 2022 21:11
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Feb 10 01:11:28 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020822W.M
Quant Title : SW846 8260
QLast Update : Tue Feb 08 07:07:46 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	1.058	1.032	2.5	98	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6