

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX020922\
 Data File : VX026868.D
 Acq On : 09 Feb 2022 14:39
 Operator : JC/MD
 Sample : VX0209WBSD01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0209WBSD01

Manual Integrations
 APPROVED

Reviewed By : Semsettin Yesilyurt 02/10/2022
 Supervised By : Mahesh Dadoda 02/10/2022

Quant Time: Feb 10 01:05:30 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020822W.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 08 07:07:46 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	97311	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	157407	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	135238	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	63222	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	65007	55.651	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery = 111.300%			
35) Dibromofluoromethane	5.391	113	50842	54.214	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 108.420%			
50) Toluene-d8	8.647	98	187621	54.503	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery = 109.000%			
62) 4-Bromofluorobenzene	11.079	95	67994	56.103	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 112.200%			
Target Compounds					Qvalue	
2) Dichlorodifluoromethane	1.172	85	22064	20.852	ug/l	100
3) Chloromethane	1.294	50	19853	20.169	ug/l	99
4) Vinyl Chloride	1.380	62	20958	20.339	ug/l	97
5) Bromomethane	1.617	94	9549	21.839	ug/l	96
6) Chloroethane	1.691	64	12432	20.197	ug/l	95
7) Trichlorofluoromethane	1.892	101	31933	21.059	ug/l	99
8) Diethyl Ether	2.136	74	11772	20.957	ug/l	98
9) 1,1,2-Trichlorotrifluo...	2.331	101	21477	21.608	ug/l	99
10) Methyl Iodide	2.459	142	24409	19.161	ug/l	99
11) Tert butyl alcohol	2.977	59	21477	101.775	ug/l	97
12) 1,1-Dichloroethene	2.325	96	20400	20.484	ug/l	96
13) Acrolein	2.239	56	19422	99.592	ug/l	98
14) Allyl chloride	2.666	41	36622	20.760	ug/l	96
15) Acrylonitrile	3.068	53	61282	101.340	ug/l	100
16) Acetone	2.386	43	59107	115.208	ug/l	97
17) Carbon Disulfide	2.514	76	54684	19.397	ug/l	99
18) Methyl Acetate	2.703	43	28810	19.517	ug/l	99
19) Methyl tert-butyl Ether	3.117	73	71431	20.264	ug/l	100
20) Methylene Chloride	2.794	84	22096	19.876	ug/l	96
21) trans-1,2-Dichloroethene	3.093	96	21911	20.908	ug/l	94
22) Diisopropyl ether	3.757	45	74471	20.968	ug/l #	84
23) Vinyl Acetate	3.721	43	317192	106.224	ug/l	99
24) 1,1-Dichloroethane	3.611	63	39774	20.696	ug/l	98
25) 2-Butanone	4.556	43	87414	105.751	ug/l	99
26) 2,2-Dichloropropane	4.477	77	29911	23.035	ug/l	98
27) cis-1,2-Dichloroethene	4.489	96	25169	20.918	ug/l	97
28) Bromochloromethane	4.897	49	17175	21.559	ug/l	100
29) Tetrahydrofuran	5.007	42	55464	98.699	ug/l	98
30) Chloroform	5.092	83	40798	20.432	ug/l	98
31) Cyclohexane	5.470	56	38306	20.638	ug/l	99
32) 1,1,1-Trichloroethane	5.385	97	36444	20.702	ug/l	99
36) 1,1-Dichloropropene	5.696	75	31084	21.011	ug/l	99
37) Ethyl Acetate	4.715	43	34629	20.596	ug/l	100
38) Carbon Tetrachloride	5.684	117	32231	20.775	ug/l	98
39) Methylcyclohexane	7.379	83	38964	21.476	ug/l	97
40) Benzene	6.037	78	88751	20.583	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX020922\
 Data File : VX026868.D
 Acq On : 09 Feb 2022 14:39
 Operator : JC/MD
 Sample : VX0209WBSD01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0209WBSD01

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 02/10/2022
 Supervised By :Mahesh Dadoda 02/10/2022

Quant Time: Feb 10 01:05:30 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020822W.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 08 07:07:46 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	4.922	41	18686	20.506	ug/l	97
42) 1,2-Dichloroethane	6.086	62	32780	21.574	ug/l	98
43) Isopropyl Acetate	6.336	43	53624	20.729	ug/l	99
44) Trichloroethene	7.129	130	26083	20.147	ug/l	95
45) 1,2-Dichloropropane	7.434	63	22363	20.499	ug/l	97
46) Dibromomethane	7.580	93	15714	20.817	ug/l	98
47) Bromodichloromethane	7.824	83	30409	21.036	ug/l	98
48) Methyl methacrylate	7.696	41	26242	20.473	ug/l	97
49) 1,4-Dioxane	7.677	88	9568	433.180	ug/l	98
51) 4-Methyl-2-Pentanone	8.574	43	165134	105.340	ug/l	99
52) Toluene	8.720	92	54839	20.088	ug/l	100
53) t-1,3-Dichloropropene	8.982	75	29632	19.774	ug/l	99
54) cis-1,3-Dichloropropene	8.366	75	34543	21.640	ug/l	95
55) 1,1,2-Trichloroethane	9.153	97	21773	21.068	ug/l	99
56) Ethyl methacrylate	9.116	69	34525	21.075	ug/l	97
57) 1,3-Dichloropropane	9.311	76	37352	21.047	ug/l	99
58) 2-Chloroethyl Vinyl ether	8.244	63	89021	105.871	ug/l	99
59) 2-Hexanone	9.433	43	124295	108.401	ug/l	98
60) Dibromochloromethane	9.519	129	22679	21.327	ug/l	99
61) 1,2-Dibromoethane	9.610	107	22979	21.045	ug/l	98
64) Tetrachloroethene	9.275	164	27338	20.127	ug/l	97
65) Chlorobenzene	10.079	112	57070	20.405	ug/l	99
66) 1,1,1,2-Tetrachloroethane	10.165	131	21143	20.550	ug/l	100
67) Ethyl Benzene	10.195	91	102359	20.678	ug/l	97
68) m/p-Xylenes	10.305	106	79205	41.588	ug/l	99
69) o-Xylene	10.640	106	38509	20.635	ug/l	97
70) Styrene	10.659	104	63958	21.131	ug/l	98
71) Bromoform	10.799	173	14829	19.778	ug/l #	100
73) Isopropylbenzene	10.963	105	101175	20.317	ug/l	99
74) N-amyl acetate	10.841	43	39927	20.066	ug/l	98
75) 1,1,2,2-Tetrachloroethane	11.213	83	27199	20.126	ug/l	98
76) 1,2,3-Trichloropropane	11.238	75	28730m	20.759	ug/l	
77) Bromobenzene	11.201	156	23519	19.869	ug/l	97
78) n-propylbenzene	11.305	91	114922	20.817	ug/l	100
79) 2-Chlorotoluene	11.366	91	69327	20.324	ug/l	100
80) 1,3,5-Trimethylbenzene	11.451	105	84496	20.616	ug/l	100
81) trans-1,4-Dichloro-2-b...	11.018	75	7707	19.623	ug/l	98
82) 4-Chlorotoluene	11.457	91	79103	20.672	ug/l	98
83) tert-Butylbenzene	11.713	119	79073	20.242	ug/l	98
84) 1,2,4-Trimethylbenzene	11.756	105	84031	20.327	ug/l	97
85) sec-Butylbenzene	11.890	105	103061	20.814	ug/l	99
86) p-Isopropyltoluene	12.012	119	87306	20.813	ug/l	99
87) 1,3-Dichlorobenzene	11.969	146	44430	20.484	ug/l	99
88) 1,4-Dichlorobenzene	12.042	146	44019	20.123	ug/l	97
89) n-Butylbenzene	12.335	91	73122	21.305	ug/l	99
90) Hexachloroethane	12.542	117	12521	20.652	ug/l	95
91) 1,2-Dichlorobenzene	12.335	146	43056	20.572	ug/l	99
92) 1,2-Dibromo-3-Chloropr...	12.945	75	6430	19.348	ug/l	96
93) 1,2,4-Trichlorobenzene	13.591	180	26943	20.398	ug/l	98
94) Hexachlorobutadiene	13.725	225	11747	21.026	ug/l	96
95) Naphthalene	13.774	128	94290	20.380	ug/l	100
96) 1,2,3-Trichlorobenzene	13.963	180	26669	19.940	ug/l	98

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX020922\
Data File : VX026868.D
Acq On : 09 Feb 2022 14:39
Operator : JC/MD
Sample : VX0209WBSD01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0209WBSD01

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 02/10/2022
Supervised By :Mahesh Dadoda 02/10/2022

Quant Time: Feb 10 01:05:30 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020822W.M
Quant Title : SW846 8260
QLast Update : Tue Feb 08 07:07:46 2022
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX020922\
 Data File : VX026868.D
 Acq On : 09 Feb 2022 14:39
 Operator : JC/MD
 Sample : VX0209WBSD01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 6 Sample Multiplier: 1


Instrument :
 MSVOA_X
 ClientSampleId :
 VX0209WBSD01

Manual Integrations APPROVED

Reviewed By : Semsettin Yesilyurt 02/10/2022
 Supervised By : Mahesh Dadoda 02/10/2022

Quant Time: Feb 10 01:05:30 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020822W.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 08 07:07:46 2022
 Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.556 min Scan# 71
Delta R.T. -0.006 min
Lab File: VX026868.D
Acq: 09 Feb 2022 14:39

Instrument :
MSVOA_X
ClientSampleId :
VX0209WBSD01

Tgt Ion:168 Resp: 9731
Ion Ratio Lower Upper
168 100
99 59.1 50.3 75.5




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 02/10/2022
Supervised By :Mahesh Dadoda 02/10/2022



#2
Dichlorodifluoromethane
Concen: 20.852 ug/l
RT: 1.172 min Scan# 14
Delta R.T. 0.000 min
Lab File: VX026868.D
Acq: 09 Feb 2022 14:39


Tgt Ion: 85 Resp: 22064
Ion Ratio Lower Upper
85 100
87 32.9 16.4 49.1





#3
Chloromethane
Concen: 20.169 ug/l
RT: 1.294 min Scan# 34
Delta R.T. 0.000 min
Lab File: VX026868.D
Acq: 09 Feb 2022 14:39




Instrument :
MSVOA_X
ClientSampleId :
VX0209WBSD01




Tgt Ion: 50 Resp: 1985
Ion Ratio Lower Upper
50 100
52 32.2 26.1 39.1

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 02/10/2022
Supervised By :Mahesh Dadoda 02/10/2022




#4
Vinyl Chloride
Concen: 20.339 ug/l
RT: 1.380 min Scan# 48
Delta R.T. 0.000 min
Lab File: VX026868.D
Acq: 09 Feb 2022 14:39




Tgt Ion: 62 Resp: 20958
Ion Ratio Lower Upper
62 100
64 31.0 26.2 39.2





#5
Bromomethane
Concen: 21.839 ug/l
RT: 1.617 min Scan# 81
Delta R.T. 0.006 min
Lab File: VX026868.D
Acq: 09 Feb 2022 14:39




Instrument :
MSVOA_X
ClientSampleId :
VX0209WBSD01



Tgt Ion: 94 Resp: 9549
Ion Ratio Lower Upper
94 100
96 98.6 75.9 113.9




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 02/10/2022
Supervised By :Mahesh Dadoda 02/10/2022



#6
Chloroethane
Concen: 20.197 ug/l
RT: 1.691 min Scan# 99
Delta R.T. -0.000 min
Lab File: VX026868.D
Acq: 09 Feb 2022 14:39

Tgt Ion: 64 Resp: 12432
Ion Ratio Lower Upper
64 100
66 34.4 25.4 38.2









#7
Trichlorofluoromethane
Concen: 21.059 ug/l
RT: 1.892 min Scan# 11
Delta R.T. 0.000 min
Lab File: VX026868.D
Acq: 09 Feb 2022 14:39

Instrument :
MSVOA_X
ClientSampleId :
VX0209WBSD01

Tgt Ion:101 Resp: 3193
Ion Ratio Lower Upper
101 100
103 64.9 51.3 76.9




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 02/10/2022
Supervised By :Mahesh Dadoda 02/10/2022



#8
Diethyl Ether
Concen: 20.957 ug/l
RT: 2.136 min Scan# 172
Delta R.T. 0.000 min
Lab File: VX026868.D
Acq: 09 Feb 2022 14:39

Tgt Ion: 74 Resp: 11772
Ion Ratio Lower Upper
74 100
45 99.3 50.7 152.1









#9
1,1,2-Trichlorotrifluoroethane
Concen: 21.608 ug/l
RT: 2.331 min Scan# 204
Delta R.T. -0.006 min
Lab File: VX026868.D
Acq: 09 Feb 2022 14:39

Instrument :
MSVOA_X
ClientSampleId :
VX0209WBSD01

Tgt Ion:101 Resp: 2147
Ion Ratio Lower Upper
101 100
85 42.9 35.3 52.9
151 77.7 61.4 92.2




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 02/10/2022
Supervised By :Mahesh Dadoda 02/10/2022



#10
Methyl Iodide
Concen: 19.161 ug/l
RT: 2.459 min Scan# 225
Delta R.T. -0.000 min
Lab File: VX026868.D
Acq: 09 Feb 2022 14:39

Tgt Ion:142 Resp: 24409
Ion Ratio Lower Upper
142 100
127 46.7 38.1 57.1
141 14.8 11.2 16.8





#11

Tert butyl alcohol

Concen: 101.775 ug/l

RT: 2.977 min Scan# 314

Delta R.T. -0.024 min

Lab File: VX026868.D

Acq: 09 Feb 2022 14:39

Instrument :

MSVOA_X

ClientSampleId :

VX0209WBSD01

Tgt Ion: 59 Resp: 2147

Ion Ratio Lower Upper

59 100





57 10.1 7.2 10.8

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 02/10/2022

Supervised By :Mahesh Dadoda 02/10/2022



#12

1,1-Dichloroethene

Concen: 20.484 ug/l

RT: 2.325 min Scan# 203

Delta R.T. 0.000 min

Lab File: VX026868.D

Acq: 09 Feb 2022 14:39




Tgt Ion: 96 Resp: 20400


Ion Ratio Lower Upper

96 100

61 166.3 138.5 207.7

98 65.1 51.8 77.6





#13

Acrolein

Concen: 99.592 ug/l

RT: 2.239 min Scan# 11

Delta R.T. 0.000 min

Lab File: VX026868.D

Acq: 09 Feb 2022 14:39

Instrument :

MSVOA_X

ClientSampleId :

VX0209WBSD01

Tgt Ion: 56 Resp: 1942

Ion Ratio Lower Upper

56 100





55 72.5 56.9 85.3

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 02/10/2022

Supervised By :Mahesh Dadoda 02/10/2022



#14

Allyl chloride

Concen: 20.760 ug/l

RT: 2.666 min Scan# 259

Delta R.T. 0.000 min

Lab File: VX026868.D

Acq: 09 Feb 2022 14:39




Tgt Ion: 41 Resp: 36622


Ion Ratio Lower Upper

41 100

39 63.9 54.2 81.4


76 32.6 25.2 37.8





#15
Acrylonitrile
Concen: 101.340 ug/l
RT: 3.068 min Scan# 31
Delta R.T. 0.000 min
Lab File: VX026868.D
Acq: 09 Feb 2022 14:39




Instrument :
MSVOA_X
ClientSampleId :
VX0209WBSD01



Tgt Ion: 53 Resp: 6128
Ion Ratio Lower Upper
53 100
52 83.3 66.6 100.0
51 36.0 29.5 44.3




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 02/10/2022
Supervised By :Mahesh Dadoda 02/10/2022



#16
Acetone
Concen: 115.208 ug/l
RT: 2.386 min Scan# 213
Delta R.T. 0.000 min
Lab File: VX026868.D
Acq: 09 Feb 2022 14:39


Tgt Ion: 43 Resp: 59107
Ion Ratio Lower Upper
43 100
58 30.5 23.0 34.6





#17
Carbon Disulfide
Concen: 19.397 ug/l
RT: 2.514 min Scan# 21
Delta R.T. 0.000 min
Lab File: VX026868.D
Acq: 09 Feb 2022 14:39




Instrument :
MSVOA_X
ClientSampleId :
VX0209WBSD01



Tgt Ion: 76 Resp: 54684
Ion Ratio Lower Upper
76 100
78 9.0 7.5 11.3




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 02/10/2022
Supervised By :Mahesh Dadoda 02/10/2022



#18
Methyl Acetate
Concen: 19.517 ug/l
RT: 2.703 min Scan# 265
Delta R.T. -0.006 min
Lab File: VX026868.D
Acq: 09 Feb 2022 14:39

Tgt Ion: 43 Resp: 28810
Ion Ratio Lower Upper
43 100
74 22.3 18.3 27.5









#19
Methyl tert-butyl Ether
Concen: 20.264 ug/l
RT: 3.117 min Scan# 311
Delta R.T. 0.000 min
Lab File: VX026868.D
Acq: 09 Feb 2022 14:39

Instrument :
MSVOA_X
ClientSampleId :
VX0209WBSD01

Tgt Ion: 73 Resp: 7143
Ion Ratio Lower Upper
73 100
57 21.6 17.4 26.2




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 02/10/2022
Supervised By :Mahesh Dadoda 02/10/2022



#20
Methylene Chloride
Concen: 19.876 ug/l
RT: 2.794 min Scan# 280
Delta R.T. 0.000 min
Lab File: VX026868.D
Acq: 09 Feb 2022 14:39


Tgt Ion: 84 Resp: 22096
Ion Ratio Lower Upper
84 100
49 121.0 101.5 152.3
51 34.4 31.2 46.8
86 64.5 51.3 76.9





#21
trans-1,2-Dichloroethene
Concen: 20.908 ug/l
RT: 3.093 min Scan# 31
Delta R.T. -0.006 min
Lab File: VX026868.D
Acq: 09 Feb 2022 14:39




Instrument :
MSVOA_X
ClientSampleId :
VX0209WBSD01




Tgt Ion: 96 Resp: 2191
Ion Ratio Lower Upper
96 100
61 132.7 112.9 169.3
98 59.5 49.5 74.3

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 02/10/2022
Supervised By :Mahesh Dadoda 02/10/2022




#22
Diisopropyl ether
Concen: 20.968 ug/l
RT: 3.757 min Scan# 438
Delta R.T. -0.006 min
Lab File: VX026868.D
Acq: 09 Feb 2022 14:39



Tgt Ion: 45 Resp: 74471
Ion Ratio Lower Upper
45 100
43 95.9 60.6 91.0#
87 27.8 22.0 33.0
59 11.0 9.0 13.4









#23
 Vinyl Acetate
 Concen: 106.224 ug/l
 RT: 3.721 min Scan# 41
 Delta R.T. -0.006 min
 Lab File: VX026868.D
 Acq: 09 Feb 2022 14:39

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0209WBSD01

Tgt Ion: 43 Resp: 317192
 Ion Ratio Lower Upper
 43 100
 86 10.4 8.1 12.1




Manual Integrations
 APPROVED


Reviewed By :Semsettin Yesilyurt 02/10/2022
 Supervised By :Mahesh Dadoda 02/10/2022



#24
 1,1-Dichloroethane
 Concen: 20.696 ug/l
 RT: 3.611 min Scan# 414
 Delta R.T. -0.006 min
 Lab File: VX026868.D
 Acq: 09 Feb 2022 14:39


Tgt Ion: 63 Resp: 39774
 Ion Ratio Lower Upper
 63 100
 98 6.7 3.7 11.1
 100 4.4 2.4 7.2





#25
2-Butanone
Concen: 105.751 ug/l
RT: 4.556 min Scan# 570
Delta R.T. -0.006 min
Lab File: VX026868.D
Acq: 09 Feb 2022 14:39




Instrument :
MSVOA_X
ClientSampleId :
VX0209WBSD01



Tgt Ion: 43 Resp: 87414
Ion Ratio Lower Upper
43 100
72 25.9 20.4 30.6




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 02/10/2022
Supervised By :Mahesh Dadoda 02/10/2022



#26
2,2-Dichloropropane
Concen: 23.035 ug/l
RT: 4.477 min Scan# 556
Delta R.T. -0.006 min
Lab File: VX026868.D
Acq: 09 Feb 2022 14:39

Tgt Ion: 77 Resp: 29911
Ion Ratio Lower Upper
77 100
97 22.9 11.0 33.0





#27

cis-1,2-Dichloroethene

Concen: 20.918 ug/l

RT: 4.489 min Scan# 51

Delta R.T. -0.006 min

Lab File: VX026868.D

Acq: 09 Feb 2022 14:39

Instrument :

MSVOA_X

ClientSampleId :

VX0209WBSD01

Tgt Ion: 96 Resp: 25169

Ion Ratio Lower Upper

96 100

61 143.2 0.0 296.6





98 65.4 0.0 128.8

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 02/10/2022

Supervised By :Mahesh Dadoda 02/10/2022



#28

Bromochloromethane

Concen: 21.559 ug/l

RT: 4.897 min Scan# 625

Delta R.T. -0.006 min

Lab File: VX026868.D

Acq: 09 Feb 2022 14:39




Tgt Ion: 49 Resp: 17175


Ion Ratio Lower Upper

49 100

129 2.1 0.0 4.2

130 78.1 62.4 93.6





#29

Tetrahydrofuran

Concen: 98.699 ug/l

RT: 5.007 min Scan# 64

Delta R.T. 0.000 min

Lab File: VX026868.D

Acq: 09 Feb 2022 14:39

Instrument :

MSVOA_X

ClientSampleId :





VX0209WBSD01

Tgt Ion:	42	Resp:	5546
Ion Ratio	Lower	Upper	
42	100		
72	45.0	34.6	52.0
71	41.2	31.8	47.8

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 02/10/2022
Supervised By :Mahesh Dadoda 02/10/2022



#30

Chloroform

Concen: 20.432 ug/l




RT: 5.092 min Scan# 657


Delta R.T. -0.006 min

Lab File: VX026868.D

Acq: 09 Feb 2022 14:39


Tgt Ion:	83	Resp:	40798
Ion Ratio	Lower	Upper	
83	100		
85	66.4	51.6	77.4





#31
Cyclohexane
Concen: 20.638 ug/l
RT: 5.470 min Scan# 719
Delta R.T. -0.006 min
Lab File: VX026868.D
Acq: 09 Feb 2022 14:39




Instrument :
MSVOA_X
ClientSampleId :
VX0209WBSD01




Tgt Ion: 56 Resp: 38300
Ion Ratio Lower Upper
56 100
69 31.8 25.8 38.6
84 86.6 68.9 103.3

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 02/10/2022
Supervised By :Mahesh Dadoda 02/10/2022




#32
1,1,1-Trichloroethane
Concen: 20.702 ug/l
RT: 5.385 min Scan# 705
Delta R.T. -0.006 min
Lab File: VX026868.D
Acq: 09 Feb 2022 14:39



Tgt Ion: 97 Resp: 36444
Ion Ratio Lower Upper
97 100
99 64.1 51.3 76.9
61 42.7 35.0 52.6





#33

1,2-Dichloroethane-d4

Concen: 55.651 ug/l

RT: 5.958 min Scan# 799

Delta R.T. -0.006 min

Lab File: VX026868.D

Acq: 09 Feb 2022 14:39

Instrument :

MSVOA_X

ClientSampleId :

VX0209WBSD01

Tgt Ion: 65 Resp: 6500

Ion Ratio Lower Upper

65 100





67 51.4 0.0 100.8

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 02/10/2022

Supervised By :Mahesh Dadoda 02/10/2022



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 6.763 min Scan# 931

Delta R.T. -0.006 min

Lab File: VX026868.D

Acq: 09 Feb 2022 14:39




Tgt Ion:114 Resp: 157407


Ion Ratio Lower Upper

114 100

63 19.8 0.0 42.8


88 16.1 0.0 33.0





#35
Dibromofluoromethane
Concen: 54.214 ug/l
RT: 5.391 min Scan# 706
Delta R.T. 0.000 min
Lab File: VX026868.D
Acq: 09 Feb 2022 14:39




Instrument :
MSVOA_X
ClientSampleId :
VX0209WBSD01




Tgt Ion: 113 Resp: 5084
Ion Ratio Lower Upper
113 100
111 104.3 81.2 121.8
192 19.7 15.0 22.4

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 02/10/2022
Supervised By :Mahesh Dadoda 02/10/2022




#36
1,1-Dichloropropene
Concen: 21.011 ug/l
RT: 5.696 min Scan# 756
Delta R.T. 0.000 min
Lab File: VX026868.D
Acq: 09 Feb 2022 14:39



Tgt Ion: 75 Resp: 31084
Ion Ratio Lower Upper
75 100
110 35.6 17.2 51.6
77 30.6 24.6 36.8









#37
Ethyl Acetate
Concen: 20.596 ug/l
RT: 4.715 min Scan# 595
Delta R.T. 0.000 min
Lab File: VX026868.D
Acq: 09 Feb 2022 14:39

Instrument :
MSVOA_X
ClientSampleId :
VX0209WBSD01

Tgt Ion: 43 Resp: 34629
Ion Ratio Lower Upper
43 100
61 13.4 10.6 16.0
70 10.6 8.5 12.7




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 02/10/2022
Supervised By :Mahesh Dadoda 02/10/2022



#38
Carbon Tetrachloride
Concen: 20.775 ug/l
RT: 5.684 min Scan# 754
Delta R.T. 0.000 min
Lab File: VX026868.D
Acq: 09 Feb 2022 14:39

Tgt Ion:117 Resp: 32231
Ion Ratio Lower Upper
117 100
119 94.9 74.7 112.1
121 30.6 25.4 38.0





#39

Methylcyclohexane

Concen: 21.476 ug/l

RT: 7.379 min Scan# 1032

Delta R.T. -0.006 min

Lab File: VX026868.D

Acq: 09 Feb 2022 14:39

Instrument :

MSVOA_X

ClientSampleId :





VX0209WBSD01

Tgt Ion: 83	Resp: 38964
Ion Ratio	Lower Upper
83 100	
55 81.3	63.6 95.4
98 51.6	38.4 57.6

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 02/10/2022
Supervised By :Mahesh Dadoda 02/10/2022



#40

Benzene

Concen: 20.583 ug/l




RT: 6.037 min Scan# 812


Delta R.T. -0.006 min

Lab File: VX026868.D

Acq: 09 Feb 2022 14:39

Tgt Ion: 78	Resp: 88751
Ion Ratio	Lower Upper
78 100	
77 23.3	18.9 28.3






#41
Methacrylonitrile
Concen: 20.506 ug/l
RT: 4.922 min Scan# 61
Delta R.T. 0.000 min
Lab File: VX026868.D
Acq: 09 Feb 2022 14:39

Instrument :

MSVOA_X

ClientSampleId :

VX0209WBSD01






Tgt Ion: 41 Resp: 18680
Ion Ratio Lower Upper
41 100
39 56.2 45.4 68.0
67 70.4 53.2 79.8
52 29.5 23.0 34.6

Manual Integrations




APPROVED


Reviewed By :Semsettin Yesilyurt 02/10/2022
Supervised By :Mahesh Dadoda 02/10/2022



#42
1,2-Dichloroethane
Concen: 21.574 ug/l
RT: 6.086 min Scan# 820
Delta R.T. -0.006 min
Lab File: VX026868.D
Acq: 09 Feb 2022 14:39


Tgt Ion: 62 Resp: 32780
Ion Ratio Lower Upper
62 100
98 9.6 0.0 17.8





#43
Isopropyl Acetate
Concen: 20.729 ug/l
RT: 6.336 min Scan# 861
Delta R.T. -0.006 min
Lab File: VX026868.D
Acq: 09 Feb 2022 14:39




Instrument :
MSVOA_X
ClientSampleId :
VX0209WBSD01



Tgt Ion: 43 Resp: 53624
Ion Ratio Lower Upper
43 100
61 19.9 15.4 23.0
87 13.3 10.2 15.2




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 02/10/2022
Supervised By :Mahesh Dadoda 02/10/2022



#44
Trichloroethene
Concen: 20.147 ug/l
RT: 7.129 min Scan# 991
Delta R.T. 0.000 min
Lab File: VX026868.D
Acq: 09 Feb 2022 14:39

Tgt Ion:130 Resp: 26083
Ion Ratio Lower Upper
130 100
95 96.5 0.0 202.6





#45

1,2-Dichloropropane

Concen: 20.499 ug/l

RT: 7.434 min Scan# 1041

Delta R.T. 0.000 min

Lab File: VX026868.D

Acq: 09 Feb 2022 14:39

Instrument :

MSVOA_X

ClientSampleId :

VX0209WBSD01

Tgt Ion: 63 Resp: 2236

Ion Ratio Lower Upper

63 100





65 31.1 23.8 35.6

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 02/10/2022

Supervised By :Mahesh Dadoda 02/10/2022



#46

Dibromomethane

Concen: 20.817 ug/l

RT: 7.580 min Scan# 1065

Delta R.T. -0.006 min

Lab File: VX026868.D

Acq: 09 Feb 2022 14:39




Tgt Ion: 93 Resp: 15714


Ion Ratio Lower Upper

93 100

95 80.7 66.1 99.1


174 100.1 78.7 118.1





#47
Bromodichloromethane
Concen: 21.036 ug/l
RT: 7.824 min Scan# 1105
Delta R.T. 0.000 min
Lab File: VX026868.D
Acq: 09 Feb 2022 14:39




Instrument :
MSVOA_X
ClientSampleId :
VX0209WBSD01




Tgt Ion: 83 Resp: 30409
Ion Ratio Lower Upper
83 100
85 66.0 51.4 77.0
127 8.4 6.9 10.3

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 02/10/2022
Supervised By :Mahesh Dadoda 02/10/2022




#48
Methyl methacrylate
Concen: 20.473 ug/l
RT: 7.696 min Scan# 1084
Delta R.T. 0.000 min
Lab File: VX026868.D
Acq: 09 Feb 2022 14:39




Tgt Ion: 41 Resp: 26242
Ion Ratio Lower Upper
41 100
69 80.7 62.2 93.4
39 55.0 45.4 68.0





#49
1,4-Dioxane
Concen: 433.180 ug/l
RT: 7.677 min Scan# 1081
Delta R.T. 0.000 min
Lab File: VX026868.D
Acq: 09 Feb 2022 14:39




Instrument :
MSVOA_X
ClientSampleId :
VX0209WBSD01



Tgt Ion: 88 Resp: 956
Ion Ratio Lower Upper
88 100
43 36.2 29.9 44.9
58 70.7 57.8 86.6




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 02/10/2022
Supervised By :Mahesh Dadoda 02/10/2022



#50
Toluene-d8
Concen: 54.503 ug/l
RT: 8.647 min Scan# 1240
Delta R.T. -0.006 min
Lab File: VX026868.D
Acq: 09 Feb 2022 14:39


Tgt Ion: 98 Resp: 187621
Ion Ratio Lower Upper
98 100
100 66.2 53.2 79.8





#51
4-Methyl-2-Pentanone
Concen: 105.340 ug/l
RT: 8.574 min Scan# 1228
Delta R.T. 0.000 min
Lab File: VX026868.D
Acq: 09 Feb 2022 14:39




Instrument :
MSVOA_X
ClientSampleId :
VX0209WBSD01



Tgt Ion: 43 Resp: 165134
Ion Ratio Lower Upper
43 100
58 39.0 30.5 45.7




Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 02/10/2022
Supervised By :Mahesh Dadoda 02/10/2022



#52
Toluene
Concen: 20.088 ug/l
RT: 8.720 min Scan# 1252
Delta R.T. 0.000 min
Lab File: VX026868.D
Acq: 09 Feb 2022 14:39

Tgt Ion: 92 Resp: 54839
Ion Ratio Lower Upper
92 100
91 170.2 135.7 203.5





#53

t-1,3-Dichloropropene

Concen: 19.774 ug/l

RT: 8.982 min Scan# 1194

Delta R.T. 0.000 min

Lab File: VX026868.D

Acq: 09 Feb 2022 14:39

Instrument :

MSVOA_X

ClientSampleId :

VX0209WBSD01

Tgt Ion: 75 Resp: 2963

Ion Ratio Lower Upper

75 100





77 32.8 25.7 38.5

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 02/10/2022

Supervised By :Mahesh Dadoda 02/10/2022



#54

cis-1,3-Dichloropropene

Concen: 21.640 ug/l

RT: 8.366 min Scan# 1194

Delta R.T. 0.000 min

Lab File: VX026868.D

Acq: 09 Feb 2022 14:39




Tgt Ion: 75 Resp: 34543


Ion Ratio Lower Upper

75 100

77 31.3 26.2 39.2


39 50.1 43.5 65.3





#55
1,1,2-Trichloroethane
Concen: 21.068 ug/l
RT: 9.153 min Scan# 1323
Delta R.T. 0.000 min
Lab File: VX026868.D
Acq: 09 Feb 2022 14:39




Instrument :
MSVOA_X
ClientSampleId :
VX0209WBSD01




Tgt Ion: 97 Resp: 2177
Ion Ratio Lower Upper
97 100
83 84.7 68.4 102.6
85 54.8 44.0 66.0
99 61.0 49.8 74.6

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 02/10/2022
Supervised By :Mahesh Dadoda 02/10/2022




#56
Ethyl methacrylate
Concen: 21.075 ug/l
RT: 9.116 min Scan# 1317
Delta R.T. 0.000 min
Lab File: VX026868.D
Acq: 09 Feb 2022 14:39



Tgt Ion: 69 Resp: 34525
Ion Ratio Lower Upper
69 100
41 68.7 57.2 85.8
39 36.6 30.9 46.3





#57

1,3-Dichloropropane

Concen: 21.047 ug/l

RT: 9.311 min Scan# 11349

Delta R.T. 0.000 min

Lab File: VX026868.D

Acq: 09 Feb 2022 14:39

Instrument :

MSVOA_X

ClientSampleId :

VX0209WBSD01

Tgt Ion: 76 Resp: 3735

Ion Ratio Lower Upper

76 100





78 32.5 25.7 38.5

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 02/10/2022

Supervised By :Mahesh Dadoda 02/10/2022



#58

2-Chloroethyl Vinyl ether

Concen: 105.871 ug/l

RT: 8.244 min Scan# 1174

Delta R.T. 0.000 min

Lab File: VX026868.D




Acq: 09 Feb 2022 14:39


Tgt Ion: 63 Resp: 89021

Ion Ratio Lower Upper

63 100

106 29.1 23.0 34.6









#59
2-Hexanone
Concen: 108.401 ug/l
RT: 9.433 min Scan# 11
Delta R.T. 0.000 min
Lab File: VX026868.D
Acq: 09 Feb 2022 14:39

Instrument :
MSVOA_X
ClientSampleId :
VX0209WBSD01

Tgt Ion: 43 Resp: 12429
Ion Ratio Lower Upper
43 100
58 53.9 26.4 79.2




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 02/10/2022
Supervised By :Mahesh Dadoda 02/10/2022



#60
Dibromochloromethane
Concen: 21.327 ug/l
RT: 9.519 min Scan# 1383
Delta R.T. -0.006 min
Lab File: VX026868.D
Acq: 09 Feb 2022 14:39


Tgt Ion:129 Resp: 22679
Ion Ratio Lower Upper
129 100
127 76.0 38.5 115.5





#61
1,2-Dibromoethane
Concen: 21.045 ug/l
RT: 9.610 min Scan# 1398
Delta R.T. 0.000 min
Lab File: VX026868.D
Acq: 09 Feb 2022 14:39




Instrument :
MSVOA_X
ClientSampleId :
VX0209WBSD01




Tgt Ion:107 Resp: 22979
Ion Ratio Lower Upper
107 100
109 92.2 75.5 113.3

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 02/10/2022
Supervised By :Mahesh Dadoda 02/10/2022




#62
4-Bromofluorobenzene
Concen: 56.103 ug/l
RT: 11.079 min Scan# 1639
Delta R.T. 0.000 min
Lab File: VX026868.D
Acq: 09 Feb 2022 14:39




Tgt Ion: 95 Resp: 67994
Ion Ratio Lower Upper
95 100
174 76.0 0.0 148.6
176 72.7 0.0 141.4





#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.055 min Scan# 1471
Delta R.T. 0.000 min
Lab File: VX026868.D
Acq: 09 Feb 2022 14:39




Instrument :
MSVOA_X
ClientSampleId :
VX0209WBSD01



Tgt Ion:117 Resp: 135238
Ion Ratio Lower Upper
117 100
82 54.6 45.2 67.8
119 31.2 26.2 39.2




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 02/10/2022
Supervised By :Mahesh Dadoda 02/10/2022



#64
Tetrachloroethene
Concen: 20.127 ug/l
RT: 9.275 min Scan# 1343
Delta R.T. 0.000 min
Lab File: VX026868.D
Acq: 09 Feb 2022 14:39


Tgt Ion:164 Resp: 27338
Ion Ratio Lower Upper
164 100
166 128.0 101.8 152.6
129 96.5 83.3 124.9
131 97.7 78.0 117.0





#65
Chlorobenzene
Concen: 20.405 ug/l
RT: 10.079 min Scan# 1475
Delta R.T. 0.000 min
Lab File: VX026868.D
Acq: 09 Feb 2022 14:39




Instrument :
MSVOA_X
ClientSampleId :
VX0209WBSD01



Tgt Ion:112 Resp: 57070
Ion Ratio Lower Upper
112 100
114 31.0 25.3 37.9




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 02/10/2022
Supervised By :Mahesh Dadoda 02/10/2022



#66
1,1,1,2-Tetrachloroethane
Concen: 20.550 ug/l
RT: 10.165 min Scan# 1489
Delta R.T. 0.000 min
Lab File: VX026868.D
Acq: 09 Feb 2022 14:39


Tgt Ion:131 Resp: 21143
Ion Ratio Lower Upper
131 100
133 95.5 47.9 143.8
119 63.1 31.7 95.1





#67
Ethyl Benzene
Concen: 20.678 ug/l
RT: 10.195 min Scan# 1494
Delta R.T. 0.000 min
Lab File: VX026868.D
Acq: 09 Feb 2022 14:39




Instrument :
MSVOA_X
ClientSampleId :
VX0209WBSD01




Tgt Ion: 91 Resp: 10235
Ion Ratio Lower Upper
91 100
106 32.3 24.7 37.1

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 02/10/2022
Supervised By :Mahesh Dadoda 02/10/2022




#68
m/p-Xylenes
Concen: 41.588 ug/l
RT: 10.305 min Scan# 1512
Delta R.T. 0.000 min
Lab File: VX026868.D
Acq: 09 Feb 2022 14:39



Tgt Ion:106 Resp: 79205
Ion Ratio Lower Upper
106 100
91 204.9 165.3 247.9









#69
o-Xylene
Concen: 20.635 ug/l
RT: 10.640 min Scan# 1567
Delta R.T. 0.000 min
Lab File: VX026868.D
Acq: 09 Feb 2022 14:39

Instrument :
MSVOA_X
ClientSampleId :
VX0209WBSD01

Tgt Ion:106 Resp: 38509
Ion Ratio Lower Upper
106 100
91 213.9 109.7 329.0




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 02/10/2022
Supervised By :Mahesh Dadoda 02/10/2022



#70
Styrene
Concen: 21.131 ug/l
RT: 10.659 min Scan# 1570
Delta R.T. 0.000 min
Lab File: VX026868.D
Acq: 09 Feb 2022 14:39


Tgt Ion:104 Resp: 63958
Ion Ratio Lower Upper
104 100
78 51.1 42.5 63.7
103 54.4 43.3 64.9





#71
Bromoform
Concen: 19.778 ug/l
RT: 10.799 min Scan# 1593
Delta R.T. -0.000 min
Lab File: VX026868.D
Acq: 09 Feb 2022 14:39




Instrument :
MSVOA_X
ClientSampleId :
VX0209WBSD01




Tgt Ion:173 Resp: 14829
Ion Ratio Lower Upper
173 100
175 48.8 24.3 72.9
254 0.0 0.0 0.0

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 02/10/2022
Supervised By :Mahesh Dadoda 02/10/2022




#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.024 min Scan# 1794
Delta R.T. 0.000 min
Lab File: VX026868.D
Acq: 09 Feb 2022 14:39



Tgt Ion:152 Resp: 63222
Ion Ratio Lower Upper
152 100
115 70.9 44.2 132.6
150 163.6 0.0 348.2





#73

Isopropylbenzene

Concen: 20.317 ug/l

RT: 10.963 min Scan# 1620

Delta R.T. 0.000 min

Lab File: VX026868.D

Acq: 09 Feb 2022 14:39

Instrument :

MSVOA_X

ClientSampleId :

VX0209WBSD01

Tgt Ion:105 Resp: 10117

Ion Ratio Lower Upper

105 100





120 26.7 13.1 39.1

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 02/10/2022

Supervised By :Mahesh Dadoda 02/10/2022



#74

N-ethyl acetate

Concen: 20.066 ug/l

RT: 10.841 min Scan# 1600

Delta R.T. -0.006 min

Lab File: VX026868.D

Acq: 09 Feb 2022 14:39

Tgt Ion: 43 Resp: 39927




Ion Ratio Lower Upper


43 100

70 42.6 32.9 49.3

55 24.3 18.3 27.5

61 24.0 18.9 28.3









#75
1,1,2,2-Tetrachloroethane
Concen: 20.126 ug/l
RT: 11.213 min Scan# 1661
Delta R.T. 0.000 min
Lab File: VX026868.D
Acq: 09 Feb 2022 14:39

Instrument :
MSVOA_X
ClientSampleId :
VX0209WBSD01

Tgt Ion: 83 Resp: 27199
Ion Ratio Lower Upper
83 100
131 10.5 5.3 15.9
85 65.5 32.0 96.2




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 02/10/2022
Supervised By :Mahesh Dadoda 02/10/2022



#76
1,2,3-Trichloropropane
Concen: 20.759 ug/l m
RT: 11.238 min Scan# 1665
Delta R.T. -0.006 min
Lab File: VX026868.D
Acq: 09 Feb 2022 14:39


Tgt Ion: 75 Resp: 28730
Ion Ratio Lower Upper
75 100
77 37.5 18.3 54.8





#77
Bromobenzene
Concen: 19.869 ug/l
RT: 11.201 min Scan# 1659
Delta R.T. 0.000 min
Lab File: VX026868.D
Acq: 09 Feb 2022 14:39




Instrument :
MSVOA_X
ClientSampleId :
VX0209WBSD01




Tgt Ion:156 Resp: 23519
Ion Ratio Lower Upper
156 100
77 164.0 84.5 253.4
158 98.4 48.1 144.4

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 02/10/2022
Supervised By :Mahesh Dadoda 02/10/2022




#78
n-propylbenzene
Concen: 20.817 ug/l
RT: 11.305 min Scan# 1676
Delta R.T. 0.000 min
Lab File: VX026868.D
Acq: 09 Feb 2022 14:39



Tgt Ion: 91 Resp: 114922
Ion Ratio Lower Upper
91 100
120 23.1 11.6 34.7





#79

2-Chlorotoluene

Concen: 20.324 ug/l

RT: 11.366 min Scan# 1

Delta R.T. 0.000 min

Lab File: VX026868.D

Acq: 09 Feb 2022 14:39

Instrument :

MSVOA_X

ClientSampleId :

VX0209WBSD01

Tgt Ion: 91 Resp: 6932

Ion Ratio Lower Upper

91 100





126 33.0 16.4 49.2

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 02/10/2022

Supervised By :Mahesh Dadoda 02/10/2022



#80

1,3,5-Trimethylbenzene

Concen: 20.616 ug/l

RT: 11.451 min Scan# 1700

Delta R.T. 0.000 min

Lab File: VX026868.D




Acq: 09 Feb 2022 14:39


Tgt Ion:105 Resp: 84496

Ion Ratio Lower Upper

105 100


120 48.2 24.0 72.0





#81
trans-1,4-Dichloro-2-butene
Concen: 19.623 ug/l
RT: 11.018 min Scan# 1629
Delta R.T. 0.000 min
Lab File: VX026868.D
Acq: 09 Feb 2022 14:39




Instrument :
MSVOA_X
ClientSampleId :
VX0209WBSD01




Tgt Ion: 75 Resp: 770
Ion Ratio Lower Upper
75 100
53 114.5 94.0 141.0
89 45.3 37.0 55.4

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 02/10/2022
Supervised By :Mahesh Dadoda 02/10/2022




#82
4-Chlorotoluene
Concen: 20.672 ug/l
RT: 11.457 min Scan# 1701
Delta R.T. 0.000 min
Lab File: VX026868.D
Acq: 09 Feb 2022 14:39



Tgt Ion: 91 Resp: 79103
Ion Ratio Lower Upper
91 100
126 29.4 14.1 42.4





#83

tert-Butylbenzene

Concen: 20.242 ug/l

RT: 11.713 min Scan# 1743

Delta R.T. 0.000 min

Lab File: VX026868.D

Acq: 09 Feb 2022 14:39

Instrument :

MSVOA_X

ClientSampleId :

VX0209WBSD01





Tgt Ion:119 Resp: 7907

Ion	Ratio	Lower	Upper
119	100		
91	61.3	31.4	94.0
134	24.1	11.8	35.3

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 02/10/2022
Supervised By :Mahesh Dadoda 02/10/2022



#84

1,2,4-Trimethylbenzene

Concen: 20.327 ug/l

RT: 11.756 min Scan# 1750




Delta R.T. 0.006 min


Lab File: VX026868.D

Acq: 09 Feb 2022 14:39

Tgt Ion:105 Resp: 84031


Ion	Ratio	Lower	Upper
105	100		
120	45.5	21.9	65.5





#85
 sec-Butylbenzene
 Concen: 20.814 ug/l
 RT: 11.890 min Scan# 1772
 Delta R.T. 0.000 min
 Lab File: VX026868.D
 Acq: 09 Feb 2022 14:39




Instrument :
 MSVOA_X
 ClientSampleId :
 VX0209WBSD01




Tgt Ion:105 Resp: 10306
 Ion Ratio Lower Upper
 105 100
 134 20.3 10.1 30.1

Manual Integrations
 APPROVED



Reviewed By :Semsettin Yesilyurt 02/10/2022
 Supervised By :Mahesh Dadoda 02/10/2022




#86
 p-Isopropyltoluene
 Concen: 20.813 ug/l
 RT: 12.012 min Scan# 1792
 Delta R.T. 0.000 min
 Lab File: VX026868.D
 Acq: 09 Feb 2022 14:39




Tgt Ion:119 Resp: 87306
 Ion Ratio Lower Upper
 119 100
 134 25.7 13.0 39.0
 91 24.2 12.6 37.6





#87
1,3-Dichlorobenzene
Concen: 20.484 ug/l
RT: 11.969 min Scan# 1785
Delta R.T. 0.000 min
Lab File: VX026868.D
Acq: 09 Feb 2022 14:39



Instrument :
MSVOA_X
ClientSampleId :
VX0209WBSD01




Tgt Ion:146 Resp: 44430
Ion Ratio Lower Upper
146 100
111 41.5 21.3 63.7
148 63.1 31.6 95.0

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 02/10/2022
Supervised By :Mahesh Dadoda 02/10/2022




#88
1,4-Dichlorobenzene
Concen: 20.123 ug/l
RT: 12.042 min Scan# 1797
Delta R.T. 0.000 min
Lab File: VX026868.D
Acq: 09 Feb 2022 14:39




Tgt Ion:146 Resp: 44019
Ion Ratio Lower Upper
146 100
111 42.1 20.2 60.6
148 65.1 31.4 94.0





#89
n-Butylbenzene
Concen: 21.305 ug/l
RT: 12.335 min Scan# 1845
Delta R.T. 0.000 min
Lab File: VX026868.D
Acq: 09 Feb 2022 14:39




Instrument :
MSVOA_X
ClientSampleId :
VX0209WBSD01



Tgt Ion: 91 Resp: 7312
Ion Ratio Lower Upper
91 100
92 54.7 27.2 81.6
134 26.9 13.1 39.1




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 02/10/2022
Supervised By :Mahesh Dadoda 02/10/2022



#90
Hexachloroethane
Concen: 20.652 ug/l
RT: 12.542 min Scan# 1879
Delta R.T. 0.000 min
Lab File: VX026868.D
Acq: 09 Feb 2022 14:39


Tgt Ion:117 Resp: 12521
Ion Ratio Lower Upper
117 100
201 75.0 35.6 106.8





#91
1,2-Dichlorobenzene
Concen: 20.572 ug/l
RT: 12.335 min Scan# 1845
Delta R.T. 0.000 min
Lab File: VX026868.D
Acq: 09 Feb 2022 14:39




Instrument :
MSVOA_X
ClientSampleId :
VX0209WBSD01




Tgt Ion:146 Resp: 43050
Ion Ratio Lower Upper
146 100
111 42.4 21.8 65.3
148 63.9 32.0 96.2

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 02/10/2022
Supervised By :Mahesh Dadoda 02/10/2022




#92
1,2-Dibromo-3-Chloropropane
Concen: 19.348 ug/l
RT: 12.945 min Scan# 1945
Delta R.T. 0.000 min
Lab File: VX026868.D
Acq: 09 Feb 2022 14:39




Tgt Ion: 75 Resp: 6430
Ion Ratio Lower Upper
75 100
155 85.0 40.0 119.9
157 105.8 51.6 154.8





#93
1,2,4-Trichlorobenzene
Concen: 20.398 ug/l
RT: 13.591 min Scan# 2051
Delta R.T. 0.000 min
Lab File: VX026868.D
Acq: 09 Feb 2022 14:39




Instrument :
MSVOA_X
ClientSampleId :
VX0209WBSD01




Tgt Ion:180 Resp: 2694
Ion Ratio Lower Upper
180 100
182 95.3 46.7 140.0
145 32.3 16.3 48.8

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 02/10/2022
Supervised By :Mahesh Dadoda 02/10/2022




#94
Hexachlorobutadiene
Concen: 21.026 ug/l
RT: 13.725 min Scan# 2073
Delta R.T. 0.000 min
Lab File: VX026868.D
Acq: 09 Feb 2022 14:39




Tgt Ion:225 Resp: 11747
Ion Ratio Lower Upper
225 100
223 62.0 32.0 96.0
227 60.3 32.3 96.9





#95
Naphthalene
Concen: 20.380 ug/l
RT: 13.774 min Scan# 2081
Delta R.T. -0.006 min
Lab File: VX026868.D
Acq: 09 Feb 2022 14:39




Instrument :
MSVOA_X
ClientSampleId :
VX0209WBSD01



Tgt Ion:128 Resp: 94290
Ion Ratio Lower Upper
128 100
127 12.8 10.3 15.5
129 10.6 8.7 13.1

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 02/10/2022
Supervised By :Mahesh Dadoda 02/10/2022



#96
1,2,3-Trichlorobenzene
Concen: 19.940 ug/l
RT: 13.963 min Scan# 2112
Delta R.T. 0.000 min
Lab File: VX026868.D
Acq: 09 Feb 2022 14:39

Tgt Ion:180 Resp: 26669
Ion Ratio Lower Upper
180 100
182 97.1 47.2 141.6
145 34.6 17.5 52.5

