

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX020922\
 Data File : VX026889.D
 Acq On : 10 Feb 2022 00:15
 Operator : JC/MD
 Sample : VX0209WBSD02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 31 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0209WBSD02

Quant Time: Feb 10 01:13:41 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020822W.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 08 07:07:46 2022
 Response via : Initial Calibration

Manual Integrations
 APPROVED

Reviewed By : Semsettin Yesilyurt 02/10/2022
 Supervised By : Mahesh Dadoda 02/10/2022

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	95329	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	158370	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	143009	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	63876	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	62418	54.546	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery = 109.100%			
35) Dibromofluoromethane	5.391	113	48890	51.816	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 103.640%			
50) Toluene-d8	8.653	98	188717	54.488	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery = 108.980%			
62) 4-Bromofluorobenzene	11.079	95	68190	55.922	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 111.840%			
Target Compounds						
						Qvalue
2) Dichlorodifluoromethane	1.173	85	22148	21.367	ug/l	98
3) Chloromethane	1.294	50	21259	22.046	ug/l	98
4) Vinyl Chloride	1.380	62	21971	21.766	ug/l	98
5) Bromomethane	1.618	94	10008	23.364	ug/l	99
6) Chloroethane	1.691	64	12982	21.627	ug/l	99
7) Trichlorofluoromethane	1.892	101	31000	20.869	ug/l	97
8) Diethyl Ether	2.136	74	11616	21.110	ug/l	96
9) 1,1,2-Trichlorotrifluo...	2.337	101	20052	20.593	ug/l	98
10) Methyl Iodide	2.459	142	25609	20.361	ug/l	98
11) Tert butyl alcohol	2.977	59	19567	94.651	ug/l	95
12) 1,1-Dichloroethene	2.325	96	19740	20.233	ug/l	93
13) Acrolein	2.239	56	19283	100.936	ug/l	99
14) Allyl chloride	2.672	41	36160	20.924	ug/l	95
15) Acrylonitrile	3.069	53	63117	106.544	ug/l	100
16) Acetone	2.386	43	53192	105.834	ug/l	97
17) Carbon Disulfide	2.520	76	54104	19.590	ug/l	99
18) Methyl Acetate	2.709	43	28937	20.011	ug/l	99
19) Methyl tert-butyl Ether	3.117	73	70167	20.319	ug/l	99
20) Methylene Chloride	2.794	84	22577	20.731	ug/l	99
21) trans-1,2-Dichloroethene	3.099	96	21402	20.847	ug/l	95
22) Diisopropyl ether	3.764	45	72840	20.936	ug/l	100
23) Vinyl Acetate	3.727	43	309141	105.680	ug/l	98
24) 1,1-Dichloroethane	3.617	63	39243	20.844	ug/l	99
25) 2-Butanone	4.562	43	87540	108.105	ug/l	98
26) 2,2-Dichloropropane	4.483	77	26870	21.124	ug/l	97
27) cis-1,2-Dichloroethene	4.495	96	25138	21.326	ug/l	96
28) Bromochloromethane	4.904	49	16423	21.044	ug/l	98
29) Tetrahydrofuran	5.007	42	57314	104.111	ug/l	98
30) Chloroform	5.099	83	40353	20.630	ug/l	96
31) Cyclohexane	5.477	56	37212	20.466	ug/l	94
32) 1,1,1-Trichloroethane	5.391	97	35648	20.670	ug/l	99
36) 1,1-Dichloropropene	5.702	75	30419	20.436	ug/l	98
37) Ethyl Acetate	4.715	43	33209	19.632	ug/l	99
38) Carbon Tetrachloride	5.684	117	31291	20.047	ug/l	96
39) Methylcyclohexane	7.385	83	36771	20.144	ug/l	98
40) Benzene	6.044	78	88650	20.435	ug/l	98

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX020922\
 Data File : VX026889.D
 Acq On : 10 Feb 2022 00:15
 Operator : JC/MD
 Sample : VX0209WBSD02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 31 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0209WBSD02

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 02/10/2022
 Supervised By :Mahesh Dadoda 02/10/2022

Quant Time: Feb 10 01:13:41 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020822W.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 08 07:07:46 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	4.928	41	19261	21.008	ug/l	96
42) 1,2-Dichloroethane	6.092	62	32818	21.468	ug/l	99
43) Isopropyl Acetate	6.342	43	53439	20.532	ug/l	99
44) Trichloroethene	7.129	130	26102	20.039	ug/l	97
45) 1,2-Dichloropropane	7.434	63	22668	20.652	ug/l	98
46) Dibromomethane	7.586	93	15744	20.730	ug/l	98
47) Bromodichloromethane	7.824	83	31074	21.365	ug/l	99
48) Methyl methacrylate	7.696	41	26648	20.663	ug/l	97
49) 1,4-Dioxane	7.684	88	8963	403.322	ug/l	97
51) 4-Methyl-2-Pentanone	8.574	43	173214	109.822	ug/l	99
52) Toluene	8.720	92	57448	20.916	ug/l	99
53) t-1,3-Dichloropropene	8.982	75	30479	20.185	ug/l	99
54) cis-1,3-Dichloropropene	8.366	75	34376	21.404	ug/l	96
55) 1,1,2-Trichloroethane	9.153	97	22687	21.819	ug/l	99
56) Ethyl methacrylate	9.116	69	35060	21.271	ug/l	98
57) 1,3-Dichloropropane	9.311	76	38921	21.798	ug/l	99
58) 2-Chloroethyl Vinyl ether	8.244	63	84705	100.125	ug/l	99
59) 2-Hexanone	9.433	43	127061	110.139	ug/l	98
60) Dibromochloromethane	9.525	129	23293	21.771	ug/l	100
61) 1,2-Dibromoethane	9.610	107	24065	21.905	ug/l	98
64) Tetrachloroethene	9.275	164	27409	19.083	ug/l	97
65) Chlorobenzene	10.079	112	61122	20.666	ug/l	100
66) 1,1,1,2-Tetrachloroethane	10.165	131	21802	20.039	ug/l	99
67) Ethyl Benzene	10.195	91	109283	20.877	ug/l	100
68) m/p-Xylenes	10.305	106	83350	41.386	ug/l	97
69) o-Xylene	10.640	106	41038	20.795	ug/l	96
70) Styrene	10.659	104	67166	20.985	ug/l	98
71) Bromoform	10.799	173	15211	19.240	ug/l #	98
73) Isopropylbenzene	10.963	105	107644	21.395	ug/l	99
74) N-amyl acetate	10.842	43	41236	20.512	ug/l	96
75) 1,1,2,2-Tetrachloroethane	11.213	83	28132	20.604	ug/l	99
76) 1,2,3-Trichloropropane	11.238	75	29609m	21.175	ug/l	
77) Bromobenzene	11.201	156	24834	20.765	ug/l	97
78) n-propylbenzene	11.305	91	119404	21.407	ug/l	99
79) 2-Chlorotoluene	11.366	91	72436	21.018	ug/l	99
80) 1,3,5-Trimethylbenzene	11.451	105	88323	21.329	ug/l	98
81) trans-1,4-Dichloro-2-b...	11.018	75	7322	18.655	ug/l	97
82) 4-Chlorotoluene	11.457	91	82695	21.389	ug/l	98
83) tert-Butylbenzene	11.713	119	83258	21.095	ug/l	99
84) 1,2,4-Trimethylbenzene	11.750	105	88471	21.181	ug/l	100
85) sec-Butylbenzene	11.890	105	106180	21.224	ug/l	100
86) p-Isopropyltoluene	12.012	119	89533	21.125	ug/l	99
87) 1,3-Dichlorobenzene	11.969	146	44505	20.308	ug/l	99
88) 1,4-Dichlorobenzene	12.043	146	44546	20.156	ug/l	99
89) n-Butylbenzene	12.335	91	72855	21.010	ug/l	98
90) Hexachloroethane	12.542	117	12790	20.880	ug/l	97
91) 1,2-Dichlorobenzene	12.335	146	43413	20.531	ug/l	98
92) 1,2-Dibromo-3-Chloropr...	12.945	75	6387	19.022	ug/l	95
93) 1,2,4-Trichlorobenzene	13.591	180	25949	19.445	ug/l	98
94) Hexachlorobutadiene	13.725	225	11224	19.884	ug/l	98
95) Naphthalene	13.774	128	93194	19.937	ug/l	99
96) 1,2,3-Trichlorobenzene	13.963	180	26525	19.630	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX020922\
Data File : VX026889.D
Acq On : 10 Feb 2022 00:15
Operator : JC/MD
Sample : VX0209WBSD02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 31 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0209WBSD02

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 02/10/2022
Supervised By :Mahesh Dadoda 02/10/2022

Quant Time: Feb 10 01:13:41 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020822W.M
Quant Title : SW846 8260
QLast Update : Tue Feb 08 07:07:46 2022
Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Manual Integrations

Quant Time: Feb 10 01:13:41 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020822W.M
Quant Title : SW846 8260
QLast Update : Tue Feb 08 07:07:46 2022
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.556 min Scan# 71
Delta R.T. -0.006 min
Lab File: VX026889.D
Acq: 10 Feb 2022 00:15

Instrument :
MSVOA_X
ClientSampleId :
VX0209WBSD02

Tgt Ion:168 Resp: 95329
Ion Ratio Lower Upper
168 100
99 60.7 50.3 75.5




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 02/10/2022
Supervised By :Mahesh Dadoda 02/10/2022



#2
Dichlorodifluoromethane
Concen: 21.367 ug/l
RT: 1.173 min Scan# 14
Delta R.T. 0.000 min
Lab File: VX026889.D
Acq: 10 Feb 2022 00:15


Tgt Ion: 85 Resp: 22148
Ion Ratio Lower Upper
85 100
87 31.6 16.4 49.1





#3
Chloromethane
Concen: 22.046 ug/l
RT: 1.294 min Scan# 34
Delta R.T. 0.000 min
Lab File: VX026889.D
Acq: 10 Feb 2022 00:15




Instrument :
MSVOA_X
ClientSampleId :
VX0209WBSD02




Tgt Ion: 50 Resp: 21259
Ion Ratio Lower Upper
50 100
52 31.6 26.1 39.1

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 02/10/2022
Supervised By :Mahesh Dadoda 02/10/2022




#4
Vinyl Chloride
Concen: 21.766 ug/l
RT: 1.380 min Scan# 48
Delta R.T. 0.000 min
Lab File: VX026889.D
Acq: 10 Feb 2022 00:15



Tgt Ion: 62 Resp: 21971
Ion Ratio Lower Upper
62 100
64 33.7 26.2 39.2









#5
Bromomethane
Concen: 23.364 ug/l
RT: 1.618 min Scan# 81
Delta R.T. 0.006 min
Lab File: VX026889.D
Acq: 10 Feb 2022 00:15

Instrument :
MSVOA_X
ClientSampleId :
VX0209WBSD02

Tgt Ion: 94 Resp: 1000
Ion Ratio Lower Upper
94 100
96 94.2 75.9 113.9




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 02/10/2022
Supervised By :Mahesh Dadoda 02/10/2022



#6
Chloroethane
Concen: 21.627 ug/l
RT: 1.691 min Scan# 99
Delta R.T. -0.000 min
Lab File: VX026889.D
Acq: 10 Feb 2022 00:15


Tgt Ion: 64 Resp: 12982
Ion Ratio Lower Upper
64 100
66 32.1 25.4 38.2





#7
 Trichlorofluoromethane
 Concen: 20.869 ug/l
 RT: 1.892 min Scan# 11
 Delta R.T. 0.000 min
 Lab File: VX026889.D
 Acq: 10 Feb 2022 00:15




Instrument :
 MSVOA_X
 ClientSampleId :
 VX0209WBSD02



Tgt Ion:101 Resp: 31000
 Ion Ratio Lower Upper
 101 100
 103 66.2 51.3 76.9




Manual Integrations
 APPROVED


Reviewed By :Semsettin Yesilyurt 02/10/2022
 Supervised By :Mahesh Dadoda 02/10/2022



#8
 Diethyl Ether
 Concen: 21.110 ug/l
 RT: 2.136 min Scan# 172
 Delta R.T. 0.000 min
 Lab File: VX026889.D
 Acq: 10 Feb 2022 00:15

Tgt Ion: 74 Resp: 11616
 Ion Ratio Lower Upper
 74 100
 45 97.6 50.7 152.1









#9
1,1,2-Trichlorotrifluoroethane
Concen: 20.593 ug/l
RT: 2.337 min Scan# 205
Delta R.T. 0.000 min
Lab File: VX026889.D
Acq: 10 Feb 2022 00:15

Instrument :
MSVOA_X
ClientSampleId :
VX0209WBSD02

Tgt Ion:101 Resp: 2005
Ion Ratio Lower Upper
101 100
85 42.4 35.3 52.9
151 78.8 61.4 92.2




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 02/10/2022
Supervised By :Mahesh Dadoda 02/10/2022



#10
Methyl Iodide
Concen: 20.361 ug/l
RT: 2.459 min Scan# 225
Delta R.T. -0.000 min
Lab File: VX026889.D
Acq: 10 Feb 2022 00:15

Tgt Ion:142 Resp: 25609
Ion Ratio Lower Upper
142 100
127 46.0 38.1 57.1
141 14.2 11.2 16.8





#11

Tert butyl alcohol

Concen: 94.651 ug/l

RT: 2.977 min Scan# 31

Delta R.T. -0.024 min

Lab File: VX026889.D

Acq: 10 Feb 2022 00:15

Instrument :

MSVOA_X

ClientSampleId :

VX0209WBSD02

Tgt Ion: 59 Resp: 1956

Ion Ratio Lower Upper

59 100





57 10.8 7.2 10.8

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 02/10/2022

Supervised By :Mahesh Dadoda 02/10/2022



#12

1,1-Dichloroethene

Concen: 20.233 ug/l

RT: 2.325 min Scan# 203

Delta R.T. 0.000 min

Lab File: VX026889.D

Acq: 10 Feb 2022 00:15




Tgt Ion: 96 Resp: 19740


Ion Ratio Lower Upper

96 100

61 161.5 138.5 207.7


98 62.5 51.8 77.6





#13
Acrolein
Concen: 100.936 ug/l
RT: 2.239 min Scan# 11
Delta R.T. 0.000 min
Lab File: VX026889.D
Acq: 10 Feb 2022 00:15




Instrument :
MSVOA_X
ClientSampleId :
VX0209WBSD02




Tgt Ion: 56 Resp: 1928
Ion Ratio Lower Upper
56 100
55 70.4 56.9 85.3

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 02/10/2022
Supervised By :Mahesh Dadoda 02/10/2022




#14
Allyl chloride
Concen: 20.924 ug/l
RT: 2.672 min Scan# 260
Delta R.T. 0.006 min
Lab File: VX026889.D
Acq: 10 Feb 2022 00:15




Tgt Ion: 41 Resp: 36160
Ion Ratio Lower Upper
41 100
39 62.8 54.2 81.4
76 30.9 25.2 37.8





#15
Acrylonitrile
Concen: 106.544 ug/l
RT: 3.069 min Scan# 31
Delta R.T. 0.000 min
Lab File: VX026889.D
Acq: 10 Feb 2022 00:15




Instrument :
MSVOA_X
ClientSampleId :
VX0209WBSD02




Tgt Ion: 53 Resp: 6311
Ion Ratio Lower Upper
53 100
52 83.3 66.6 100.0
51 36.1 29.5 44.3

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 02/10/2022
Supervised By :Mahesh Dadoda 02/10/2022




#16
Acetone
Concen: 105.834 ug/l
RT: 2.386 min Scan# 213
Delta R.T. 0.000 min
Lab File: VX026889.D
Acq: 10 Feb 2022 00:15



Tgt Ion: 43 Resp: 53192
Ion Ratio Lower Upper
43 100
58 30.4 23.0 34.6









#17
Carbon Disulfide
Concen: 19.590 ug/l
RT: 2.520 min Scan# 211
Delta R.T. 0.006 min
Lab File: VX026889.D
Acq: 10 Feb 2022 00:15

Instrument :
MSVOA_X
ClientSampleId :
VX0209WBSD02

Tgt Ion: 76 Resp: 54104
Ion Ratio Lower Upper
76 100
78 9.2 7.5 11.3




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 02/10/2022
Supervised By :Mahesh Dadoda 02/10/2022



#18
Methyl Acetate
Concen: 20.011 ug/l
RT: 2.709 min Scan# 266
Delta R.T. 0.000 min
Lab File: VX026889.D
Acq: 10 Feb 2022 00:15

Tgt Ion: 43 Resp: 28937
Ion Ratio Lower Upper
43 100
74 23.6 18.3 27.5









#19
Methyl tert-butyl Ether
Concen: 20.319 ug/l
RT: 3.117 min Scan# 311
Delta R.T. 0.000 min
Lab File: VX026889.D
Acq: 10 Feb 2022 00:15

Instrument :
MSVOA_X
ClientSampleId :
VX0209WBSD02

Tgt Ion: 73 Resp: 7016
Ion Ratio Lower Upper
73 100
57 21.2 17.4 26.2




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 02/10/2022
Supervised By :Mahesh Dadoda 02/10/2022



#20
Methylene Chloride
Concen: 20.731 ug/l
RT: 2.794 min Scan# 280
Delta R.T. 0.000 min
Lab File: VX026889.D
Acq: 10 Feb 2022 00:15


Tgt Ion: 84 Resp: 22577
Ion Ratio Lower Upper
84 100
49 124.3 101.5 152.3
51 38.7 31.2 46.8
86 63.9 51.3 76.9





#21
trans-1,2-Dichloroethene
Concen: 20.847 ug/l
RT: 3.099 min Scan# 311
Delta R.T. 0.000 min
Lab File: VX026889.D
Acq: 10 Feb 2022 00:15




Instrument :
MSVOA_X
ClientSampleId :
VX0209WBSD02




Tgt Ion: 96 Resp: 2140
Ion Ratio Lower Upper
96 100
61 133.2 112.9 169.3
98 62.7 49.5 74.3

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 02/10/2022
Supervised By :Mahesh Dadoda 02/10/2022




#22
Diisopropyl ether
Concen: 20.936 ug/l
RT: 3.764 min Scan# 439
Delta R.T. 0.000 min
Lab File: VX026889.D
Acq: 10 Feb 2022 00:15




Tgt Ion: 45 Resp: 72840
Ion Ratio Lower Upper
45 100
43 75.5 60.6 91.0
87 27.6 22.0 33.0
59 11.7 9.0 13.4





#23
 Vinyl Acetate
 Concen: 105.680 ug/l
 RT: 3.727 min Scan# 41
 Delta R.T. 0.000 min
 Lab File: VX026889.D
 Acq: 10 Feb 2022 00:15




Instrument :
 MSVOA_X
 ClientSampleId :
 VX0209WBSD02




Tgt Ion: 43 Resp: 30914
 Ion Ratio Lower Upper
 43 100
 86 10.9 8.1 12.1

Manual Integrations
 APPROVED



Reviewed By :Semsettin Yesilyurt 02/10/2022
 Supervised By :Mahesh Dadoda 02/10/2022




#24
 1,1-Dichloroethane
 Concen: 20.844 ug/l
 RT: 3.617 min Scan# 415
 Delta R.T. 0.000 min
 Lab File: VX026889.D
 Acq: 10 Feb 2022 00:15




Tgt Ion: 63 Resp: 39243
 Ion Ratio Lower Upper
 63 100
 98 7.7 3.7 11.1
 100 4.7 2.4 7.2





#25
2-Butanone
Concen: 108.105 ug/l
RT: 4.562 min Scan# 51
Delta R.T. 0.000 min
Lab File: VX026889.D
Acq: 10 Feb 2022 00:15




Instrument :
MSVOA_X
ClientSampleId :
VX0209WBSD02




Tgt Ion: 43 Resp: 87540
Ion Ratio Lower Upper
43 100
72 24.5 20.4 30.6

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 02/10/2022
Supervised By :Mahesh Dadoda 02/10/2022




#26
2,2-Dichloropropane
Concen: 21.124 ug/l
RT: 4.483 min Scan# 557
Delta R.T. 0.000 min
Lab File: VX026889.D
Acq: 10 Feb 2022 00:15



Tgt Ion: 77 Resp: 26870
Ion Ratio Lower Upper
77 100
97 23.2 11.0 33.0





#27

cis-1,2-Dichloroethene

Concen: 21.326 ug/l

RT: 4.495 min Scan# 559

Delta R.T. 0.000 min

Lab File: VX026889.D

Acq: 10 Feb 2022 00:15

Instrument :

MSVOA_X

ClientSampleId :

VX0209WBSD02

Tgt Ion: 96 Resp: 25138

Ion Ratio Lower Upper

96 100

61 141.3 0.0 296.6





98 64.8 0.0 128.8

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 02/10/2022

Supervised By :Mahesh Dadoda 02/10/2022



#28

Bromochloromethane

Concen: 21.044 ug/l

RT: 4.904 min Scan# 626

Delta R.T. 0.000 min

Lab File: VX026889.D

Acq: 10 Feb 2022 00:15




Tgt Ion: 49 Resp: 16423


Ion Ratio Lower Upper

49 100

129 1.5 0.0 4.2

130 79.7 62.4 93.6





#29

Tetrahydrofuran

Concen: 104.111 ug/l

RT: 5.007 min Scan# 643

Delta R.T. 0.000 min

Lab File: VX026889.D

Acq: 10 Feb 2022 00:15

Instrument :

MSVOA_X

ClientSampleId :





VX0209WBSD02

Tgt Ion:	42	Resp:	57314
Ion	Ratio	Lower	Upper
42	100		
72	44.6	34.6	52.0
71	41.5	31.8	47.8

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 02/10/2022
Supervised By :Mahesh Dadoda 02/10/2022



#30

Chloroform

Concen: 20.630 ug/l




RT: 5.099 min Scan# 658


Delta R.T. 0.000 min

Lab File: VX026889.D

Acq: 10 Feb 2022 00:15


Tgt Ion:	83	Resp:	40353
Ion	Ratio	Lower	Upper
83	100		
85	67.4	51.6	77.4





#31
Cyclohexane
Concen: 20.466 ug/l
RT: 5.477 min Scan# 720
Delta R.T. 0.000 min
Lab File: VX026889.D
Acq: 10 Feb 2022 00:15




Instrument :
MSVOA_X
ClientSampleId :
VX0209WBSD02




Tgt Ion: 56 Resp: 3721
Ion Ratio Lower Upper
56 100
69 30.3 25.8 38.6
84 92.8 68.9 103.3

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 02/10/2022
Supervised By :Mahesh Dadoda 02/10/2022




#32
1,1,1-Trichloroethane
Concen: 20.670 ug/l
RT: 5.391 min Scan# 706
Delta R.T. 0.000 min
Lab File: VX026889.D
Acq: 10 Feb 2022 00:15



Tgt Ion: 97 Resp: 35648
Ion Ratio Lower Upper
97 100
99 65.0 51.3 76.9
61 43.8 35.0 52.6





#33

1,2-Dichloroethane-d4

Concen: 54.546 ug/l

RT: 5.964 min Scan# 800

Delta R.T. 0.000 min

Lab File: VX026889.D

Acq: 10 Feb 2022 00:15

Instrument :

MSVOA_X

ClientSampleId :

VX0209WBSD02

Tgt Ion: 65 Resp: 62418

Ion Ratio Lower Upper

65 100





67 51.7 0.0 100.8

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 02/10/2022

Supervised By :Mahesh Dadoda 02/10/2022



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 6.769 min Scan# 932

Delta R.T. 0.000 min

Lab File: VX026889.D

Acq: 10 Feb 2022 00:15




Tgt Ion:114 Resp: 158370


Ion Ratio Lower Upper

114 100

63 20.2 0.0 42.8

88 15.2 0.0 33.0





#35

Dibromofluoromethane

Concen: 51.816 ug/l

RT: 5.391 min Scan# 706

Delta R.T. 0.000 min

Lab File: VX026889.D


Acq: 10 Feb 2022 00:15

Instrument :

MSVOA_X

ClientSampleId :

VX0209WBSD02






Tgt Ion:	113	Resp:	48890
Ion	Ratio	Lower	Upper
113	100		
111	103.7	81.2	121.8
192	19.6	15.0	22.4

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 02/10/2022
Supervised By :Mahesh Dadoda 02/10/2022



#36

1,1-Dichloropropene

Concen: 20.436 ug/l




RT: 5.702 min Scan# 757


Delta R.T. 0.006 min

Lab File: VX026889.D

Acq: 10 Feb 2022 00:15


Tgt Ion:	75	Resp:	30419
Ion	Ratio	Lower	Upper
75	100		
110	36.1	17.2	51.6
77	30.5	24.6	36.8





#37
Ethyl Acetate
Concen: 19.632 ug/l
RT: 4.715 min Scan# 595
Delta R.T. 0.000 min
Lab File: VX026889.D
Acq: 10 Feb 2022 00:15




Instrument :
MSVOA_X
ClientSampleId :
VX0209WBSD02



Tgt Ion: 43 Resp: 33209
Ion Ratio Lower Upper
43 100
61 13.8 10.6 16.0
70 11.2 8.5 12.7




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 02/10/2022
Supervised By :Mahesh Dadoda 02/10/2022



#38
Carbon Tetrachloride
Concen: 20.047 ug/l
RT: 5.684 min Scan# 754
Delta R.T. 0.000 min
Lab File: VX026889.D
Acq: 10 Feb 2022 00:15

Tgt Ion:117 Resp: 31291
Ion Ratio Lower Upper
117 100
119 96.7 74.7 112.1
121 29.1 25.4 38.0









#39
Methylcyclohexane
Concen: 20.144 ug/l
RT: 7.385 min Scan# 1033
Delta R.T. 0.000 min
Lab File: VX026889.D
Acq: 10 Feb 2022 00:15

Instrument :
MSVOA_X
ClientSampleId :
VX0209WBSD02

Tgt Ion: 83 Resp: 3677
Ion Ratio Lower Upper
83 100
55 81.4 63.6 95.4
98 49.2 38.4 57.6




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 02/10/2022
Supervised By :Mahesh Dadoda 02/10/2022



#40
Benzene
Concen: 20.435 ug/l
RT: 6.044 min Scan# 813
Delta R.T. 0.000 min
Lab File: VX026889.D
Acq: 10 Feb 2022 00:15


Tgt Ion: 78 Resp: 88650
Ion Ratio Lower Upper
78 100
77 24.6 18.9 28.3





#41
Methacrylonitrile
Concen: 21.008 ug/l
RT: 4.928 min Scan# 61
Delta R.T. 0.006 min
Lab File: VX026889.D
Acq: 10 Feb 2022 00:15




Instrument :
MSVOA_X
ClientSampleId :
VX0209WBSD02



Tgt Ion: 41 Resp: 1926
Ion Ratio Lower Upper
41 100
39 52.4 45.4 68.0
67 68.7 53.2 79.8
52 27.6 23.0 34.6




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 02/10/2022
Supervised By :Mahesh Dadoda 02/10/2022



#42
1,2-Dichloroethane
Concen: 21.468 ug/l
RT: 6.092 min Scan# 821
Delta R.T. 0.000 min
Lab File: VX026889.D
Acq: 10 Feb 2022 00:15

Tgt Ion: 62 Resp: 32818
Ion Ratio Lower Upper
62 100
98 9.3 0.0 17.8





#43

Isopropyl Acetate

Concen: 20.532 ug/l

RT: 6.342 min Scan# 862

Delta R.T. 0.000 min

Lab File: VX026889.D


Acq: 10 Feb 2022 00:15

Instrument :

MSVOA_X

ClientSampleId :

VX0209WBSD02






Tgt Ion: 43	Resp: 53439
Ion Ratio	Lower Upper
43	100
61	19.4 15.4 23.0
87	13.1 10.2 15.2

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 02/10/2022
Supervised By :Mahesh Dadoda 02/10/2022



#44

Trichloroethene

Concen: 20.039 ug/l




RT: 7.129 min Scan# 991

Delta R.T. 0.000 min

Lab File: VX026889.D

Acq: 10 Feb 2022 00:15

Tgt Ion:130	Resp: 26102
Ion Ratio	Lower Upper
130	100
95	98.3 0.0 202.6





#45

1,2-Dichloropropane

Concen: 20.652 ug/l

RT: 7.434 min Scan# 1041

Delta R.T. 0.000 min

Lab File: VX026889.D

Acq: 10 Feb 2022 00:15

Instrument :

MSVOA_X

ClientSampleId :

VX0209WBSD02

Tgt Ion: 63 Resp: 2266

Ion Ratio Lower Upper

63 100





65 30.8 23.8 35.6

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 02/10/2022

Supervised By :Mahesh Dadoda 02/10/2022



#46

Dibromomethane

Concen: 20.730 ug/l

RT: 7.586 min Scan# 1066

Delta R.T. 0.000 min

Lab File: VX026889.D

Acq: 10 Feb 2022 00:15




Tgt Ion: 93 Resp: 15744


Ion Ratio Lower Upper

93 100

95 83.2 66.1 99.1


174 101.3 78.7 118.1





#47
Bromodichloromethane
Concen: 21.365 ug/l
RT: 7.824 min Scan# 1105
Delta R.T. 0.000 min
Lab File: VX026889.D
Acq: 10 Feb 2022 00:15




Instrument :
MSVOA_X
ClientSampleId :
VX0209WBSD02




Tgt Ion: 83 Resp: 31074
Ion Ratio Lower Upper
83 100
85 63.7 51.4 77.0
127 9.4 6.9 10.3

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 02/10/2022
Supervised By :Mahesh Dadoda 02/10/2022




#48
Methyl methacrylate
Concen: 20.663 ug/l
RT: 7.696 min Scan# 1084
Delta R.T. 0.000 min
Lab File: VX026889.D
Acq: 10 Feb 2022 00:15



Tgt Ion: 41 Resp: 26648
Ion Ratio Lower Upper
41 100
69 81.2 62.2 93.4
39 55.2 45.4 68.0





#49

1,4-Dioxane

Concen: 403.322 ug/l

RT: 7.684 min Scan# 1081

Delta R.T. 0.006 min

Lab File: VX026889.D

Acq: 10 Feb 2022 00:15

Instrument :

MSVOA_X

ClientSampleId :





VX0209WBSD02

Tgt Ion: 88 Resp: 896
 Ion Ratio Lower Upper
 88 100
 43 34.2 29.9 44.9
 58 71.3 57.8 86.6

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 02/10/2022
 Supervised By :Mahesh Dadoda 02/10/2022



#50

Toluene-d8

Concen: 54.488 ug/l




RT: 8.653 min Scan# 1241


Delta R.T. 0.000 min

Lab File: VX026889.D

Acq: 10 Feb 2022 00:15


Tgt Ion: 98 Resp: 188717
 Ion Ratio Lower Upper
 98 100
 100 66.5 53.2 79.8





#51
4-Methyl-2-Pentanone
Concen: 109.822 ug/l
RT: 8.574 min Scan# 1228
Delta R.T. 0.000 min
Lab File: VX026889.D
Acq: 10 Feb 2022 00:15




Instrument :
MSVOA_X
ClientSampleId :
VX0209WBSD02




Tgt Ion: 43 Resp: 173214
Ion Ratio Lower Upper
43 100
58 38.8 30.5 45.7

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 02/10/2022
Supervised By :Mahesh Dadoda 02/10/2022



#52
Toluene
Concen: 20.916 ug/l
RT: 8.720 min Scan# 1252
Delta R.T. 0.000 min
Lab File: VX026889.D
Acq: 10 Feb 2022 00:15



Tgt Ion: 92 Resp: 57448
Ion Ratio Lower Upper
92 100
91 167.7 135.7 203.5





#53

t-1,3-Dichloropropene

Concen: 20.185 ug/l

RT: 8.982 min Scan# 1194

Delta R.T. 0.000 min

Lab File: VX026889.D

Acq: 10 Feb 2022 00:15

Instrument :

MSVOA_X

ClientSampleId :

VX0209WBSD02

Tgt Ion: 75 Resp: 30479

Ion Ratio Lower Upper

75 100





77 31.4 25.7 38.5

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 02/10/2022

Supervised By :Mahesh Dadoda 02/10/2022



#54

cis-1,3-Dichloropropene

Concen: 21.404 ug/l

RT: 8.366 min Scan# 1194

Delta R.T. 0.000 min

Lab File: VX026889.D

Acq: 10 Feb 2022 00:15




Tgt Ion: 75 Resp: 34376


Ion Ratio Lower Upper

75 100

77 31.6 26.2 39.2


39 50.6 43.5 65.3





#55
1,1,2-Trichloroethane
Concen: 21.819 ug/l
RT: 9.153 min Scan# 1323
Delta R.T. 0.000 min
Lab File: VX026889.D
Acq: 10 Feb 2022 00:15




Instrument :
MSVOA_X
ClientSampleId :
VX0209WBSD02




Tgt Ion: 97 Resp: 2268
Ion Ratio Lower Upper
97 100
83 87.3 68.4 102.6
85 54.4 44.0 66.0
99 63.2 49.8 74.6

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 02/10/2022
Supervised By :Mahesh Dadoda 02/10/2022




#56
Ethyl methacrylate
Concen: 21.271 ug/l
RT: 9.116 min Scan# 1317
Delta R.T. 0.000 min
Lab File: VX026889.D
Acq: 10 Feb 2022 00:15



Tgt Ion: 69 Resp: 35060
Ion Ratio Lower Upper
69 100
41 69.6 57.2 85.8
39 37.8 30.9 46.3





#57

1,3-Dichloropropane

Concen: 21.798 ug/l

RT: 9.311 min Scan# 11

Delta R.T. 0.000 min

Lab File: VX026889.D

Acq: 10 Feb 2022 00:15

Instrument :

MSVOA_X

ClientSampleId :

VX0209WBSD02

Tgt Ion: 76 Resp: 3892

Ion Ratio Lower Upper

76 100





78 31.8 25.7 38.5

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 02/10/2022

Supervised By :Mahesh Dadoda 02/10/2022



#58

2-Chloroethyl Vinyl ether

Concen: 100.125 ug/l

RT: 8.244 min Scan# 1174

Delta R.T. 0.000 min

Lab File: VX026889.D




Acq: 10 Feb 2022 00:15


Tgt Ion: 63 Resp: 84705

Ion Ratio Lower Upper

63 100

106 29.6 23.0 34.6





#59

2-Hexanone

Concen: 110.139 ug/l

RT: 9.433 min Scan# 1369

Delta R.T. 0.000 min

Lab File: VX026889.D

Acq: 10 Feb 2022 00:15

Instrument :

MSVOA_X

ClientSampleId :

VX0209WBSD02

Tgt Ion: 43 Resp: 12706

Ion Ratio Lower Upper

43 100





58 53.9 26.4 79.2

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 02/10/2022

Supervised By :Mahesh Dadoda 02/10/2022



#60

Dibromochloromethane

Concen: 21.771 ug/l

RT: 9.525 min Scan# 1384

Delta R.T. 0.000 min

Lab File: VX026889.D




Acq: 10 Feb 2022 00:15


Tgt Ion:129 Resp: 23293

Ion Ratio Lower Upper

129 100


127 77.1 38.5 115.5





#61
1,2-Dibromoethane
Concen: 21.905 ug/l
RT: 9.610 min Scan# 1398
Delta R.T. 0.000 min
Lab File: VX026889.D
Acq: 10 Feb 2022 00:15




Instrument :
MSVOA_X
ClientSampleId :
VX0209WBSD02




Tgt Ion:107 Resp: 2406
Ion Ratio Lower Upper
107 100
109 92.4 75.5 113.3

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 02/10/2022
Supervised By :Mahesh Dadoda 02/10/2022




#62
4-Bromofluorobenzene
Concen: 55.922 ug/l
RT: 11.079 min Scan# 1639
Delta R.T. 0.000 min
Lab File: VX026889.D
Acq: 10 Feb 2022 00:15




Tgt Ion: 95 Resp: 68190
Ion Ratio Lower Upper
95 100
174 76.2 0.0 148.6
176 72.9 0.0 141.4





#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.055 min Scan# 1471
Delta R.T. 0.000 min
Lab File: VX026889.D
Acq: 10 Feb 2022 00:15




Instrument :
MSVOA_X
ClientSampleId :
VX0209WBSD02



Tgt Ion:117 Resp: 143009
Ion Ratio Lower Upper
117 100
82 54.6 45.2 67.8
119 31.6 26.2 39.2




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 02/10/2022
Supervised By :Mahesh Dadoda 02/10/2022



#64
Tetrachloroethene
Concen: 19.083 ug/l
RT: 9.275 min Scan# 1343
Delta R.T. 0.000 min
Lab File: VX026889.D
Acq: 10 Feb 2022 00:15

Tgt Ion:164 Resp: 27409
Ion Ratio Lower Upper
164 100
166 125.2 101.8 152.6
129 98.1 83.3 124.9
131 94.4 78.0 117.0









#65
Chlorobenzene
Concen: 20.666 ug/l
RT: 10.079 min Scan# 1475
Delta R.T. 0.000 min
Lab File: VX026889.D
Acq: 10 Feb 2022 00:15

Instrument :
MSVOA_X
ClientSampleId :
VX0209WBSD02

Tgt Ion:112 Resp: 6112
Ion Ratio Lower Upper
112 100
114 31.4 25.3 37.9




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 02/10/2022
Supervised By :Mahesh Dadoda 02/10/2022



#66
1,1,1,2-Tetrachloroethane
Concen: 20.039 ug/l
RT: 10.165 min Scan# 1489
Delta R.T. 0.000 min
Lab File: VX026889.D
Acq: 10 Feb 2022 00:15


Tgt Ion:131 Resp: 21802
Ion Ratio Lower Upper
131 100
133 95.2 47.9 143.8
119 63.9 31.7 95.1





#67
Ethyl Benzene
Concen: 20.877 ug/l
RT: 10.195 min Scan# 1494
Delta R.T. 0.000 min
Lab File: VX026889.D
Acq: 10 Feb 2022 00:15




Instrument :
MSVOA_X
ClientSampleId :
VX0209WBSD02




Tgt Ion: 91 Resp: 10928
Ion Ratio Lower Upper
91 100
106 30.7 24.7 37.1

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 02/10/2022
Supervised By :Mahesh Dadoda 02/10/2022




#68
m/p-Xylenes
Concen: 41.386 ug/l
RT: 10.305 min Scan# 1512
Delta R.T. 0.000 min
Lab File: VX026889.D
Acq: 10 Feb 2022 00:15



Tgt Ion:106 Resp: 83350
Ion Ratio Lower Upper
106 100
91 202.5 165.3 247.9









#69
o-Xylene
Concen: 20.795 ug/l
RT: 10.640 min Scan# 11
Delta R.T. 0.000 min
Lab File: VX026889.D
Acq: 10 Feb 2022 00:15

Instrument :
MSVOA_X
ClientSampleId :
VX0209WBSD02

Tgt Ion:106 Resp: 41038
Ion Ratio Lower Upper
106 100
91 213.7 109.7 329.0




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 02/10/2022
Supervised By :Mahesh Dadoda 02/10/2022



#70
Styrene
Concen: 20.985 ug/l
RT: 10.659 min Scan# 1570
Delta R.T. 0.000 min
Lab File: VX026889.D
Acq: 10 Feb 2022 00:15

Tgt Ion:104 Resp: 67166
Ion Ratio Lower Upper
104 100
78 50.1 42.5 63.7
103 54.1 43.3 64.9





#71

Bromoform

Concen: 19.240 ug/l

RT: 10.799 min Scan# 1593

Delta R.T. -0.000 min

Lab File: VX026889.D

Acq: 10 Feb 2022 00:15

Instrument :

MSVOA_X

ClientSampleId :





VX0209WBSD02

Tgt Ion:173	Resp: 1521
Ion Ratio	Lower Upper
173	100
175	47.6 24.3 72.9
254	0.0 0.0 0.0

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 02/10/2022
Supervised By :Mahesh Dadoda 02/10/2022



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l




RT: 12.024 min Scan# 1794


Delta R.T. 0.000 min

Lab File: VX026889.D

Acq: 10 Feb 2022 00:15

Tgt Ion:152	Resp: 63876
Ion Ratio	Lower Upper
152	100
115	70.3 44.2 132.6
150	161.8 0.0 348.2









#73
Isopropylbenzene
Concen: 21.395 ug/l
RT: 10.963 min Scan# 1620
Delta R.T. 0.000 min
Lab File: VX026889.D
Acq: 10 Feb 2022 00:15

Instrument :
MSVOA_X
ClientSampleId :
VX0209WBSD02

Tgt Ion:105 Resp: 107644
Ion Ratio Lower Upper
105 100
120 26.5 13.1 39.1




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 02/10/2022
Supervised By :Mahesh Dadoda 02/10/2022



#74
N-ethyl acetate
Concen: 20.512 ug/l
RT: 10.842 min Scan# 1600
Delta R.T. -0.006 min
Lab File: VX026889.D
Acq: 10 Feb 2022 00:15

Tgt Ion: 43 Resp: 41236
Ion Ratio Lower Upper
43 100
70 43.9 32.9 49.3
55 25.1 18.3 27.5
61 24.3 18.9 28.3









#75
1,1,2,2-Tetrachloroethane
Concen: 20.604 ug/l
RT: 11.213 min Scan# 1661
Delta R.T. 0.000 min
Lab File: VX026889.D
Acq: 10 Feb 2022 00:15

Instrument :
MSVOA_X
ClientSampleId :
VX0209WBSD02

Tgt Ion: 83 Resp: 2813
Ion Ratio Lower Upper
83 100
131 10.2 5.3 15.9
85 63.6 32.0 96.2




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 02/10/2022
Supervised By :Mahesh Dadoda 02/10/2022



#76
1,2,3-Trichloropropane
Concen: 21.175 ug/l m
RT: 11.238 min Scan# 1665
Delta R.T. -0.006 min
Lab File: VX026889.D
Acq: 10 Feb 2022 00:15

Tgt Ion: 75 Resp: 29609
Ion Ratio Lower Upper
75 100
77 37.7 18.3 54.8





#77

Bromobenzene

Concen: 20.765 ug/l

RT: 11.201 min Scan# 1659

Delta R.T. 0.000 min

Lab File: VX026889.D


Acq: 10 Feb 2022 00:15

Instrument :

MSVOA_X

ClientSampleId :

VX0209WBSD02





Tgt Ion:	156	Resp:	24834
Ion Ratio	100	Lower	Upper
156	100		
77	163.0	84.5	253.4
158	96.4	48.1	144.4

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 02/10/2022
Supervised By :Mahesh Dadoda 02/10/2022



#78

n-propylbenzene

Concen: 21.407 ug/l




RT: 11.305 min Scan# 1676


Delta R.T. 0.000 min

Lab File: VX026889.D

Acq: 10 Feb 2022 00:15

Tgt Ion:	91	Resp:	119404
Ion Ratio	100	Lower	Upper
91	100		
120	23.4	11.6	34.7









#79
2-Chlorotoluene
Concen: 21.018 ug/l
RT: 11.366 min Scan# 1686
Delta R.T. 0.000 min
Lab File: VX026889.D
Acq: 10 Feb 2022 00:15

Instrument :
MSVOA_X
ClientSampleId :
VX0209WBSD02

Tgt Ion: 91 Resp: 72436
Ion Ratio Lower Upper
91 100
126 33.3 16.4 49.2




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 02/10/2022
Supervised By :Mahesh Dadoda 02/10/2022



#80
1,3,5-Trimethylbenzene
Concen: 21.329 ug/l
RT: 11.451 min Scan# 1700
Delta R.T. 0.000 min
Lab File: VX026889.D
Acq: 10 Feb 2022 00:15


Tgt Ion:105 Resp: 88323
Ion Ratio Lower Upper
105 100
120 49.1 24.0 72.0





#81
trans-1,4-Dichloro-2-butene
Concen: 18.655 ug/l
RT: 11.018 min Scan# 1629
Delta R.T. 0.000 min
Lab File: VX026889.D
Acq: 10 Feb 2022 00:15




Instrument :
MSVOA_X
ClientSampleId :
VX0209WBSD02




Tgt Ion: 75 Resp: 732
Ion Ratio Lower Upper
75 100
53 121.4 94.0 141.0
89 47.9 37.0 55.4

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 02/10/2022
Supervised By :Mahesh Dadoda 02/10/2022




#82
4-Chlorotoluene
Concen: 21.389 ug/l
RT: 11.457 min Scan# 1701
Delta R.T. 0.000 min
Lab File: VX026889.D
Acq: 10 Feb 2022 00:15




Tgt Ion: 91 Resp: 82695
Ion Ratio Lower Upper
91 100
126 29.2 14.1 42.4





#83
tert-Butylbenzene
Concen: 21.095 ug/l
RT: 11.713 min Scan# 1743
Delta R.T. 0.000 min
Lab File: VX026889.D
Acq: 10 Feb 2022 00:15




Instrument :
MSVOA_X
ClientSampleId :
VX0209WBSD02




Tgt Ion:119 Resp: 83258
Ion Ratio Lower Upper
119 100
91 62.0 31.4 94.0
134 23.8 11.8 35.3

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 02/10/2022
Supervised By :Mahesh Dadoda 02/10/2022




#84
1,2,4-Trimethylbenzene
Concen: 21.181 ug/l
RT: 11.750 min Scan# 1749
Delta R.T. 0.000 min
Lab File: VX026889.D
Acq: 10 Feb 2022 00:15




Tgt Ion:105 Resp: 88471
Ion Ratio Lower Upper
105 100
120 43.8 21.9 65.5





#85
 sec-Butylbenzene
 Concen: 21.224 ug/l
 RT: 11.890 min Scan# 1772
 Delta R.T. 0.000 min
 Lab File: VX026889.D
 Acq: 10 Feb 2022 00:15




Instrument :
 MSVOA_X
 ClientSampleId :
 VX0209WBSD02




Tgt Ion:105 Resp: 106180
 Ion Ratio Lower Upper
 105 100
 134 20.1 10.1 30.1

Manual Integrations
 APPROVED



Reviewed By :Semsettin Yesilyurt 02/10/2022
 Supervised By :Mahesh Dadoda 02/10/2022




#86
 p-Isopropyltoluene
 Concen: 21.125 ug/l
 RT: 12.012 min Scan# 1792
 Delta R.T. 0.000 min
 Lab File: VX026889.D
 Acq: 10 Feb 2022 00:15




Tgt Ion:119 Resp: 89533
 Ion Ratio Lower Upper
 119 100
 134 25.8 13.0 39.0
 91 24.0 12.6 37.6





#87
1,3-Dichlorobenzene
Concen: 20.308 ug/l
RT: 11.969 min Scan# 1785
Delta R.T. 0.000 min
Lab File: VX026889.D
Acq: 10 Feb 2022 00:15




Instrument :
MSVOA_X
ClientSampleId :
VX0209WBSD02




Tgt Ion:146 Resp: 4450
Ion Ratio Lower Upper
146 100
111 41.8 21.3 63.7
148 64.5 31.6 95.0

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 02/10/2022
Supervised By :Mahesh Dadoda 02/10/2022




#88
1,4-Dichlorobenzene
Concen: 20.156 ug/l
RT: 12.043 min Scan# 1797
Delta R.T. 0.000 min
Lab File: VX026889.D
Acq: 10 Feb 2022 00:15




Tgt Ion:146 Resp: 44546
Ion Ratio Lower Upper
146 100
111 41.2 20.2 60.6
148 63.7 31.4 94.0





#89
n-Butylbenzene
Concen: 21.010 ug/l
RT: 12.335 min Scan# 1845
Delta R.T. 0.000 min
Lab File: VX026889.D
Acq: 10 Feb 2022 00:15




Instrument :
MSVOA_X
ClientSampleId :
VX0209WBSD02



Tgt Ion: 91 Resp: 7285
Ion Ratio Lower Upper
91 100
92 53.1 27.2 81.6
134 26.7 13.1 39.1




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 02/10/2022
Supervised By :Mahesh Dadoda 02/10/2022



#90
Hexachloroethane
Concen: 20.880 ug/l
RT: 12.542 min Scan# 1879
Delta R.T. 0.000 min
Lab File: VX026889.D
Acq: 10 Feb 2022 00:15


Tgt Ion:117 Resp: 12790
Ion Ratio Lower Upper
117 100
201 74.0 35.6 106.8





#91
1,2-Dichlorobenzene
Concen: 20.531 ug/l
RT: 12.335 min Scan# 1845
Delta R.T. 0.000 min
Lab File: VX026889.D
Acq: 10 Feb 2022 00:15




Instrument :
MSVOA_X
ClientSampleId :
VX0209WBSD02




Tgt Ion:146 Resp: 4341
Ion Ratio Lower Upper
146 100
111 42.1 21.8 65.3
148 62.5 32.0 96.2

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 02/10/2022
Supervised By :Mahesh Dadoda 02/10/2022




#92
1,2-Dibromo-3-Chloropropane
Concen: 19.022 ug/l
RT: 12.945 min Scan# 1945
Delta R.T. 0.000 min
Lab File: VX026889.D
Acq: 10 Feb 2022 00:15




Tgt Ion: 75 Resp: 6387
Ion Ratio Lower Upper
75 100
155 86.5 40.0 119.9
157 106.7 51.6 154.8





#93
1,2,4-Trichlorobenzene
Concen: 19.445 ug/l
RT: 13.591 min Scan# 2051
Delta R.T. 0.000 min
Lab File: VX026889.D
Acq: 10 Feb 2022 00:15




Instrument :
MSVOA_X
ClientSampleId :
VX0209WBSD02




Tgt Ion:180 Resp: 25949
Ion Ratio Lower Upper
180 100
182 95.2 46.7 140.0
145 32.1 16.3 48.8

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 02/10/2022
Supervised By :Mahesh Dadoda 02/10/2022




#94
Hexachlorobutadiene
Concen: 19.884 ug/l
RT: 13.725 min Scan# 2073
Delta R.T. 0.000 min
Lab File: VX026889.D
Acq: 10 Feb 2022 00:15




Tgt Ion:225 Resp: 11224
Ion Ratio Lower Upper
225 100
223 62.2 32.0 96.0
227 62.6 32.3 96.9





#95
Naphthalene
Concen: 19.937 ug/l
RT: 13.774 min Scan# 2081
Delta R.T. -0.006 min
Lab File: VX026889.D
Acq: 10 Feb 2022 00:15




Instrument :
MSVOA_X
ClientSampleId :
VX0209WBSD02




Tgt Ion:128 Resp: 93194
Ion Ratio Lower Upper
128 100
127 12.6 10.3 15.5
129 10.7 8.7 13.1

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 02/10/2022
Supervised By :Mahesh Dadoda 02/10/2022



#96
1,2,3-Trichlorobenzene
Concen: 19.630 ug/l
RT: 13.963 min Scan# 2112
Delta R.T. 0.000 min
Lab File: VX026889.D
Acq: 10 Feb 2022 00:15



Tgt Ion:180 Resp: 26525
Ion Ratio Lower Upper
180 100
182 94.4 47.2 141.6
145 33.4 17.5 52.5

