

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX020923\
 Data File : VX034067.D
 Acq On : 07 Feb 2023 17:38
 Operator : JC/MD
 Sample : 01416-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 EB-2023-2-2

Quant Time: Feb 08 00:05:45 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020423W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 03 23:25:45 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	130214	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	180166	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	162387	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	79935	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	63274	50.404	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 100.800%	
35) Dibromofluoromethane	5.385	113	56868	49.596	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 99.200%	
50) Toluene-d8	8.647	98	207094	50.129	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 100.260%	
62) 4-Bromofluorobenzene	11.079	95	73886	50.076	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 100.160%	
Target Compounds						Qvalue
16) Acetone	2.373	43	14034	26.469	ug/l	99
20) Methylene Chloride	2.788	84	1572	1.296	ug/l #	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX020923\
Data File : VX034067.D
Acq On : 07 Feb 2023 17:38
Operator : JC/MD
Sample : 01416-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
EB-2023-2-2

Quant Time: Feb 08 00:05:45 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020423W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 03 23:25:45 2023
Response via : Initial Calibration

