

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX020923\
 Data File : VX034069.D
 Acq On : 07 Feb 2023 18:24
 Operator : JC/MD
 Sample : 01467-01 50X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 33906

Quant Time: Feb 08 00:06:35 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020423W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 03 23:25:45 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	131882	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	182272	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	163560	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	82007	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	63790	50.173	ug/l	0.00
Spiked Amount	50.000	Range 78 - 117	Recovery	= 100.340%		
35) Dibromofluoromethane	5.385	113	57026	49.159	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 98.320%		
50) Toluene-d8	8.647	98	208209	49.816	ug/l	0.00
Spiked Amount	50.000	Range 92 - 112	Recovery	= 99.640%		
62) 4-Bromofluorobenzene	11.079	95	74357	49.813	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	= 99.620%		
Target Compounds						
					Qvalue	
52) Toluene	8.720	92	5206	1.746	ug/l	100
67) Ethyl Benzene	10.195	91	2610	0.456	ug/l	92
68) m/p-Xylenes	10.305	106	3493	1.528	ug/l	93
69) o-Xylene	10.640	106	2403	1.095	ug/l	97
78) n-propylbenzene	11.305	91	1242	0.205	ug/l	95
80) 1,3,5-Trimethylbenzene	11.451	105	1908	0.418	ug/l	96
84) 1,2,4-Trimethylbenzene	11.756	105	6089	1.336	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX020923\
Data File : VX034069.D
Acq On : 07 Feb 2023 18:24
Operator : JC/MD
Sample : 01467-01 50X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
33906

Quant Time: Feb 08 00:06:35 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020423W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 03 23:25:45 2023
Response via : Initial Calibration

