

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\WX020923\
 Data File : VX034061.D
 Acq On : 07 Feb 2023 15:12
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Feb 07 15:41:49 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020423W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 03 23:25:45 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	50.000	50.000	0.0	95	0.00
2 T	Dichlorodifluoromethane	50.000	51.525	-3.0	91	0.00
3 P	Chloromethane	50.000	47.794	4.4	88	0.00
4 C	Vinyl Chloride	50.000	44.976	10.0#	82	0.00
5 T	Bromomethane	50.000	47.895	4.2	90	0.00
6 T	Chloroethane	50.000	48.119	3.8	86	0.00
7 T	Trichlorofluoromethane	50.000	48.885	2.2	89	0.00
8 T	Diethyl Ether	50.000	45.255	9.5	83	0.00
9 T	1,1,2-Trichlorotrifluoroeth	50.000	48.057	3.9	88	0.00
10 T	Methyl Iodide	50.000	47.406	5.2	84	0.00
11 T	Tert butyl alcohol	250.000	261.169	-4.5	100	-0.01
12 CM	1,1-Dichloroethene	50.000	46.370	7.3#	85	0.00
13 T	Acrolein	250.000	230.520	7.8	86	0.00
14 T	Allyl chloride	50.000	46.594	6.8	84	0.00
15 T	Acrylonitrile	250.000	229.867	8.1	84	0.00
16 T	Acetone	250.000	223.707	10.5	88	0.00
17 T	Carbon Disulfide	50.000	44.754	10.5	82	0.00
18 T	Methyl Acetate	50.000	46.773	6.5	84	0.00
19 T	Methyl tert-butyl Ether	50.000	48.103	3.8	87	0.00
20 T	Methylene Chloride	50.000	45.690	8.6	88	0.00
21 T	trans-1,2-Dichloroethene	50.000	46.651	6.7	87	0.00
22 T	Diisopropyl ether	50.000	46.249	7.5	85	0.00
23 T	Vinyl Acetate	250.000	234.495	6.2	83	0.00
24 P	1,1-Dichloroethane	50.000	47.485	5.0	87	0.00
25 T	2-Butanone	250.000	227.177	9.1	87	0.00
26 T	2,2-Dichloropropane	50.000	48.475	3.0	88	0.00
27 T	cis-1,2-Dichloroethene	50.000	47.196	5.6	87	0.00
28 T	Bromochloromethane	50.000	42.380	15.2	81	0.00
29 T	Tetrahydrofuran	250.000	224.553	10.2	83	0.00
30 C	Chloroform	50.000	48.202	3.6#	89	0.00
31 T	Cyclohexane	50.000	46.395	7.2	84	0.00
32 T	1,1,1-Trichloroethane	50.000	48.867	2.3	89	0.00
33 S	1,2-Dichloroethane-d4	50.000	45.424	9.2	89	0.00
34 I	1,4-Difluorobenzene	50.000	50.000	0.0	93	0.00
35 S	Dibromofluoromethane	50.000	45.951	8.1	87	0.00
36 T	1,1-Dichloropropene	50.000	48.208	3.6	86	0.00
37 T	Ethyl Acetate	50.000	47.158	5.7	83	0.00
38 T	Carbon Tetrachloride	50.000	49.197	1.6	88	0.00
39 T	Methylcyclohexane	50.000	48.089	3.8	86	0.00
40 TM	Benzene	50.000	48.032	3.9	86	0.00
41 T	Methacrylonitrile	50.000	48.022	4.0	86	0.00
42 TM	1,2-Dichloroethane	50.000	50.794	-1.6	91	0.00
43 T	Isopropyl Acetate	50.000	47.684	4.6	82	0.00
44 TM	Trichloroethene	50.000	48.321	3.4	87	0.00
45 C	1,2-Dichloropropane	50.000	48.361	3.3#	86	0.00
46 T	Dibromomethane	50.000	49.477	1.0	89	0.00
47 T	Bromodichloromethane	50.000	50.089	-0.2	87	0.00
48 T	Methyl methacrylate	50.000	49.070	1.9	85	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX020923\
 Data File : VX034061.D
 Acq On : 07 Feb 2023 15:12
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Feb 07 15:41:49 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020423W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 03 23:25:45 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	1000.000	1197.063	-19.7	110	0.00
50 S	Toluene-d8	50.000	46.238	7.5	88	0.00
51 T	4-Methyl-2-Pentanone	250.000	244.511	2.2	86	0.00
52 CM	Toluene	50.000	48.932	2.1#	88	0.00
53 T	t-1,3-Dichloropropene	50.000	51.723	-3.4	88	0.00
54 T	cis-1,3-Dichloropropene	50.000	50.316	-0.6	86	0.00
55 T	1,1,2-Trichloroethane	50.000	49.622	0.8	88	0.00
56 T	Ethyl methacrylate	50.000	51.084	-2.2	86	0.00
57 T	1,3-Dichloropropane	50.000	49.600	0.8	88	0.00
58 T	2-Chloroethyl Vinyl ether	250.000	243.523	2.6	85	0.00
59 T	2-Hexanone	250.000	247.248	1.1	87	0.00
60 T	Dibromochloromethane	50.000	50.861	-1.7	88	0.00
61 T	1,2-Dibromoethane	50.000	49.447	1.1	88	0.00
62 S	4-Bromofluorobenzene	50.000	48.222	3.6	92	0.00
63 I	Chlorobenzene-d5	50.000	50.000	0.0	95	0.00
64 T	Tetrachloroethene	50.000	47.613	4.8	89	0.00
65 PM	Chlorobenzene	50.000	49.074	1.9	90	0.00
66 T	1,1,1,2-Tetrachloroethane	50.000	49.909	0.2	90	0.00
67 C	Ethyl Benzene	50.000	49.739	0.5#	91	0.00
68 T	m/p-Xylenes	100.000	99.166	0.8	91	0.00
69 T	o-Xylene	50.000	50.457	-0.9	91	0.00
70 T	Styrene	50.000	51.431	-2.9	92	0.00
71 P	Bromoform	50.000	52.101	-4.2	91	0.00
72 I	1,4-Dichlorobenzene-d4	50.000	50.000	0.0	98	0.00
73 T	Isopropylbenzene	50.000	49.004	2.0	92	0.00
74 T	N-ethyl acetate	50.000	48.520	3.0	87	0.00
75 P	1,1,2,2-Tetrachloroethane	50.000	46.144	7.7	90	0.00
76 T	1,2,3-Trichloropropane	50.000	49.691	0.6	102	0.00
77 T	Bromobenzene	50.000	48.154	3.7	92	0.00
78 T	n-propylbenzene	50.000	49.608	0.8	93	0.00
79 T	2-Chlorotoluene	50.000	47.774	4.5	92	0.00
80 T	1,3,5-Trimethylbenzene	50.000	49.566	0.9	93	0.00
81 T	trans-1,4-Dichloro-2-butene	50.000	47.023	6.0	86	0.00
82 T	4-Chlorotoluene	50.000	49.255	1.5	93	0.00
83 T	tert-Butylbenzene	50.000	49.060	1.9	94	0.00
84 T	1,2,4-Trimethylbenzene	50.000	49.865	0.3	94	0.00
85 T	sec-Butylbenzene	50.000	49.526	0.9	92	0.00
86 T	p-Isopropyltoluene	50.000	50.183	-0.4	94	0.00
87 T	1,3-Dichlorobenzene	50.000	48.400	3.2	93	0.00
88 T	1,4-Dichlorobenzene	50.000	47.608	4.8	93	0.00
89 T	n-Butylbenzene	50.000	51.071	-2.1	94	0.00
90 T	Hexachloroethane	50.000	49.646	0.7	93	0.00
91 T	1,2-Dichlorobenzene	50.000	48.435	3.1	93	0.00
92 T	1,2-Dibromo-3-Chloropropane	50.000	48.510	3.0	90	0.00
93 T	1,2,4-Trichlorobenzene	50.000	50.509	-1.0	94	0.00
94 T	Hexachlorobutadiene	50.000	48.187	3.6	96	0.00
95 T	Naphthalene	50.000	48.738	2.5	89	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX020923\
Data File : VX034061.D
Acq On : 07 Feb 2023 15:12
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Feb 07 15:41:49 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020423W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 03 23:25:45 2023
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	50.000	48.535	2.9	93	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6