


Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX020923\
Data File : VX034060.D
Acq On : 07 Feb 2023 14:44
Operator : JC/MD
Sample : BFB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020423W.M
Title : SW846 8260
Last Update : Fri Feb 03 23:25:45 2023



AutoFind: Scans 1638, 1639, 1640; Background Corrected with Scan 1632

Target	Rel. to	Lower	Upper	Rel.	Raw	Result
Mass	Mass	Limit%	Limit%	Abn%	Abn	Pass/Fail
50	95	15	40	16.6	6344	PASS
75	95	30	60	46.7	17883	PASS
95	95	100	100	100.0	38307	PASS
96	95	5	9	6.6	2510	PASS
173	174	0.00	2	0.5	170	PASS
174	95	50	100	94.0	35992	PASS
175	174	5	9	8.0	2883	PASS
176	174	95	101	96.3	34667	PASS
177	176	5	9	6.5	2239	PASS