

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX021020\
 Data File : VX014888.D
 Acq On : 10 Feb 2020 15:23
 Operator : JC/SP
 Sample : L1448-02
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 STORAGE-BLANK-WATER-REF-5

Quant Time: Feb 11 07:09:33 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X021020W.M
 Quant Title : SW846 8260
 QLast Update : Mon Feb 10 12:09:20 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	448793	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	692275	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	642462	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	341125	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.05	65	255911	50.53	ug/l	0.00
Spiked Amount			Recovery	=	101.06%	
35) Dibromofluoromethane	5.49	113	207501	48.76	ug/l	0.00
Spiked Amount			Recovery	=	97.52%	
50) Toluene-d8	8.71	98	811078	49.93	ug/l	0.00
Spiked Amount			Recovery	=	99.86%	
62) 4-Bromofluorobenzene	11.13	95	295419	50.48	ug/l	0.00
Spiked Amount			Recovery	=	100.96%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX021020\
Data File : VX014888.D
Acq On : 10 Feb 2020 15:23
Operator : JC/SP
Sample : L1448-02
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-WATER-REF-5

Quant Time: Feb 11 07:09:33 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X021020W.M
Quant Title : SW846 8260
QLast Update : Mon Feb 10 12:09:20 2020
Response via : Initial Calibration

