

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX021020\
 Data File : VX014896.D
 Acq On : 10 Feb 2020 18:23
 Operator : JC/SP
 Sample : L1456-01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FM-5213

Quant Time: Feb 11 07:29:22 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X021020W.M
 Quant Title : SW846 8260
 QLast Update : Mon Feb 10 12:09:20 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	452791	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	695569	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	666063	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	366081	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	253360	49.58	ug/l	0.00
Spiked Amount	50.000		Recovery	=	99.16%	
35) Dibromofluoromethane	5.49	113	211012	49.35	ug/l	0.00
Spiked Amount	50.000		Recovery	=	98.70%	
50) Toluene-d8	8.71	98	826562	50.64	ug/l	0.00
Spiked Amount	50.000		Recovery	=	101.28%	
62) 4-Bromofluorobenzene	11.13	95	321523	54.68	ug/l	0.00
Spiked Amount	50.000		Recovery	=	109.36%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX021020\
Data File : VX014896.D
Acq On : 10 Feb 2020 18:23
Operator : JC/SP
Sample : L1456-01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FM-5213

Quant Time: Feb 11 07:29:22 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X021020W.M
Quant Title : SW846 8260
QLast Update : Mon Feb 10 12:09:20 2020
Response via : Initial Calibration

