

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021022\
 Data File : VX026913.D
 Acq On : 10 Feb 2022 12:04
 Operator : JC/MD
 Sample : VX0210WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0210WBL01

Quant Time: Feb 11 03:39:55 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X011022W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Mon Jan 10 12:50:55 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	4.904	128	17910	30.000	ug/l	0.00
28) 1,4-Difluorobenzene	6.763	114	100112	30.000	ug/l	0.00
57) Chlorobenzene-d5	10.055	117	85277	30.000	ug/l	0.00

System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	5.958	65	40541	28.150	ug/l	0.00
Spiked Amount	30.000	Range	91 - 110	Recovery	=	93.833%
60) 4-Bromofluorobenzene	11.079	95	38679	27.178	ug/l	0.00
Spiked Amount	30.000	Range	63 - 112	Recovery	=	90.600%
63) Toluene-d8	8.647	98	116224	29.668	ug/l	0.00
Spiked Amount	30.000	Range	91 - 112	Recovery	=	98.900%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021022\
Data File : VX026913.D
Acq On : 10 Feb 2022 12:04
Operator : JC/MD
Sample : VX0210WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0210WBL01

Quant Time: Feb 11 03:39:55 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X011022W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Mon Jan 10 12:50:55 2022
Response via : Initial Calibration

