

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021022\
 Data File : VX026914.D
 Acq On : 10 Feb 2022 12:31
 Operator : JC/MD
 Sample : VX0210WBL02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0210WBL02

Quant Time: Feb 11 03:58:40 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020822W.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 08 07:07:46 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	121707	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	205644	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	180628	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	74766	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	79556	54.454	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 108.900%	
35) Dibromofluoromethane	5.391	113	64784	52.877	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 105.760%	
50) Toluene-d8	8.653	98	238908	53.122	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 106.240%	
62) 4-Bromofluorobenzene	11.079	95	83267	52.589	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 105.180%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021022\
Data File : VX026914.D
Acq On : 10 Feb 2022 12:31
Operator : JC/MD
Sample : VX0210WBL02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0210WBL02

Quant Time: Feb 11 03:58:40 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020822W.M
Quant Title : SW846 8260
QLast Update : Tue Feb 08 07:07:46 2022
Response via : Initial Calibration

