

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021022\
 Data File : VX026917.D
 Acq On : 10 Feb 2022 13:56
 Operator : JC/MD
 Sample : N1416-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 001-WILLETS-PT-BLVD(FEB)

Quant Time: Feb 11 03:40:39 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X011022W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Mon Jan 10 12:50:55 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	4.904	128	15059	30.000	ug/l	0.00
28) 1,4-Difluorobenzene	6.769	114	82611	30.000	ug/l	0.00
57) Chlorobenzene-d5	10.055	117	69253	30.000	ug/l	0.00
System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	5.958	65	34594	28.568	ug/l	0.00
Spiked Amount	30.000	Range	91 - 110	Recovery	=	95.233%
60) 4-Bromofluorobenzene	11.079	95	32243	27.897	ug/l	0.00
Spiked Amount	30.000	Range	63 - 112	Recovery	=	93.000%
63) Toluene-d8	8.653	98	95602	30.050	ug/l	0.00
Spiked Amount	30.000	Range	91 - 112	Recovery	=	100.167%
Target Compounds						
					Qvalue	
15) Acetone	2.386	58	14167	95.346	ug/l	81
16) Carbon Disulfide	2.514	76	2270	0.993	ug/l #	91
24) Methyl tert-Butyl Ether	3.123	73	1672	0.555	ug/l	99
30) 2-Butanone	4.568	43	7729	10.431	ug/l #	96
58) 4-Methyl-2-Pentanone	8.574	43	3156	2.443	ug/l	97
62) Toluene	8.720	91	1380	0.373	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021022\
Data File : VX026917.D
Acq On : 10 Feb 2022 13:56
Operator : JC/MD
Sample : N1416-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
001-WILLETS-PT-BLVD(FEB)

Quant Time: Feb 11 03:40:39 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X011022W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Mon Jan 10 12:50:55 2022
Response via : Initial Calibration

