

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021022\
 Data File : VX026924.D
 Acq On : 10 Feb 2022 16:45
 Operator : JC/MD
 Sample : N1446-07DL 10X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-136ADL

Quant Time: Feb 11 04:00:46 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020822W.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 08 07:07:46 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	116193	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	197439	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	172875	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	70791	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	76372	54.756	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 109.520%	
35) Dibromofluoromethane	5.391	113	62086	52.781	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 105.560%	
50) Toluene-d8	8.653	98	226982	52.568	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 105.140%	
62) 4-Bromofluorobenzene	11.085	95	80851	53.185	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 106.380%	
Target Compounds						Qvalue
11) Tert butyl alcohol	2.971	59	15834	62.840	ug/l	# 95
19) Methyl tert-butyl Ether	3.117	73	174474	41.452	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021022\
Data File : VX026924.D
Acq On : 10 Feb 2022 16:45
Operator : JC/MD
Sample : N1446-07DL 10X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-136ADL

Quant Time: Feb 11 04:00:46 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020822W.M
Quant Title : SW846 8260
QLast Update : Tue Feb 08 07:07:46 2022
Response via : Initial Calibration

