

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021022\
 Data File : VX026925.D
 Acq On : 10 Feb 2022 17:08
 Operator : JC/MD
 Sample : N1446-08DL 10X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-136BDL

Quant Time: Feb 11 04:01:06 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020822W.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 08 07:07:46 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	114756	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	194166	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	167996	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	72177	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	76723	55.696	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 111.400%	
35) Dibromofluoromethane	5.391	113	60880	52.628	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 105.260%	
50) Toluene-d8	8.653	98	222536	52.407	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 104.820%	
62) 4-Bromofluorobenzene	11.079	95	79501	53.179	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 106.360%	
Target Compounds						Qvalue
11) Tert butyl alcohol	2.971	59	11376	45.713	ug/l	97
19) Methyl tert-butyl Ether	3.117	73	152646	36.720	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021022\
Data File : VX026925.D
Acq On : 10 Feb 2022 17:08
Operator : JC/MD
Sample : N1446-08DL 10X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-136BDL

Quant Time: Feb 11 04:01:06 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020822W.M
Quant Title : SW846 8260
QLast Update : Tue Feb 08 07:07:46 2022
Response via : Initial Calibration

