

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021022\
 Data File : VX026927.D
 Acq On : 10 Feb 2022 17:55
 Operator : JC/MD
 Sample : N1446-03DL 20X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-118DDL

Quant Time: Feb 11 04:01:41 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020822W.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 08 07:07:46 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	121494	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	207658	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	183679	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	77091	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	80974	55.522	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	111.040%
35) Dibromofluoromethane	5.391	113	64485	52.122	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	104.240%
50) Toluene-d8	8.653	98	241940	53.275	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	106.540%
62) 4-Bromofluorobenzene	11.079	95	86227	53.930	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	107.860%
Target Compounds						
						Qvalue
11) Tert butyl alcohol	2.971	59	10032	38.077	ug/l	# 75
19) Methyl tert-butyl Ether	3.117	73	193591	43.987	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021022\
Data File : VX026927.D
Acq On : 10 Feb 2022 17:55
Operator : JC/MD
Sample : N1446-03DL 20X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-118DDL

Quant Time: Feb 11 04:01:41 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020822W.M
Quant Title : SW846 8260
QLast Update : Tue Feb 08 07:07:46 2022
Response via : Initial Calibration

