

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021022\
 Data File : VX026931.D
 Acq On : 10 Feb 2022 19:27
 Operator : JC/MD
 Sample : N1438-03DL 20X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-25-R-DDL

Quant Time: Feb 11 04:02:45 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020822W.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 08 07:07:46 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	119376	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	202674	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	178772	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	74726	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	79084	55.188	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	110.380%
35) Dibromofluoromethane	5.391	113	63895	52.916	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	105.840%
50) Toluene-d8	8.653	98	233764	52.740	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	105.480%
62) 4-Bromofluorobenzene	11.079	95	82543	52.896	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	105.800%
Target Compounds						
						Qvalue
4) Vinyl Chloride	1.380	62	1567	1.240	ug/l	95
27) cis-1,2-Dichloroethene	4.495	96	58853	39.871	ug/l	89
44) Trichloroethene	7.129	130	7498	4.498	ug/l	97
64) Tetrachloroethene	9.275	164	2870	1.598	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021022\
Data File : VX026931.D
Acq On : 10 Feb 2022 19:27
Operator : JC/MD
Sample : N1438-03DL 20X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-25-R-DDL

Quant Time: Feb 11 04:02:45 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020822W.M
Quant Title : SW846 8260
QLast Update : Tue Feb 08 07:07:46 2022
Response via : Initial Calibration

