

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021022\
 Data File : VX026933.D
 Acq On : 10 Feb 2022 20:13
 Operator : JC/MD
 Sample : N1425-04 10X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 20422DL

Quant Time: Feb 11 04:03:19 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020822W.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 08 07:07:46 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	115802	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	195253	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	180836	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	78948	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	75064	54.000	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery = 108.000%			
35) Dibromofluoromethane	5.391	113	61191	52.602	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 105.200%			
50) Toluene-d8	8.653	98	230320	53.938	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery = 107.880%			
62) 4-Bromofluorobenzene	11.079	95	86507	57.543	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 115.080%			
Target Compounds					Qvalue	
11) Tert butyl alcohol	2.959	59	4441	17.684	ug/l	99
16) Acetone	2.386	43	8052	13.188	ug/l #	88
25) 2-Butanone	4.568	43	4220	4.290	ug/l	95
40) Benzene	6.044	78	132016	24.683	ug/l	100
51) 4-Methyl-2-Pentanone	8.580	43	541	0.278	ug/l	94
52) Toluene	8.720	92	67742	20.005	ug/l	99
67) Ethyl Benzene	10.195	91	14550	2.198	ug/l	96
68) m/p-Xylenes	10.299	106	27473	10.788	ug/l	100
69) o-Xylene	10.646	106	23097	9.256	ug/l	99
70) Styrene	10.659	104	13151	3.249	ug/l #	82
73) Isopropylbenzene	10.963	105	2150	0.346	ug/l	98
78) n-propylbenzene	11.305	91	7074	1.026	ug/l	97
80) 1,3,5-Trimethylbenzene	11.451	105	10833	2.117	ug/l	98
84) 1,2,4-Trimethylbenzene	11.756	105	42345	8.203	ug/l	98
85) sec-Butylbenzene	11.890	105	2403	0.389	ug/l	88
89) n-Butylbenzene	12.335	91	4360	1.017	ug/l #	43
95) Naphthalene	13.780	128	95034	16.449	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021022\
Data File : VX026933.D
Acq On : 10 Feb 2022 20:13
Operator : JC/MD
Sample : N1425-04 10X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
20422DL

Quant Time: Feb 11 04:03:19 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020822W.M
Quant Title : SW846 8260
QLast Update : Tue Feb 08 07:07:46 2022
Response via : Initial Calibration

