

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021022\
 Data File : VX026934.D
 Acq On : 10 Feb 2022 20:36
 Operator : JC/MD
 Sample : N1425-03 40X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 31377

Quant Time: Feb 11 04:03:40 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020822W.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 08 07:07:46 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	121359	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	206805	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	185907	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	81554	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	80538	55.284	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 110.560%	
35) Dibromofluoromethane	5.391	113	64339	52.219	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 104.440%	
50) Toluene-d8	8.653	98	241562	53.411	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 106.820%	
62) 4-Bromofluorobenzene	11.079	95	88382	55.506	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 111.020%	
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	16126	25.203	ug/l	96
39) Methylcyclohexane	7.379	83	3968	1.665	ug/l	89
40) Benzene	6.050	78	5526	0.975	ug/l	95
52) Toluene	8.720	92	32534	9.071	ug/l	100
67) Ethyl Benzene	10.195	91	28292	4.158	ug/l	97
68) m/p-Xylenes	10.305	106	34622	13.224	ug/l	100
69) o-Xylene	10.640	106	19710	7.683	ug/l	97
73) Isopropylbenzene	10.963	105	6154	0.958	ug/l	98
78) n-propylbenzene	11.305	91	19550	2.745	ug/l	99
80) 1,3,5-Trimethylbenzene	11.451	105	20655	3.907	ug/l	97
84) 1,2,4-Trimethylbenzene	11.756	105	80356	15.068	ug/l	98
85) sec-Butylbenzene	11.890	105	6623	1.037	ug/l	87
89) n-Butylbenzene	12.329	91	12223	2.761	ug/l #	50
95) Naphthalene	13.780	128	9683	1.622	ug/l #	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021022\
Data File : VX026934.D
Acq On : 10 Feb 2022 20:36
Operator : JC/MD
Sample : N1425-03 40X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
31377

Quant Time: Feb 11 04:03:40 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020822W.M
Quant Title : SW846 8260
QLast Update : Tue Feb 08 07:07:46 2022
Response via : Initial Calibration

