

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021022\
 Data File : VX026935.D
 Acq On : 10 Feb 2022 20:59
 Operator : JC/MD
 Sample : N1425-05
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 20722

Quant Time: Feb 11 04:04:02 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020822W.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 08 07:07:46 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	118347	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	200331	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	177570	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	74772	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	77374	54.464	ug/l	0.00
Spiked Amount	50.000	Range 78 - 117	Recovery	= 108.920%		
35) Dibromofluoromethane	5.391	113	63050	52.826	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 105.660%		
50) Toluene-d8	8.653	98	233618	53.324	ug/l	0.00
Spiked Amount	50.000	Range 92 - 112	Recovery	= 106.640%		
62) 4-Bromofluorobenzene	11.079	95	82545	53.516	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	= 107.040%		
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	14141	22.664	ug/l	95
25) 2-Butanone	4.568	43	3361	3.343	ug/l	91
39) Methylcyclohexane	7.379	83	1185	0.513	ug/l #	87
40) Benzene	6.044	78	4947	0.901	ug/l	95
51) 4-Methyl-2-Pentanone	8.580	43	1975	0.990	ug/l	95
52) Toluene	8.720	92	3675	1.058	ug/l	96
67) Ethyl Benzene	10.195	91	8221	1.265	ug/l	97
68) m/p-Xylenes	10.305	106	11947	4.777	ug/l	100
69) o-Xylene	10.646	106	8663	3.535	ug/l	99
73) Isopropylbenzene	10.964	105	3554	0.603	ug/l	99
78) n-propylbenzene	11.305	91	12371	1.895	ug/l	99
80) 1,3,5-Trimethylbenzene	11.451	105	17044	3.516	ug/l	98
84) 1,2,4-Trimethylbenzene	11.756	105	64413	13.174	ug/l	99
85) sec-Butylbenzene	11.890	105	5436	0.928	ug/l	80
89) n-Butylbenzene	12.329	91	9132	2.250	ug/l #	46
95) Naphthalene	13.780	128	14635	2.675	ug/l #	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021022\
Data File : VX026935.D
Acq On : 10 Feb 2022 20:59
Operator : JC/MD
Sample : N1425-05
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
20722

Quant Time: Feb 11 04:04:02 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020822W.M
Quant Title : SW846 8260
QLast Update : Tue Feb 08 07:07:46 2022
Response via : Initial Calibration

