

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021022\
 Data File : VX026947.D
 Acq On : 11 Feb 2022 02:45
 Operator : JC/MD
 Sample : N1467-05
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 41 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-8

Quant Time: Feb 11 04:08:38 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020822W.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 08 07:07:46 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	114659	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	193063	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	184071	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	81061	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	74223	53.927	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 107.860%	
35) Dibromofluoromethane	5.391	113	60235	52.368	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 104.740%	
50) Toluene-d8	8.653	98	230686	54.637	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 109.280%	
62) 4-Bromofluorobenzene	11.085	95	86932	58.482	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 116.960%	
Target Compounds						
					Qvalue	
4) Vinyl Chloride	1.380	62	15742	12.966	ug/l	95
16) Acetone	2.386	43	4425	7.320	ug/l	99
19) Methyl tert-butyl Ether	3.117	73	13381	3.222	ug/l #	85
21) trans-1,2-Dichloroethene	3.099	96	2030	1.644	ug/l	90
27) cis-1,2-Dichloroethene	4.495	96	86230	60.822	ug/l	89
31) Cyclohexane	5.476	56	4809	2.199	ug/l	93
39) Methylcyclohexane	7.385	83	22984	10.328	ug/l	99
40) Benzene	6.043	78	3976	0.752	ug/l #	89
44) Trichloroethene	7.141	130	1140	0.718	ug/l	98
52) Toluene	8.720	92	130442	38.958	ug/l	99
64) Tetrachloroethene	9.275	164	1590	0.860	ug/l	92
67) Ethyl Benzene	10.195	91	638840	94.818	ug/l	100
68) m/p-Xylenes	10.305	106	1322578	510.206	ug/l	96
69) o-Xylene	10.646	106	869295	342.233	ug/l	96
73) Isopropylbenzene	10.963	105	365731	57.280	ug/l	99
78) n-propylbenzene	11.305	91	876929	123.887	ug/l	100
80) 1,3,5-Trimethylbenzene	11.457	105	2830619	538.651	ug/l	97
83) tert-Butylbenzene	11.719	119	56930	11.367	ug/l	93
84) 1,2,4-Trimethylbenzene	11.768	105	6956556	1312.429	ug/l	91
85) sec-Butylbenzene	11.896	105	219216	34.529	ug/l	100
86) p-Isopropyltoluene	12.012	119	267096	49.660	ug/l	99
95) Naphthalene	13.774	128	481569	81.181	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021022\
Data File : VX026947.D
Acq On : 11 Feb 2022 02:45
Operator : JC/MD
Sample : N1467-05
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 41 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-8

Quant Time: Feb 11 04:08:38 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020822W.M
Quant Title : SW846 8260
QLast Update : Tue Feb 08 07:07:46 2022
Response via : Initial Calibration

