

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021022\
 Data File : VX026949.D
 Acq On : 11 Feb 2022 03:31
 Operator : JC/MD
 Sample : N1467-09
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 43 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-23S

Quant Time: Feb 11 04:09:24 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020822W.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 08 07:07:46 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	125521	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	210626	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	189879	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	79127	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	79926	53.045	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 106.100%	
35) Dibromofluoromethane	5.391	113	65921	52.532	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 105.060%	
50) Toluene-d8	8.653	98	245606	53.320	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 106.640%	
62) 4-Bromofluorobenzene	11.079	95	88929	54.836	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 109.680%	
Target Compounds						
						Qvalue
4) Vinyl Chloride	1.380	62	26023	19.579	ug/l	99
12) 1,1-Dichloroethene	2.331	96	1971	1.534	ug/l	93
16) Acetone	2.386	43	1689	2.552	ug/l	93
21) trans-1,2-Dichloroethene	3.099	96	6981	5.164	ug/l	97
27) cis-1,2-Dichloroethene	4.495	96	1552175	1000.075	ug/l	88
29) Tetrahydrofuran	5.026	42	2542	3.507	ug/l	97
44) Trichloroethene	7.129	130	429004	247.641	ug/l	95
64) Tetrachloroethene	9.275	164	389473	204.227	ug/l	97
68) m/p-Xylenes	10.305	106	882	0.330	ug/l	97
69) o-Xylene	10.647	106	548	0.209	ug/l	87
80) 1,3,5-Trimethylbenzene	11.451	105	1909	0.372	ug/l	99
84) 1,2,4-Trimethylbenzene	11.756	105	6416	1.240	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021022\
Data File : VX026949.D
Acq On : 11 Feb 2022 03:31
Operator : JC/MD
Sample : N1467-09
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 43 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-23S

Quant Time: Feb 11 04:09:24 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020822W.M
Quant Title : SW846 8260
QLast Update : Tue Feb 08 07:07:46 2022
Response via : Initial Calibration

