

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021022\
 Data File : VX026954.D
 Acq On : 11 Feb 2022 05:26
 Operator : JC/MD
 Sample : N1497-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 48 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-7D

Quant Time: Feb 11 07:36:52 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020822W.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 08 07:07:46 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	118759	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	199696	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	184938	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	79608	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	77192	54.148	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 108.300%	
35) Dibromofluoromethane	5.391	113	63160	53.087	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 106.180%	
50) Toluene-d8	8.652	98	236240	54.094	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 108.180%	
62) 4-Bromofluorobenzene	11.079	95	87951	57.202	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 114.400%	
Target Compounds						
					Qvalue	
4) Vinyl Chloride	1.379	62	722	0.574	ug/l	96
16) Acetone	2.385	43	2043	3.263	ug/l	98
27) cis-1,2-Dichloroethene	4.495	96	9064	6.173	ug/l	87
44) Trichloroethene	7.134	130	1234	0.751	ug/l	86
64) Tetrachloroethene	9.274	164	2306	1.241	ug/l	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021022\
Data File : VX026954.D
Acq On : 11 Feb 2022 05:26
Operator : JC/MD
Sample : N1497-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 48 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-7D

Quant Time: Feb 11 07:36:52 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020822W.M
Quant Title : SW846 8260
QLast Update : Tue Feb 08 07:07:46 2022
Response via : Initial Calibration

