

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021022\
 Data File : VX026955.D
 Acq On : 11 Feb 2022 05:49
 Operator : JC/MD
 Sample : N1497-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 49 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-5

Quant Time: Feb 11 07:37:13 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020822W.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 08 07:07:46 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	126615	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	213889	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	187938	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	78718	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	81910	53.892	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 107.780%	
35) Dibromofluoromethane	5.391	113	67163	52.706	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 105.420%	
50) Toluene-d8	8.653	98	246605	52.720	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 105.440%	
62) 4-Bromofluorobenzene	11.079	95	87528	53.149	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 106.300%	
Target Compounds						
					Qvalue	
27) cis-1,2-Dichloroethene	4.501	96	1206	0.770	ug/l	92
44) Trichloroethene	7.129	130	527	0.300	ug/l	88
64) Tetrachloroethene	9.275	164	826	0.438	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021022\
Data File : VX026955.D
Acq On : 11 Feb 2022 05:49
Operator : JC/MD
Sample : N1497-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 49 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-5

Quant Time: Feb 11 07:37:13 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020822W.M
Quant Title : SW846 8260
QLast Update : Tue Feb 08 07:07:46 2022
Response via : Initial Calibration

