

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\WX021022\
 Data File : VX026957.D
 Acq On : 11 Feb 2022 06:35
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 51 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Feb 11 07:37:49 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020822W.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 08 07:07:46 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	50.000	50.000	0.0	88	0.00
2 T	Dichlorodifluoromethane	50.000	51.035	-2.1	91	0.00
3 P	Chloromethane	50.000	53.983	-8.0	99	0.00
4 C	Vinyl Chloride	50.000	51.855	-3.7#	93	0.00
5 T	Bromomethane	50.000	49.564	0.9	96	0.00
6 T	Chloroethane	50.000	52.757	-5.5	88	0.00
7 T	Trichlorofluoromethane	50.000	51.270	-2.5	91	0.00
8 T	Diethyl Ether	50.000	49.727	0.5	92	0.00
9 T	1,1,2-Trichlorotrifluoroeth	50.000	49.899	0.2	90	0.00
10 T	Methyl Iodide	50.000	46.869	6.3	87	0.00
11 T	Tert butyl alcohol	250.000	208.683	16.5	78	0.01
12 CM	1,1-Dichloroethene	50.000	48.479	3.0#	88	0.00
13 T	Acrolein	250.000	250.275	-0.1	91	0.00
14 T	Allyl chloride	50.000	49.603	0.8	89	0.00
15 T	Acrylonitrile	250.000	255.079	-2.0	92	0.00
16 T	Acetone	250.000	244.706	2.1	93	0.00
17 T	Carbon Disulfide	50.000	46.390	7.2	85	0.00
18 T	Methyl Acetate	50.000	48.381	3.2	91	0.00
19 T	Methyl tert-butyl Ether	50.000	48.008	4.0	88	0.00
20 T	Methylene Chloride	50.000	49.996	0.0	94	0.00
21 T	trans-1,2-Dichloroethene	50.000	49.951	0.1	89	0.00
22 T	Diisopropyl ether	50.000	50.722	-1.4	92	0.00
23 T	Vinyl Acetate	250.000	257.416	-3.0	91	0.00
24 P	1,1-Dichloroethane	50.000	50.708	-1.4	92	0.00
25 T	2-Butanone	250.000	252.375	-0.9	91	0.00
26 T	2,2-Dichloropropane	50.000	45.549	8.9	80	0.00
27 T	cis-1,2-Dichloroethene	50.000	51.181	-2.4	91	0.00
28 T	Bromochloromethane	50.000	51.196	-2.4	93	0.00
29 T	Tetrahydrofuran	250.000	250.799	-0.3	91	0.00
30 C	Chloroform	50.000	50.688	-1.4#	94	0.00
31 T	Cyclohexane	50.000	49.682	0.6	90	0.00
32 T	1,1,1-Trichloroethane	50.000	49.164	1.7	88	0.00
33 S	1,2-Dichloroethane-d4	50.000	53.524	-7.0	102	0.00
34 I	1,4-Difluorobenzene	50.000	50.000	0.0	89	0.00
35 S	Dibromofluoromethane	50.000	51.574	-3.1	101	0.00
36 T	1,1-Dichloropropene	50.000	48.926	2.1	91	0.00
37 T	Ethyl Acetate	50.000	48.812	2.4	91	0.00
38 T	Carbon Tetrachloride	50.000	48.779	2.4	90	0.00
39 T	Methylcyclohexane	50.000	49.093	1.8	89	0.00
40 TM	Benzene	50.000	49.719	0.6	92	0.00
41 T	Methacrylonitrile	50.000	50.419	-0.8	92	0.00
42 TM	1,2-Dichloroethane	50.000	51.617	-3.2	94	0.00
43 T	Isopropyl Acetate	50.000	49.649	0.7	89	0.00
44 TM	Trichloroethene	50.000	48.358	3.3	89	0.00
45 C	1,2-Dichloropropane	50.000	49.339	1.3#	91	0.00
46 T	Dibromomethane	50.000	50.355	-0.7	93	0.00
47 T	Bromodichloromethane	50.000	52.848	-5.7	95	0.00
48 T	Methyl methacrylate	50.000	49.451	1.1	90	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021022\
 Data File : VX026957.D
 Acq On : 11 Feb 2022 06:35
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 51 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Feb 11 07:37:49 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020822W.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 08 07:07:46 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	1000.000	920.397	8.0	84	0.01
50 S	Toluene-d8	50.000	53.064	-6.1	103	0.00
51 T	4-Methyl-2-Pentanone	250.000	254.842	-1.9	94	0.00
52 CM	Toluene	50.000	50.102	-0.2#	94	0.00
53 T	t-1,3-Dichloropropene	50.000	47.581	4.8	91	0.00
54 T	cis-1,3-Dichloropropene	50.000	51.972	-3.9	90	0.00
55 T	1,1,2-Trichloroethane	50.000	52.772	-5.5	97	0.00
56 T	Ethyl methacrylate	50.000	51.220	-2.4	92	0.00
57 T	1,3-Dichloropropane	50.000	52.430	-4.9	96	0.00
58 T	2-Chloroethyl Vinyl ether	250.000	242.090	3.2	83	0.00
59 T	2-Hexanone	250.000	250.025	-0.0	92	0.00
60 T	Dibromochloromethane	50.000	54.125	-8.3	95	0.00
61 T	1,2-Dibromoethane	50.000	52.485	-5.0	95	0.00
62 S	4-Bromofluorobenzene	50.000	52.953	-5.9	104	0.00
63 I	Chlorobenzene-d5	50.000	50.000	0.0	94	0.00
64 T	Tetrachloroethene	50.000	47.299	5.4	88	0.00
65 PM	Chlorobenzene	50.000	49.281	1.4	96	0.00
66 T	1,1,1,2-Tetrachloroethane	50.000	50.215	-0.4	96	0.00
67 C	Ethyl Benzene	50.000	50.376	-0.8#	96	0.00
68 T	m/p-Xylenes	100.000	99.155	0.8	96	0.00
69 T	o-Xylene	50.000	49.514	1.0	96	0.00
70 T	Styrene	50.000	50.077	-0.2	95	0.00
71 P	Bromoform	50.000	44.745	10.5	92	0.00
72 I	1,4-Dichlorobenzene-d4	50.000	50.000	0.0	94	0.00
73 T	Isopropylbenzene	50.000	49.732	0.5	94	0.00
74 T	N-amyl acetate	50.000	47.831	4.3	90	0.00
75 P	1,1,2,2-Tetrachloroethane	50.000	47.775	4.5	96	0.00
76 T	1,2,3-Trichloropropane	50.000	48.830	2.3	95	0.00
77 T	Bromobenzene	50.000	47.871	4.3	93	0.00
78 T	n-propylbenzene	50.000	50.979	-2.0	96	0.00
79 T	2-Chlorotoluene	50.000	49.089	1.8	96	0.00
80 T	1,3,5-Trimethylbenzene	50.000	49.794	0.4	95	0.00
81 T	trans-1,4-Dichloro-2-butene	50.000	42.883	14.2	85	0.00
82 T	4-Chlorotoluene	50.000	49.562	0.9	95	0.00
83 T	tert-Butylbenzene	50.000	48.477	3.0	94	0.00
84 T	1,2,4-Trimethylbenzene	50.000	49.046	1.9	96	0.00
85 T	sec-Butylbenzene	50.000	50.003	-0.0	96	0.00
86 T	p-Isopropyltoluene	50.000	49.521	1.0	94	0.00
87 T	1,3-Dichlorobenzene	50.000	48.322	3.4	94	0.00
88 T	1,4-Dichlorobenzene	50.000	47.777	4.4	94	0.00
89 T	n-Butylbenzene	50.000	49.765	0.5	94	0.00
90 T	Hexachloroethane	50.000	53.913	-7.8	101	0.00
91 T	1,2-Dichlorobenzene	50.000	48.281	3.4	93	0.00
92 T	1,2-Dibromo-3-Chloropropane	50.000	48.482	3.0	89	0.00
93 T	1,2,4-Trichlorobenzene	50.000	48.378	3.2	92	0.00
94 T	Hexachlorobutadiene	50.000	45.802	8.4	88	0.00
95 T	Naphthalene	50.000	48.134	3.7	88	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021022\
Data File : VX026957.D
Acq On : 11 Feb 2022 06:35
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 51 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Feb 11 07:37:49 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020822W.M
Quant Title : SW846 8260
QLast Update : Tue Feb 08 07:07:46 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	50.000	47.358	5.3	90	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6