

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX021022\  
 Data File : VX026922.D  
 Acq On : 10 Feb 2022 15:59  
 Operator : JC/MD  
 Sample : VX0210WBSD02  
 Misc : 5.0mL/MSVOA\_X/WATER  
 ALS Vial : 13 Sample Multiplier: 1

Instrument :  
 MSVOA\_X  
 ClientSampleId :  
 VX0210WBSD02

Manual Integrations  
 APPROVED

Reviewed By : Semsettin Yesilyurt 02/11/2022  
 Supervised By : Mahesh Dadoda 02/11/2022

Quant Time: Feb 11 03:59:44 2022  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X020822W.M  
 Quant Title : SW846 8260  
 QLast Update : Tue Feb 08 07:07:46 2022  
 Response via : Initial Calibration

| Compound                     | R.T.           | QIon | Response            | Conc    | Units | Dev(Min) |
|------------------------------|----------------|------|---------------------|---------|-------|----------|
| -----                        |                |      |                     |         |       |          |
| Internal Standards           |                |      |                     |         |       |          |
| 1) Pentafluorobenzene        | 5.556          | 168  | 91993               | 50.000  | ug/l  | 0.00     |
| 34) 1,4-Difluorobenzene      | 6.763          | 114  | 151434              | 50.000  | ug/l  | 0.00     |
| 63) Chlorobenzene-d5         | 10.055         | 117  | 133902              | 50.000  | ug/l  | 0.00     |
| 72) 1,4-Dichlorobenzene-d4   | 12.024         | 152  | 60077               | 50.000  | ug/l  | 0.00     |
|                              |                |      |                     |         |       |          |
| System Monitoring Compounds  |                |      |                     |         |       |          |
| 33) 1,2-Dichloroethane-d4    | 5.958          | 65   | 60018               | 54.350  | ug/l  | 0.00     |
| Spiked Amount 50.000         | Range 78 - 117 |      | Recovery = 108.700% |         |       |          |
| 35) Dibromofluoromethane     | 5.391          | 113  | 46679               | 51.738  | ug/l  | 0.00     |
| Spiked Amount 50.000         | Range 75 - 124 |      | Recovery = 103.480% |         |       |          |
| 50) Toluene-d8               | 8.653          | 98   | 176884              | 53.411  | ug/l  | 0.00     |
| Spiked Amount 50.000         | Range 92 - 112 |      | Recovery = 106.820% |         |       |          |
| 62) 4-Bromofluorobenzene     | 11.085         | 95   | 63020               | 54.050  | ug/l  | 0.00     |
| Spiked Amount 50.000         | Range 83 - 123 |      | Recovery = 108.100% |         |       |          |
|                              |                |      |                     |         |       |          |
| Target Compounds             |                |      |                     |         |       | Qvalue   |
| 2) Dichlorodifluoromethane   | 1.172          | 85   | 20363               | 20.357  | ug/l  | 99       |
| 3) Chloromethane             | 1.294          | 50   | 18677               | 20.071  | ug/l  | 100      |
| 4) Vinyl Chloride            | 1.374          | 62   | 19201               | 19.711  | ug/l  | 99       |
| 5) Bromomethane              | 1.617          | 94   | 8184                | 19.799  | ug/l  | 94       |
| 6) Chloroethane              | 1.691          | 64   | 11411               | 19.571  | ug/l  | 99       |
| 7) Trichlorofluoromethane    | 1.892          | 101  | 28246               | 19.704  | ug/l  | 98       |
| 8) Diethyl Ether             | 2.136          | 74   | 10243               | 19.289  | ug/l  | 99       |
| 9) 1,1,2-Trichlorotrifluo... | 2.337          | 101  | 19113               | 20.341  | ug/l  | 99       |
| 10) Methyl Iodide            | 2.459          | 142  | 20962               | 17.613  | ug/l  | 98       |
| 11) Tert butyl alcohol       | 2.977          | 59   | 17667               | 88.560  | ug/l  | 96       |
| 12) 1,1-Dichloroethene       | 2.325          | 96   | 17785               | 18.891  | ug/l  | 90       |
| 13) Acrolein                 | 2.239          | 56   | 17477               | 94.800  | ug/l  | 100      |
| 14) Allyl chloride           | 2.666          | 41   | 32856               | 19.702  | ug/l  | 97       |
| 15) Acrylonitrile            | 3.068          | 53   | 58859               | 102.960 | ug/l  | 97       |
| 16) Acetone                  | 2.386          | 43   | 49328               | 101.705 | ug/l  | 100      |
| 17) Carbon Disulfide         | 2.514          | 76   | 47416               | 17.791  | ug/l  | 99       |
| 18) Methyl Acetate           | 2.709          | 43   | 26520               | 19.004  | ug/l  | 100      |
| 19) Methyl tert-butyl Ether  | 3.117          | 73   | 63626               | 19.093  | ug/l  | 99       |
| 20) Methylene Chloride       | 2.794          | 84   | 20454               | 19.463  | ug/l  | 97       |
| 21) trans-1,2-Dichloroethene | 3.093          | 96   | 19549               | 19.732  | ug/l  | 96       |
| 22) Diisopropyl ether        | 3.763          | 45   | 66852               | 19.911  | ug/l  | 99       |
| 23) Vinyl Acetate            | 3.727          | 43   | 284101              | 100.642 | ug/l  | 97       |
| 24) 1,1-Dichloroethane       | 3.611          | 63   | 35636               | 19.615  | ug/l  | 98       |
| 25) 2-Butanone               | 4.562          | 43   | 79266               | 101.437 | ug/l  | 99       |
| 26) 2,2-Dichloropropane      | 4.477          | 77   | 24828               | 20.226  | ug/l  | 98       |
| 27) cis-1,2-Dichloroethene   | 4.495          | 96   | 22835               | 20.075  | ug/l  | 96       |
| 28) Bromochloromethane       | 4.903          | 49   | 16018               | 21.269  | ug/l  | 98       |
| 29) Tetrahydrofuran          | 5.013          | 42   | 53036               | 99.834  | ug/l  | 99       |
| 30) Chloroform               | 5.099          | 83   | 36891               | 19.544  | ug/l  | 98       |
| 31) Cyclohexane              | 5.477          | 56   | 34119               | 19.445  | ug/l  | 98       |
| 32) 1,1,1-Trichloroethane    | 5.385          | 97   | 32107               | 19.292  | ug/l  | 99       |
| 36) 1,1-Dichloropropene      | 5.702          | 75   | 28104               | 19.746  | ug/l  | 99       |
| 37) Ethyl Acetate            | 4.714          | 43   | 31402               | 19.414  | ug/l  | 98       |
| 38) Carbon Tetrachloride     | 5.684          | 117  | 27786               | 18.617  | ug/l  | 97       |
| 39) Methylcyclohexane        | 7.385          | 83   | 33832               | 19.383  | ug/l  | 96       |
| 40) Benzene                  | 6.043          | 78   | 80862               | 19.494  | ug/l  | 98       |

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX021022\  
 Data File : VX026922.D  
 Acq On : 10 Feb 2022 15:59  
 Operator : JC/MD  
 Sample : VX0210WBSD02  
 Misc : 5.0mL/MSVOA\_X/WATER  
 ALS Vial : 13 Sample Multiplier: 1

Instrument :  
 MSVOA\_X  
 ClientSampleId :  
 VX0210WBSD02

Manual Integrations  
 APPROVED

Reviewed By :Semsettin Yesilyurt 02/11/2022  
 Supervised By :Mahesh Dadoda 02/11/2022

Quant Time: Feb 11 03:59:44 2022  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X020822W.M  
 Quant Title : SW846 8260  
 QLast Update : Tue Feb 08 07:07:46 2022  
 Response via : Initial Calibration

| Compound                      | R.T.   | QIon | Response | Conc    | Units  | Dev(Min) |
|-------------------------------|--------|------|----------|---------|--------|----------|
| 41) Methacrylonitrile         | 4.922  | 41   | 17392    | 19.839  | ug/l   | 97       |
| 42) 1,2-Dichloroethane        | 6.092  | 62   | 30072    | 20.573  | ug/l   | 99       |
| 43) Isopropyl Acetate         | 6.342  | 43   | 48591    | 19.524  | ug/l   | 99       |
| 44) Trichloroethene           | 7.129  | 130  | 22957    | 18.432  | ug/l   | 98       |
| 45) 1,2-Dichloropropane       | 7.433  | 63   | 20827    | 19.844  | ug/l   | 99       |
| 46) Dibromomethane            | 7.586  | 93   | 14641    | 20.161  | ug/l   | 98       |
| 47) Bromodichloromethane      | 7.824  | 83   | 28866    | 20.756  | ug/l   | 98       |
| 48) Methyl methacrylate       | 7.696  | 41   | 23986    | 19.451  | ug/l   | 97       |
| 49) 1,4-Dioxane               | 7.677  | 88   | 8468     | 398.501 | ug/l   | 98       |
| 51) 4-Methyl-2-Pentanone      | 8.574  | 43   | 156251   | 103.604 | ug/l   | 99       |
| 52) Toluene                   | 8.720  | 92   | 52127    | 19.848  | ug/l   | 98       |
| 53) t-1,3-Dichloropropene     | 8.982  | 75   | 27459    | 19.096  | ug/l   | 97       |
| 54) cis-1,3-Dichloropropene   | 8.366  | 75   | 30880    | 20.108  | ug/l   | 98       |
| 55) 1,1,2-Trichloroethane     | 9.153  | 97   | 20926    | 21.047  | ug/l   | 98       |
| 56) Ethyl methacrylate        | 9.116  | 69   | 31087    | 19.725  | ug/l   | 100      |
| 57) 1,3-Dichloropropane       | 9.311  | 76   | 35944    | 21.053  | ug/l   | 100      |
| 58) 2-Chloroethyl Vinyl ether | 8.244  | 63   | 78732    | 97.328  | ug/l   | 100      |
| 59) 2-Hexanone                | 9.433  | 43   | 113848   | 103.206 | ug/l   | 98       |
| 60) Dibromochloromethane      | 9.525  | 129  | 21397    | 20.915  | ug/l   | 100      |
| 61) 1,2-Dibromoethane         | 9.610  | 107  | 21775    | 20.728  | ug/l   | 100      |
| 64) Tetrachloroethene         | 9.275  | 164  | 24854    | 18.481  | ug/l   | 96       |
| 65) Chlorobenzene             | 10.079 | 112  | 54381    | 19.638  | ug/l   | 100      |
| 66) 1,1,1,2-Tetrachloroethane | 10.165 | 131  | 19913    | 19.548  | ug/l   | 99       |
| 67) Ethyl Benzene             | 10.195 | 91   | 97027    | 19.796  | ug/l   | 99       |
| 68) m/p-Xylenes               | 10.305 | 106  | 74494    | 39.504  | ug/l   | 99       |
| 69) o-Xylene                  | 10.646 | 106  | 36186    | 19.584  | ug/l   | 98       |
| 70) Styrene                   | 10.659 | 104  | 59996    | 20.020  | ug/l   | 98       |
| 71) Bromoform                 | 10.799 | 173  | 13596    | 18.452  | ug/l # | 99       |
| 73) Isopropylbenzene          | 10.963 | 105  | 95366    | 20.153  | ug/l   | 99       |
| 74) N-amyl acetate            | 10.841 | 43   | 37780    | 19.981  | ug/l   | 99       |
| 75) 1,1,2,2-Tetrachloroethane | 11.213 | 83   | 25412    | 19.788  | ug/l   | 100      |
| 76) 1,2,3-Trichloropropane    | 11.238 | 75   | 25632m   | 19.490  | ug/l   |          |
| 77) Bromobenzene              | 11.201 | 156  | 21948    | 19.512  | ug/l   | 98       |
| 78) n-propylbenzene           | 11.305 | 91   | 107805   | 20.550  | ug/l   | 100      |
| 79) 2-Chlorotoluene           | 11.366 | 91   | 65459    | 20.195  | ug/l   | 98       |
| 80) 1,3,5-Trimethylbenzene    | 11.451 | 105  | 79292    | 20.359  | ug/l   | 99       |
| 81) trans-1,4-Dichloro-2-b... | 11.024 | 75   | 6956     | 18.809  | ug/l   | 97       |
| 82) 4-Chlorotoluene           | 11.457 | 91   | 74758    | 20.559  | ug/l   | 99       |
| 83) tert-Butylbenzene         | 11.719 | 119  | 74033    | 19.944  | ug/l   | 99       |
| 84) 1,2,4-Trimethylbenzene    | 11.750 | 105  | 78682    | 20.029  | ug/l   | 98       |
| 85) sec-Butylbenzene          | 11.890 | 105  | 96282    | 20.462  | ug/l   | 99       |
| 86) p-Isopropyltoluene        | 12.012 | 119  | 80374    | 20.163  | ug/l   | 100      |
| 87) 1,3-Dichlorobenzene       | 11.969 | 146  | 40075    | 19.443  | ug/l   | 98       |
| 88) 1,4-Dichlorobenzene       | 12.042 | 146  | 40827    | 19.641  | ug/l   | 98       |
| 89) n-Butylbenzene            | 12.335 | 91   | 67253    | 20.621  | ug/l   | 100      |
| 90) Hexachloroethane          | 12.542 | 117  | 11252    | 19.530  | ug/l   | 97       |
| 91) 1,2-Dichlorobenzene       | 12.335 | 146  | 39333    | 19.777  | ug/l   | 98       |
| 92) 1,2-Dibromo-3-Chloropr... | 12.945 | 75   | 5719     | 18.109  | ug/l   | 98       |
| 93) 1,2,4-Trichlorobenzene    | 13.591 | 180  | 23999    | 19.121  | ug/l   | 100      |
| 94) Hexachlorobutadiene       | 13.725 | 225  | 9856     | 18.565  | ug/l   | 99       |
| 95) Naphthalene               | 13.780 | 128  | 84030    | 19.113  | ug/l   | 100      |
| 96) 1,2,3-Trichlorobenzene    | 13.963 | 180  | 24116    | 18.975  | ug/l   | 100      |

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX021022\  
Data File : VX026922.D  
Acq On : 10 Feb 2022 15:59  
Operator : JC/MD  
Sample : VX0210WBSD02  
Misc : 5.0mL/MSVOA\_X/WATER  
ALS Vial : 13 Sample Multiplier: 1

Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0210WBSD02

Manual Integrations  
APPROVED

Reviewed By :Semsettin Yesilyurt 02/11/2022  
Supervised By :Mahesh Dadoda 02/11/2022

Quant Time: Feb 11 03:59:44 2022  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X020822W.M  
Quant Title : SW846 8260  
QLast Update : Tue Feb 08 07:07:46 2022  
Response via : Initial Calibration

| Compound | R.T. | QIon | Response | Conc | Units | Dev(Min) |
|----------|------|------|----------|------|-------|----------|
|----------|------|------|----------|------|-------|----------|

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX021022\  
 Data File : VX026922.D  
 Acq On : 10 Feb 2022 15:59  
 Operator : JC/MD  
 Sample : VX0210WBSD02  
 Misc : 5.0mL/MSVOA\_X/WATER  
 ALS Vial : 13 Sample Multiplier: 1


Instrument :  
 MSVOA\_X  
 ClientSampleId :  
 VX0210WBSD02

Quant Time: Feb 11 03:59:44 2022  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X020822W.M  
 Quant Title : SW846 8260  
 QLast Update : Tue Feb 08 07:07:46 2022  
 Response via : Initial Calibration

### Manual Integrations APPROVED


Reviewed By : Semsettin Yesilyurt 02/11/2022  
 Supervised By : Mahesh Dadoda 02/11/2022





#1  
Pentafluorobenzene  
Concen: 50.000 ug/l  
RT: 5.556 min Scan# 71  
Delta R.T. -0.006 min  
Lab File: VX026922.D  
Acq: 10 Feb 2022 15:59




Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0210WBSD02




Tgt Ion:168 Resp: 9199  
Ion Ratio Lower Upper  
168 100  
99 62.6 50.3 75.5

Manual Integrations  
APPROVED



Reviewed By :Semsettin Yesilyurt 02/11/2022  
Supervised By :Mahesh Dadoda 02/11/2022




#2  
Dichlorodifluoromethane  
Concen: 20.357 ug/l  
RT: 1.172 min Scan# 14  
Delta R.T. 0.000 min  
Lab File: VX026922.D  
Acq: 10 Feb 2022 15:59




Tgt Ion: 85 Resp: 20363  
Ion Ratio Lower Upper  
85 100  
87 31.9 16.4 49.1





#3  
Chloromethane  
Concen: 20.071 ug/l  
RT: 1.294 min Scan# 34  
Delta R.T. 0.000 min  
Lab File: VX026922.D  
Acq: 10 Feb 2022 15:59




Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0210WBSD02




Tgt Ion: 50 Resp: 1867  
Ion Ratio Lower Upper  
50 100  
52 32.5 26.1 39.1

Manual Integrations  
APPROVED



Reviewed By :Semsettin Yesilyurt 02/11/2022  
Supervised By :Mahesh Dadoda 02/11/2022




#4  
Vinyl Chloride  
Concen: 19.711 ug/l  
RT: 1.374 min Scan# 47  
Delta R.T. -0.006 min  
Lab File: VX026922.D  
Acq: 10 Feb 2022 15:59




Tgt Ion: 62 Resp: 19201  
Ion Ratio Lower Upper  
62 100  
64 32.0 26.2 39.2





#5  
Bromomethane  
Concen: 19.799 ug/l  
RT: 1.617 min Scan# 81  
Delta R.T. 0.006 min  
Lab File: VX026922.D  
Acq: 10 Feb 2022 15:59




Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0210WBSD02



Tgt Ion: 94 Resp: 8184  
Ion Ratio Lower Upper  
94 100  
96 88.6 75.9 113.9




Manual Integrations  
APPROVED


Reviewed By :Semsettin Yesilyurt 02/11/2022  
Supervised By :Mahesh Dadoda 02/11/2022



#6  
Chloroethane  
Concen: 19.571 ug/l  
RT: 1.691 min Scan# 99  
Delta R.T. -0.000 min  
Lab File: VX026922.D  
Acq: 10 Feb 2022 15:59

Tgt Ion: 64 Resp: 11411  
Ion Ratio Lower Upper  
64 100  
66 31.5 25.4 38.2









#7  
Trichlorofluoromethane  
Concen: 19.704 ug/l  
RT: 1.892 min Scan# 11  
Delta R.T. 0.000 min  
Lab File: VX026922.D  
Acq: 10 Feb 2022 15:59

Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0210WBSD02

Tgt Ion:101 Resp: 28240  
Ion Ratio Lower Upper  
101 100  
103 62.4 51.3 76.9




Manual Integrations  
APPROVED


Reviewed By :Semsettin Yesilyurt 02/11/2022  
Supervised By :Mahesh Dadoda 02/11/2022



#8  
Diethyl Ether  
Concen: 19.289 ug/l  
RT: 2.136 min Scan# 172  
Delta R.T. 0.000 min  
Lab File: VX026922.D  
Acq: 10 Feb 2022 15:59

Tgt Ion: 74 Resp: 10243  
Ion Ratio Lower Upper  
74 100  
45 100.7 50.7 152.1









#9  
1,1,2-Trichlorotrifluoroethane  
Concen: 20.341 ug/l  
RT: 2.337 min Scan# 205  
Delta R.T. 0.000 min  
Lab File: VX026922.D  
Acq: 10 Feb 2022 15:59

Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0210WBSD02

Tgt Ion:101 Resp: 1911  
Ion Ratio Lower Upper  
101 100  
85 43.4 35.3 52.9  
151 75.8 61.4 92.2




Manual Integrations  
APPROVED


Reviewed By :Semsettin Yesilyurt 02/11/2022  
Supervised By :Mahesh Dadoda 02/11/2022



#10  
Methyl Iodide  
Concen: 17.613 ug/l  
RT: 2.459 min Scan# 225  
Delta R.T. -0.000 min  
Lab File: VX026922.D  
Acq: 10 Feb 2022 15:59

Tgt Ion:142 Resp: 20962  
Ion Ratio Lower Upper  
142 100  
127 46.6 38.1 57.1  
141 14.7 11.2 16.8





#11

Tert butyl alcohol

Concen: 88.560 ug/l

RT: 2.977 min Scan# 31

Delta R.T. -0.024 min

Lab File: VX026922.D

Acq: 10 Feb 2022 15:59

Instrument :

MSVOA\_X

ClientSampleId :

VX0210WBSD02

Tgt Ion: 59 Resp: 1766

Ion Ratio Lower Upper

59 100





57 10.6 7.2 10.8

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 02/11/2022

Supervised By :Mahesh Dadoda 02/11/2022



#12

1,1-Dichloroethene

Concen: 18.891 ug/l

RT: 2.325 min Scan# 203

Delta R.T. 0.000 min

Lab File: VX026922.D

Acq: 10 Feb 2022 15:59




Tgt Ion: 96 Resp: 17785


Ion Ratio Lower Upper

96 100

61 155.7 138.5 207.7


98 63.1 51.8 77.6





#13  
Acrolein  
Concen: 94.800 ug/l  
RT: 2.239 min Scan# 189  
Delta R.T. 0.000 min  
Lab File: VX026922.D  
Acq: 10 Feb 2022 15:59



Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0210WBSD02



Tgt Ion: 56 Resp: 1747  
Ion Ratio Lower Upper  
56 100  
55 71.5 56.9 85.3




Manual Integrations  
APPROVED


Reviewed By :Semsettin Yesilyurt 02/11/2022  
Supervised By :Mahesh Dadoda 02/11/2022



#14  
Allyl chloride  
Concen: 19.702 ug/l  
RT: 2.666 min Scan# 259  
Delta R.T. 0.000 min  
Lab File: VX026922.D  
Acq: 10 Feb 2022 15:59

Tgt Ion: 41 Resp: 32856  
Ion Ratio Lower Upper  
41 100  
39 64.7 54.2 81.4  
76 31.7 25.2 37.8





#15

Acrylonitrile

Concen: 102.960 ug/l

RT: 3.068 min Scan# 31

Delta R.T. 0.000 min

Lab File: VX026922.D

Acq: 10 Feb 2022 15:59

Instrument :

MSVOA\_X

ClientSampleId :





VX0210WBSD02

|             |             |
|-------------|-------------|
| Tgt Ion: 53 | Resp: 5885  |
| Ion Ratio   | Lower Upper |
| 53 100      |             |
| 52 81.1     | 66.6 100.0  |
| 51 35.4     | 29.5 44.3   |

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 02/11/2022  
Supervised By :Mahesh Dadoda 02/11/2022



#16

Acetone

Concen: 101.705 ug/l




RT: 2.386 min Scan# 213


Delta R.T. 0.000 min

Lab File: VX026922.D

Acq: 10 Feb 2022 15:59


|             |             |
|-------------|-------------|
| Tgt Ion: 43 | Resp: 49328 |
| Ion Ratio   | Lower Upper |
| 43 100      |             |
| 58 29.0     | 23.0 34.6   |





#17  
Carbon Disulfide  
Concen: 17.791 ug/l  
RT: 2.514 min Scan# 211  
Delta R.T. 0.000 min  
Lab File: VX026922.D  
Acq: 10 Feb 2022 15:59




Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0210WBSD02




Tgt Ion: 76 Resp: 47410  
Ion Ratio Lower Upper  
76 100  
78 9.0 7.5 11.3

Manual Integrations  
APPROVED



Reviewed By :Semsettin Yesilyurt 02/11/2022  
Supervised By :Mahesh Dadoda 02/11/2022




#18  
Methyl Acetate  
Concen: 19.004 ug/l  
RT: 2.709 min Scan# 266  
Delta R.T. 0.000 min  
Lab File: VX026922.D  
Acq: 10 Feb 2022 15:59



Tgt Ion: 43 Resp: 26520  
Ion Ratio Lower Upper  
43 100  
74 22.9 18.3 27.5









#19  
Methyl tert-butyl Ether  
Concen: 19.093 ug/l  
RT: 3.117 min Scan# 31  
Delta R.T. 0.000 min  
Lab File: VX026922.D  
Acq: 10 Feb 2022 15:59

Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0210WBSD02

Tgt Ion: 73 Resp: 63620  
Ion Ratio Lower Upper  
73 100  
57 21.6 17.4 26.2




Manual Integrations  
APPROVED


Reviewed By :Semsettin Yesilyurt 02/11/2022  
Supervised By :Mahesh Dadoda 02/11/2022



#20  
Methylene Chloride  
Concen: 19.463 ug/l  
RT: 2.794 min Scan# 280  
Delta R.T. 0.000 min  
Lab File: VX026922.D  
Acq: 10 Feb 2022 15:59


Tgt Ion: 84 Resp: 20454  
Ion Ratio Lower Upper  
84 100  
49 125.3 101.5 152.3  
51 36.3 31.2 46.8  
86 61.3 51.3 76.9





#21  
trans-1,2-Dichloroethene  
Concen: 19.732 ug/l  
RT: 3.093 min Scan# 31  
Delta R.T. -0.006 min  
Lab File: VX026922.D  
Acq: 10 Feb 2022 15:59




Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0210WBSD02



Tgt Ion: 96 Resp: 19549  
Ion Ratio Lower Upper  
96 100  
61 135.7 112.9 169.3  
98 60.7 49.5 74.3




Manual Integrations  
APPROVED


Reviewed By :Semsettin Yesilyurt 02/11/2022  
Supervised By :Mahesh Dadoda 02/11/2022



#22  
Diisopropyl ether  
Concen: 19.911 ug/l  
RT: 3.763 min Scan# 439  
Delta R.T. 0.000 min  
Lab File: VX026922.D  
Acq: 10 Feb 2022 15:59


Tgt Ion: 45 Resp: 66852  
Ion Ratio Lower Upper  
45 100  
43 75.1 60.6 91.0  
87 26.5 22.0 33.0  
59 12.1 9.0 13.4





#23  
 Vinyl Acetate  
 Concen: 100.642 ug/l  
 RT: 3.727 min Scan# 41  
 Delta R.T. 0.000 min  
 Lab File: VX026922.D  
 Acq: 10 Feb 2022 15:59




Instrument :  
 MSVOA\_X  
 ClientSampleId :  
 VX0210WBSD02




Tgt Ion: 43 Resp: 28410  
 Ion Ratio Lower Upper  
 43 100  
 86 11.0 8.1 12.1

Manual Integrations  
 APPROVED



Reviewed By :Semsettin Yesilyurt 02/11/2022  
 Supervised By :Mahesh Dadoda 02/11/2022




#24  
 1,1-Dichloroethane  
 Concen: 19.615 ug/l  
 RT: 3.611 min Scan# 414  
 Delta R.T. -0.006 min  
 Lab File: VX026922.D  
 Acq: 10 Feb 2022 15:59




Tgt Ion: 63 Resp: 35636  
 Ion Ratio Lower Upper  
 63 100  
 98 6.8 3.7 11.1  
 100 4.1 2.4 7.2





#25  
2-Butanone  
Concen: 101.437 ug/l  
RT: 4.562 min Scan# 51  
Delta R.T. 0.000 min  
Lab File: VX026922.D  
Acq: 10 Feb 2022 15:59




Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0210WBSD02



Tgt Ion: 43 Resp: 7926  
Ion Ratio Lower Upper  
43 100  
72 26.0 20.4 30.6




Manual Integrations  
APPROVED


Reviewed By :Semsettin Yesilyurt 02/11/2022  
Supervised By :Mahesh Dadoda 02/11/2022



#26  
2,2-Dichloropropane  
Concen: 20.226 ug/l  
RT: 4.477 min Scan# 556  
Delta R.T. -0.006 min  
Lab File: VX026922.D  
Acq: 10 Feb 2022 15:59


Tgt Ion: 77 Resp: 24828  
Ion Ratio Lower Upper  
77 100  
97 23.0 11.0 33.0





#27  
 cis-1,2-Dichloroethene  
 Concen: 20.075 ug/l  
 RT: 4.495 min Scan# 511  
 Delta R.T. 0.000 min  
 Lab File: VX026922.D  
 Acq: 10 Feb 2022 15:59




Instrument :  
 MSVOA\_X  
 ClientSampleId :  
 VX0210WBSD02




Tgt Ion: 96 Resp: 2283  
 Ion Ratio Lower Upper  
 96 100  
 61 141.1 0.0 296.6  
 98 64.3 0.0 128.8

Manual Integrations  
 APPROVED



Reviewed By :Semsettin Yesilyurt 02/11/2022  
 Supervised By :Mahesh Dadoda 02/11/2022




#28  
 Bromochloromethane  
 Concen: 21.269 ug/l  
 RT: 4.903 min Scan# 626  
 Delta R.T. 0.000 min  
 Lab File: VX026922.D  
 Acq: 10 Feb 2022 15:59



Tgt Ion: 49 Resp: 16018  
 Ion Ratio Lower Upper  
 49 100  
 129 1.4 0.0 4.2  
 130 80.1 62.4 93.6





#29

Tetrahydrofuran

Concen: 99.834 ug/l

RT: 5.013 min Scan# 64

Delta R.T. 0.006 min

Lab File: VX026922.D

Acq: 10 Feb 2022 15:59

Instrument :

MSVOA\_X

ClientSampleId :





VX0210WBSD02

|             |             |
|-------------|-------------|
| Tgt Ion: 42 | Resp: 5303  |
| Ion Ratio   | Lower Upper |
| 42 100      |             |
| 72 44.2     | 34.6 52.0   |
| 71 40.3     | 31.8 47.8   |

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 02/11/2022  
Supervised By :Mahesh Dadoda 02/11/2022



#30

Chloroform

Concen: 19.544 ug/l




RT: 5.099 min Scan# 658


Delta R.T. 0.000 min

Lab File: VX026922.D

Acq: 10 Feb 2022 15:59

|             |             |
|-------------|-------------|
| Tgt Ion: 83 | Resp: 36891 |
| Ion Ratio   | Lower Upper |
| 83 100      |             |
| 85 63.0     | 51.6 77.4   |









#31  
Cyclohexane  
Concen: 19.445 ug/l  
RT: 5.477 min Scan# 705  
Delta R.T. 0.000 min  
Lab File: VX026922.D  
Acq: 10 Feb 2022 15:59

Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0210WBSD02

Tgt Ion: 56 Resp: 34119  
Ion Ratio Lower Upper  
56 100  
69 30.8 25.8 38.6  
84 84.4 68.9 103.3




Manual Integrations  
APPROVED


Reviewed By :Semsettin Yesilyurt 02/11/2022  
Supervised By :Mahesh Dadoda 02/11/2022



#32  
1,1,1-Trichloroethane  
Concen: 19.292 ug/l  
RT: 5.385 min Scan# 705  
Delta R.T. -0.006 min  
Lab File: VX026922.D  
Acq: 10 Feb 2022 15:59

Tgt Ion: 97 Resp: 32107  
Ion Ratio Lower Upper  
97 100  
99 63.7 51.3 76.9  
61 44.2 35.0 52.6





#33

1,2-Dichloroethane-d4

Concen: 54.350 ug/l

RT: 5.958 min Scan# 799

Delta R.T. -0.006 min

Lab File: VX026922.D

Acq: 10 Feb 2022 15:59

Instrument :

MSVOA\_X

ClientSampleId :

VX0210WBSD02

Tgt Ion: 65 Resp: 60018

Ion Ratio Lower Upper

65 100





67 51.5 0.0 100.8

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 02/11/2022

Supervised By :Mahesh Dadoda 02/11/2022



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 6.763 min Scan# 931

Delta R.T. -0.006 min

Lab File: VX026922.D

Acq: 10 Feb 2022 15:59




Tgt Ion:114 Resp: 151434


Ion Ratio Lower Upper

114 100

63 20.9 0.0 42.8


88 16.5 0.0 33.0





#35  
Dibromofluoromethane  
Concen: 51.738 ug/l  
RT: 5.391 min Scan# 706  
Delta R.T. 0.000 min  
Lab File: VX026922.D  
Acq: 10 Feb 2022 15:59




Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0210WBSD02



Tgt Ion: 113 Resp: 46679  
Ion Ratio Lower Upper  
113 100  
111 102.7 81.2 121.8  
192 19.2 15.0 22.4




Manual Integrations  
APPROVED


Reviewed By :Semsettin Yesilyurt 02/11/2022  
Supervised By :Mahesh Dadoda 02/11/2022



#36  
1,1-Dichloropropene  
Concen: 19.746 ug/l  
RT: 5.702 min Scan# 757  
Delta R.T. 0.006 min  
Lab File: VX026922.D  
Acq: 10 Feb 2022 15:59

Tgt Ion: 75 Resp: 28104  
Ion Ratio Lower Upper  
75 100  
110 34.9 17.2 51.6  
77 30.3 24.6 36.8





#37

Ethyl Acetate

Concen: 19.414 ug/l

RT: 4.714 min Scan# 595

Delta R.T. 0.000 min

Lab File: VX026922.D

Acq: 10 Feb 2022 15:59

Instrument :

MSVOA\_X

ClientSampleId :





VX0210WBSD02

|             |             |
|-------------|-------------|
| Tgt Ion: 43 | Resp: 3140  |
| Ion Ratio   | Lower Upper |
| 43 100      |             |
| 61 13.7     | 10.6 16.0   |
| 70 11.6     | 8.5 12.7    |

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 02/11/2022  
Supervised By :Mahesh Dadoda 02/11/2022



#38

Carbon Tetrachloride

Concen: 18.617 ug/l




RT: 5.684 min Scan# 754


Delta R.T. 0.000 min

Lab File: VX026922.D

Acq: 10 Feb 2022 15:59

|             |             |
|-------------|-------------|
| Tgt Ion:117 | Resp: 27786 |
| Ion Ratio   | Lower Upper |
| 117 100     |             |
| 119 96.2    | 74.7 112.1  |
| 121 31.0    | 25.4 38.0   |





#39

Methylcyclohexane

Concen: 19.383 ug/l

RT: 7.385 min Scan# 1033

Delta R.T. 0.000 min

Lab File: VX026922.D

Acq: 10 Feb 2022 15:59

Instrument :

MSVOA\_X

ClientSampleId :





VX0210WBSD02

|             |             |
|-------------|-------------|
| Tgt Ion: 83 | Resp: 3383  |
| Ion Ratio   | Lower Upper |
| 83 100      |             |
| 55 73.9     | 63.6 95.4   |
| 98 48.5     | 38.4 57.6   |

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 02/11/2022  
Supervised By :Mahesh Dadoda 02/11/2022



#40

Benzene

Concen: 19.494 ug/l




RT: 6.043 min Scan# 813


Delta R.T. 0.000 min

Lab File: VX026922.D

Acq: 10 Feb 2022 15:59


|             |             |
|-------------|-------------|
| Tgt Ion: 78 | Resp: 80862 |
| Ion Ratio   | Lower Upper |
| 78 100      |             |
| 77 22.6     | 18.9 28.3   |





#41  
Methacrylonitrile  
Concen: 19.839 ug/l  
RT: 4.922 min Scan# 61  
Delta R.T. 0.000 min  
Lab File: VX026922.D  
Acq: 10 Feb 2022 15:59




Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0210WBSD02




Tgt Ion: 41 Resp: 17392  
Ion Ratio Lower Upper  
41 100  
39 54.8 45.4 68.0  
67 70.0 53.2 79.8  
52 27.9 23.0 34.6

Manual Integrations  
APPROVED



Reviewed By :Semsettin Yesilyurt 02/11/2022  
Supervised By :Mahesh Dadoda 02/11/2022




#42  
1,2-Dichloroethane  
Concen: 20.573 ug/l  
RT: 6.092 min Scan# 821  
Delta R.T. 0.000 min  
Lab File: VX026922.D  
Acq: 10 Feb 2022 15:59




Tgt Ion: 62 Resp: 30072  
Ion Ratio Lower Upper  
62 100  
98 9.2 0.0 17.8





#43  
Isopropyl Acetate  
Concen: 19.524 ug/l  
RT: 6.342 min Scan# 862  
Delta R.T. 0.000 min  
Lab File: VX026922.D  
Acq: 10 Feb 2022 15:59




Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0210WBSD02




Tgt Ion: 43 Resp: 4859  
Ion Ratio Lower Upper  
43 100  
61 19.6 15.4 23.0  
87 13.1 10.2 15.2

Manual Integrations  
APPROVED



Reviewed By :Semsettin Yesilyurt 02/11/2022  
Supervised By :Mahesh Dadoda 02/11/2022



#44  
Trichloroethene  
Concen: 18.432 ug/l  
RT: 7.129 min Scan# 991  
Delta R.T. 0.000 min  
Lab File: VX026922.D  
Acq: 10 Feb 2022 15:59



Tgt Ion:130 Resp: 22957  
Ion Ratio Lower Upper  
130 100  
95 102.9 0.0 202.6





#45

1,2-Dichloropropane

Concen: 19.844 ug/l

RT: 7.433 min Scan# 1041

Delta R.T. 0.000 min

Lab File: VX026922.D

Acq: 10 Feb 2022 15:59

Instrument :

MSVOA\_X

ClientSampleId :

VX0210WBSD02

Tgt Ion: 63 Resp: 2082

Ion Ratio Lower Upper

63 100





65 29.2 23.8 35.6

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 02/11/2022

Supervised By :Mahesh Dadoda 02/11/2022



#46

Dibromomethane

Concen: 20.161 ug/l

RT: 7.586 min Scan# 1066

Delta R.T. 0.000 min

Lab File: VX026922.D

Acq: 10 Feb 2022 15:59




Tgt Ion: 93 Resp: 14641


Ion Ratio Lower Upper

93 100

95 81.6 66.1 99.1

174 100.6 78.7 118.1









#47  
Bromodichloromethane  
Concen: 20.756 ug/l  
RT: 7.824 min Scan# 1105  
Delta R.T. 0.000 min  
Lab File: VX026922.D  
Acq: 10 Feb 2022 15:59

Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0210WBSD02

Tgt Ion: 83 Resp: 28860  
Ion Ratio Lower Upper  
83 100  
85 63.1 51.4 77.0  
127 9.3 6.9 10.3




Manual Integrations  
APPROVED


Reviewed By :Semsettin Yesilyurt 02/11/2022  
Supervised By :Mahesh Dadoda 02/11/2022



#48  
Methyl methacrylate  
Concen: 19.451 ug/l  
RT: 7.696 min Scan# 1084  
Delta R.T. 0.000 min  
Lab File: VX026922.D  
Acq: 10 Feb 2022 15:59

Tgt Ion: 41 Resp: 23986  
Ion Ratio Lower Upper  
41 100  
69 80.9 62.2 93.4  
39 55.4 45.4 68.0





#49

1,4-Dioxane

Concen: 398.501 ug/l

RT: 7.677 min Scan# 1081

Delta R.T. 0.000 min

Lab File: VX026922.D

Acq: 10 Feb 2022 15:59

Instrument :

MSVOA\_X

ClientSampleId :

VX0210WBSD02





Tgt Ion: 88 Resp: 846

| Ion | Ratio | Lower | Upper |
|-----|-------|-------|-------|
| 88  | 100   |       |       |
| 43  | 35.5  | 29.9  | 44.9  |
| 58  | 73.1  | 57.8  | 86.6  |

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 02/11/2022  
Supervised By :Mahesh Dadoda 02/11/2022



#50

Toluene-d8

Concen: 53.411 ug/l

RT: 8.653 min Scan# 1241




Delta R.T. 0.000 min


Lab File: VX026922.D

Acq: 10 Feb 2022 15:59

Tgt Ion: 98 Resp: 176884


| Ion | Ratio | Lower | Upper |
|-----|-------|-------|-------|
| 98  | 100   |       |       |
| 100 | 66.2  | 53.2  | 79.8  |





#51  
4-Methyl-2-Pentanone  
Concen: 103.604 ug/l  
RT: 8.574 min Scan# 1228  
Delta R.T. 0.000 min  
Lab File: VX026922.D  
Acq: 10 Feb 2022 15:59




Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0210WBSD02




Tgt Ion: 43 Resp: 15625  
Ion Ratio Lower Upper  
43 100  
58 38.9 30.5 45.7

Manual Integrations  
APPROVED



Reviewed By :Semsettin Yesilyurt 02/11/2022  
Supervised By :Mahesh Dadoda 02/11/2022



#52  
Toluene  
Concen: 19.848 ug/l  
RT: 8.720 min Scan# 1252  
Delta R.T. 0.000 min  
Lab File: VX026922.D  
Acq: 10 Feb 2022 15:59



Tgt Ion: 92 Resp: 52127  
Ion Ratio Lower Upper  
92 100  
91 166.4 135.7 203.5





#53

t-1,3-Dichloropropene

Concen: 19.096 ug/l

RT: 8.982 min Scan# 1194

Delta R.T. -0.000 min

Lab File: VX026922.D

Acq: 10 Feb 2022 15:59

Instrument :

MSVOA\_X

ClientSampleId :

VX0210WBSD02

Tgt Ion: 75 Resp: 2745

Ion Ratio Lower Upper

75 100





77 30.5 25.7 38.5

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 02/11/2022

Supervised By :Mahesh Dadoda 02/11/2022



#54

cis-1,3-Dichloropropene

Concen: 20.108 ug/l

RT: 8.366 min Scan# 1194

Delta R.T. 0.000 min

Lab File: VX026922.D

Acq: 10 Feb 2022 15:59




Tgt Ion: 75 Resp: 30880


Ion Ratio Lower Upper

75 100

77 31.3 26.2 39.2


39 52.7 43.5 65.3





#55  
1,1,2-Trichloroethane  
Concen: 21.047 ug/l  
RT: 9.153 min Scan# 1323  
Delta R.T. 0.000 min  
Lab File: VX026922.D  
Acq: 10 Feb 2022 15:59

Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0210WBSD02






Tgt Ion: 97 Resp: 20920


| Ion | Ratio | Lower | Upper |
|-----|-------|-------|-------|
| 97  | 100   |       |       |
| 83  | 82.7  | 68.4  | 102.6 |
| 85  | 53.5  | 44.0  | 66.0  |
| 99  | 61.6  | 49.8  | 74.6  |

Manual Integrations  
APPROVED

Reviewed By :Semsettin Yesilyurt 02/11/2022  
Supervised By :Mahesh Dadoda 02/11/2022






#56  
Ethyl methacrylate  
Concen: 19.725 ug/l  
RT: 9.116 min Scan# 1317  
Delta R.T. 0.000 min  
Lab File: VX026922.D  
Acq: 10 Feb 2022 15:59



Tgt Ion: 69 Resp: 31087

| Ion | Ratio | Lower | Upper |
|-----|-------|-------|-------|
| 69  | 100   |       |       |
| 41  | 71.4  | 57.2  | 85.8  |
| 39  | 38.5  | 30.9  | 46.3  |





#57

1,3-Dichloropropane

Concen: 21.053 ug/l

RT: 9.311 min Scan# 11

Delta R.T. 0.000 min

Lab File: VX026922.D

Acq: 10 Feb 2022 15:59

Instrument :

MSVOA\_X

ClientSampleId :

VX0210WBSD02

Tgt Ion: 76 Resp: 3594

Ion Ratio Lower Upper

76 100





78 32.2 25.7 38.5

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 02/11/2022

Supervised By :Mahesh Dadoda 02/11/2022



#58

2-Chloroethyl Vinyl ether

Concen: 97.328 ug/l

RT: 8.244 min Scan# 1174

Delta R.T. 0.000 min

Lab File: VX026922.D




Acq: 10 Feb 2022 15:59


Tgt Ion: 63 Resp: 78732

Ion Ratio Lower Upper

63 100

106 28.9 23.0 34.6









#59  
2-Hexanone  
Concen: 103.206 ug/l  
RT: 9.433 min Scan# 11384  
Delta R.T. 0.000 min  
Lab File: VX026922.D  
Acq: 10 Feb 2022 15:59

Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0210WBSD02

Tgt Ion: 43 Resp: 11384  
Ion Ratio Lower Upper  
43 100  
58 54.4 26.4 79.2




Manual Integrations  
APPROVED


Reviewed By :Semsettin Yesilyurt 02/11/2022  
Supervised By :Mahesh Dadoda 02/11/2022



#60  
Dibromochloromethane  
Concen: 20.915 ug/l  
RT: 9.525 min Scan# 1384  
Delta R.T. 0.000 min  
Lab File: VX026922.D  
Acq: 10 Feb 2022 15:59


Tgt Ion:129 Resp: 21397  
Ion Ratio Lower Upper  
129 100  
127 77.4 38.5 115.5





#61  
1,2-Dibromoethane  
Concen: 20.728 ug/l  
RT: 9.610 min Scan# 11  
Delta R.T. 0.000 min  
Lab File: VX026922.D  
Acq: 10 Feb 2022 15:59




Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0210WBSD02




Tgt Ion:107 Resp: 21775  
Ion Ratio Lower Upper  
107 100  
109 94.8 75.5 113.3

Manual Integrations  
APPROVED



Reviewed By :Semsettin Yesilyurt 02/11/2022  
Supervised By :Mahesh Dadoda 02/11/2022




#62  
4-Bromofluorobenzene  
Concen: 54.050 ug/l  
RT: 11.085 min Scan# 1640  
Delta R.T. 0.006 min  
Lab File: VX026922.D  
Acq: 10 Feb 2022 15:59




Tgt Ion: 95 Resp: 63020  
Ion Ratio Lower Upper  
95 100  
174 76.7 0.0 148.6  
176 72.6 0.0 141.4





#63  
Chlorobenzene-d5  
Concen: 50.000 ug/l  
RT: 10.055 min Scan# 1471  
Delta R.T. -0.000 min  
Lab File: VX026922.D  
Acq: 10 Feb 2022 15:59




Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0210WBSD02



Tgt Ion:117 Resp: 133902  
Ion Ratio Lower Upper  
117 100  
82 56.2 45.2 67.8  
119 32.4 26.2 39.2




Manual Integrations  
APPROVED


Reviewed By :Semsettin Yesilyurt 02/11/2022  
Supervised By :Mahesh Dadoda 02/11/2022



#64  
Tetrachloroethene  
Concen: 18.481 ug/l  
RT: 9.275 min Scan# 1343  
Delta R.T. 0.000 min  
Lab File: VX026922.D  
Acq: 10 Feb 2022 15:59


Tgt Ion:164 Resp: 24854  
Ion Ratio Lower Upper  
164 100  
166 126.5 101.8 152.6  
129 97.2 83.3 124.9  
131 90.7 78.0 117.0





#65  
Chlorobenzene  
Concen: 19.638 ug/l  
RT: 10.079 min Scan# 1475  
Delta R.T. 0.000 min  
Lab File: VX026922.D  
Acq: 10 Feb 2022 15:59




Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0210WBSD02




Tgt Ion:112 Resp: 5438  
Ion Ratio Lower Upper  
112 100  
114 31.7 25.3 37.9

Manual Integrations  
APPROVED



Reviewed By :Semsettin Yesilyurt 02/11/2022  
Supervised By :Mahesh Dadoda 02/11/2022




#66  
1,1,1,2-Tetrachloroethane  
Concen: 19.548 ug/l  
RT: 10.165 min Scan# 1489  
Delta R.T. 0.000 min  
Lab File: VX026922.D  
Acq: 10 Feb 2022 15:59



Tgt Ion:131 Resp: 19913  
Ion Ratio Lower Upper  
131 100  
133 96.5 47.9 143.8  
119 64.5 31.7 95.1









#67  
Ethyl Benzene  
Concen: 19.796 ug/l  
RT: 10.195 min Scan# 1494  
Delta R.T. 0.000 min  
Lab File: VX026922.D  
Acq: 10 Feb 2022 15:59

Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0210WBSD02

Tgt Ion: 91 Resp: 9702  
Ion Ratio Lower Upper  
91 100  
106 31.6 24.7 37.1




Manual Integrations  
APPROVED


Reviewed By :Semsettin Yesilyurt 02/11/2022  
Supervised By :Mahesh Dadoda 02/11/2022



#68  
m/p-Xylenes  
Concen: 39.504 ug/l  
RT: 10.305 min Scan# 1512  
Delta R.T. 0.000 min  
Lab File: VX026922.D  
Acq: 10 Feb 2022 15:59

Tgt Ion:106 Resp: 74494  
Ion Ratio Lower Upper  
106 100  
91 204.9 165.3 247.9









#69  
o-Xylene  
Concen: 19.584 ug/l  
RT: 10.646 min Scan# 1567  
Delta R.T. 0.006 min  
Lab File: VX026922.D  
Acq: 10 Feb 2022 15:59

Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0210WBSD02

Tgt Ion:106 Resp: 36180  
Ion Ratio Lower Upper  
106 100  
91 215.6 109.7 329.0




Manual Integrations  
APPROVED


Reviewed By :Semsettin Yesilyurt 02/11/2022  
Supervised By :Mahesh Dadoda 02/11/2022



#70  
Styrene  
Concen: 20.020 ug/l  
RT: 10.659 min Scan# 1570  
Delta R.T. 0.000 min  
Lab File: VX026922.D  
Acq: 10 Feb 2022 15:59


Tgt Ion:104 Resp: 59996  
Ion Ratio Lower Upper  
104 100  
78 51.2 42.5 63.7  
103 54.7 43.3 64.9





#71  
Bromoform  
Concen: 18.452 ug/l  
RT: 10.799 min Scan# 1593  
Delta R.T. -0.000 min  
Lab File: VX026922.D  
Acq: 10 Feb 2022 15:59




Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0210WBSD02




Tgt Ion:173 Resp: 13590  
Ion Ratio Lower Upper  
173 100  
175 49.1 24.3 72.9  
254 0.0 0.0 0.0

Manual Integrations  
APPROVED



Reviewed By :Semsettin Yesilyurt 02/11/2022  
Supervised By :Mahesh Dadoda 02/11/2022




#72  
1,4-Dichlorobenzene-d4  
Concen: 50.000 ug/l  
RT: 12.024 min Scan# 1794  
Delta R.T. 0.000 min  
Lab File: VX026922.D  
Acq: 10 Feb 2022 15:59




Tgt Ion:152 Resp: 60077  
Ion Ratio Lower Upper  
152 100  
115 70.8 44.2 132.6  
150 163.9 0.0 348.2





#73  
Isopropylbenzene  
Concen: 20.153 ug/l  
RT: 10.963 min Scan# 1620  
Delta R.T. 0.000 min  
Lab File: VX026922.D  
Acq: 10 Feb 2022 15:59




Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0210WBSD02




Tgt Ion:105 Resp: 95360  
Ion Ratio Lower Upper  
105 100  
120 26.4 13.1 39.1

Manual Integrations  
APPROVED



Reviewed By :Semsettin Yesilyurt 02/11/2022  
Supervised By :Mahesh Dadoda 02/11/2022




#74  
N-ethyl acetate  
Concen: 19.981 ug/l  
RT: 10.841 min Scan# 1600  
Delta R.T. -0.006 min  
Lab File: VX026922.D  
Acq: 10 Feb 2022 15:59



Tgt Ion: 43 Resp: 37780  
Ion Ratio Lower Upper  
43 100  
70 42.4 32.9 49.3  
55 23.8 18.3 27.5  
61 23.5 18.9 28.3









#75  
1,1,2,2-Tetrachloroethane  
Concen: 19.788 ug/l  
RT: 11.213 min Scan# 1661  
Delta R.T. 0.000 min  
Lab File: VX026922.D  
Acq: 10 Feb 2022 15:59

Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0210WBSD02

Tgt Ion: 83 Resp: 2541  
Ion Ratio Lower Upper  
83 100  
131 10.9 5.3 15.9  
85 64.4 32.0 96.2




Manual Integrations  
APPROVED


Reviewed By :Semsettin Yesilyurt 02/11/2022  
Supervised By :Mahesh Dadoda 02/11/2022



#76  
1,2,3-Trichloropropane  
Concen: 19.490 ug/l m  
RT: 11.238 min Scan# 1665  
Delta R.T. -0.006 min  
Lab File: VX026922.D  
Acq: 10 Feb 2022 15:59

Tgt Ion: 75 Resp: 25632  
Ion Ratio Lower Upper  
75 100  
77 37.1 18.3 54.8





#77

Bromobenzene

Concen: 19.512 ug/l

RT: 11.201 min Scan# 1659

Delta R.T. 0.000 min

Lab File: VX026922.D


Acq: 10 Feb 2022 15:59

Instrument :

MSVOA\_X

ClientSampleId :

VX0210WBSD02






| Tgt Ion:  | 156   | Resp: | 2194  |
|-----------|-------|-------|-------|
| Ion Ratio | 100   | Lower | Upper |
| 156       | 100   |       |       |
| 77        | 166.6 | 84.5  | 253.4 |
| 158       | 97.6  | 48.1  | 144.4 |

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 02/11/2022  
Supervised By :Mahesh Dadoda 02/11/2022



#78

n-propylbenzene


Concen: 20.550 ug/l

RT: 11.305 min Scan# 1676



Delta R.T. 0.000 min


Lab File: VX026922.D

Acq: 10 Feb 2022 15:59



| Tgt Ion:  | 91   | Resp: | 107805 |
|-----------|------|-------|--------|
| Ion Ratio | 100  | Lower | Upper  |
| 91        | 100  |       |        |
| 120       | 23.2 | 11.6  | 34.7   |





#79

2-Chlorotoluene

Concen: 20.195 ug/l

RT: 11.366 min Scan# 1

Delta R.T. 0.000 min

Lab File: VX026922.D

Acq: 10 Feb 2022 15:59

Instrument :

MSVOA\_X

ClientSampleId :

VX0210WBSD02

Tgt Ion: 91 Resp: 65459

Ion Ratio Lower Upper

91 100





126 33.8 16.4 49.2

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 02/11/2022

Supervised By :Mahesh Dadoda 02/11/2022



#80

1,3,5-Trimethylbenzene

Concen: 20.359 ug/l

RT: 11.451 min Scan# 1700

Delta R.T. 0.000 min

Lab File: VX026922.D




Acq: 10 Feb 2022 15:59


Tgt Ion:105 Resp: 79292

Ion Ratio Lower Upper

105 100


120 48.7 24.0 72.0





#81  
trans-1,4-Dichloro-2-butene  
Concen: 18.809 ug/l  
RT: 11.024 min Scan# 1629  
Delta R.T. 0.006 min  
Lab File: VX026922.D  
Acq: 10 Feb 2022 15:59




Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0210WBSD02




Tgt Ion: 75 Resp: 6950  
Ion Ratio Lower Upper  
75 100  
53 114.6 94.0 141.0  
89 48.2 37.0 55.4

Manual Integrations  
APPROVED



Reviewed By :Semsettin Yesilyurt 02/11/2022  
Supervised By :Mahesh Dadoda 02/11/2022




#82  
4-Chlorotoluene  
Concen: 20.559 ug/l  
RT: 11.457 min Scan# 1701  
Delta R.T. 0.000 min  
Lab File: VX026922.D  
Acq: 10 Feb 2022 15:59



Tgt Ion: 91 Resp: 74758  
Ion Ratio Lower Upper  
91 100  
126 28.9 14.1 42.4









#83  
tert-Butylbenzene  
Concen: 19.944 ug/l  
RT: 11.719 min Scan# 1743  
Delta R.T. 0.006 min  
Lab File: VX026922.D  
Acq: 10 Feb 2022 15:59

Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0210WBSD02

Tgt Ion:119 Resp: 7403  
Ion Ratio Lower Upper  
119 100  
91 62.0 31.4 94.0  
134 23.9 11.8 35.3




Manual Integrations  
APPROVED


Reviewed By :Semsettin Yesilyurt 02/11/2022  
Supervised By :Mahesh Dadoda 02/11/2022



#84  
1,2,4-Trimethylbenzene  
Concen: 20.029 ug/l  
RT: 11.750 min Scan# 1749  
Delta R.T. -0.000 min  
Lab File: VX026922.D  
Acq: 10 Feb 2022 15:59


Tgt Ion:105 Resp: 78682  
Ion Ratio Lower Upper  
105 100  
120 45.2 21.9 65.5





#85  
 sec-Butylbenzene  
 Concen: 20.462 ug/l  
 RT: 11.890 min Scan# 1772  
 Delta R.T. 0.000 min  
 Lab File: VX026922.D  
 Acq: 10 Feb 2022 15:59




Instrument :  
 MSVOA\_X  
 ClientSampleId :  
 VX0210WBSD02




Tgt Ion:105 Resp: 9628  
 Ion Ratio Lower Upper  
 105 100  
 134 19.9 10.1 30.1

Manual Integrations  
 APPROVED



Reviewed By :Semsettin Yesilyurt 02/11/2022  
 Supervised By :Mahesh Dadoda 02/11/2022




#86  
 p-Isopropyltoluene  
 Concen: 20.163 ug/l  
 RT: 12.012 min Scan# 1792  
 Delta R.T. 0.000 min  
 Lab File: VX026922.D  
 Acq: 10 Feb 2022 15:59



Tgt Ion:119 Resp: 80374  
 Ion Ratio Lower Upper  
 119 100  
 134 25.9 13.0 39.0  
 91 24.8 12.6 37.6





#87

1,3-Dichlorobenzene

Concen: 19.443 ug/l

RT: 11.969 min Scan# 1785

Delta R.T. 0.000 min

Lab File: VX026922.D

Acq: 10 Feb 2022 15:59

Instrument :

MSVOA\_X

ClientSampleId :

VX0210WBSD02





Tgt Ion:146 Resp: 40075

| Ion | Ratio | Lower | Upper |
|-----|-------|-------|-------|
| 146 | 100   |       |       |
| 111 | 41.7  | 21.3  | 63.7  |
| 148 | 64.6  | 31.6  | 95.0  |

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 02/11/2022  
Supervised By :Mahesh Dadoda 02/11/2022



#88

1,4-Dichlorobenzene

Concen: 19.641 ug/l

RT: 12.042 min Scan# 1797




Delta R.T. 0.000 min


Lab File: VX026922.D

Acq: 10 Feb 2022 15:59

Tgt Ion:146 Resp: 40827


| Ion | Ratio | Lower | Upper |
|-----|-------|-------|-------|
| 146 | 100   |       |       |
| 111 | 40.8  | 20.2  | 60.6  |
| 148 | 64.8  | 31.4  | 94.0  |





#89  
n-Butylbenzene  
Concen: 20.621 ug/l  
RT: 12.335 min Scan# 1845  
Delta R.T. 0.000 min  
Lab File: VX026922.D  
Acq: 10 Feb 2022 15:59




Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0210WBSD02



Tgt Ion: 91 Resp: 6725  
Ion Ratio Lower Upper  
91 100  
92 54.0 27.2 81.6  
134 26.2 13.1 39.1




Manual Integrations  
APPROVED


Reviewed By :Semsettin Yesilyurt 02/11/2022  
Supervised By :Mahesh Dadoda 02/11/2022



#90  
Hexachloroethane  
Concen: 19.530 ug/l  
RT: 12.542 min Scan# 1879  
Delta R.T. 0.000 min  
Lab File: VX026922.D  
Acq: 10 Feb 2022 15:59


Tgt Ion:117 Resp: 11252  
Ion Ratio Lower Upper  
117 100  
201 74.1 35.6 106.8





#91  
1,2-Dichlorobenzene  
Concen: 19.777 ug/l  
RT: 12.335 min Scan# 1845  
Delta R.T. 0.000 min  
Lab File: VX026922.D  
Acq: 10 Feb 2022 15:59




Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0210WBSD02




Tgt Ion:146 Resp: 3933  
Ion Ratio Lower Upper  
146 100  
111 42.6 21.8 65.3  
148 62.4 32.0 96.2

Manual Integrations  
APPROVED



Reviewed By :Semsettin Yesilyurt 02/11/2022  
Supervised By :Mahesh Dadoda 02/11/2022




#92  
1,2-Dibromo-3-Chloropropane  
Concen: 18.109 ug/l  
RT: 12.945 min Scan# 1945  
Delta R.T. 0.000 min  
Lab File: VX026922.D  
Acq: 10 Feb 2022 15:59



Tgt Ion: 75 Resp: 5719  
Ion Ratio Lower Upper  
75 100  
155 81.3 40.0 119.9  
157 105.5 51.6 154.8









#93  
1,2,4-Trichlorobenzene  
Concen: 19.121 ug/l  
RT: 13.591 min Scan# 2051  
Delta R.T. 0.000 min  
Lab File: VX026922.D  
Acq: 10 Feb 2022 15:59

Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0210WBSD02

Tgt Ion:180 Resp: 23999  
Ion Ratio Lower Upper  
180 100  
182 93.9 46.7 140.0  
145 32.5 16.3 48.8




Manual Integrations  
APPROVED


Reviewed By :Semsettin Yesilyurt 02/11/2022  
Supervised By :Mahesh Dadoda 02/11/2022



#94  
Hexachlorobutadiene  
Concen: 18.565 ug/l  
RT: 13.725 min Scan# 2073  
Delta R.T. 0.000 min  
Lab File: VX026922.D  
Acq: 10 Feb 2022 15:59


Tgt Ion:225 Resp: 9856  
Ion Ratio Lower Upper  
225 100  
223 64.7 32.0 96.0  
227 65.9 32.3 96.9





#95  
Naphthalene  
Concen: 19.113 ug/l  
RT: 13.780 min Scan# 2082  
Delta R.T. 0.000 min  
Lab File: VX026922.D  
Acq: 10 Feb 2022 15:59



Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0210WBSD02



Tgt Ion:128 Resp: 84030  
Ion Ratio Lower Upper  
128 100  
127 12.6 10.3 15.5  
129 10.9 8.7 13.1

Manual Integrations  
APPROVED

Reviewed By :Semsettin Yesilyurt 02/11/2022  
Supervised By :Mahesh Dadoda 02/11/2022



#96  
1,2,3-Trichlorobenzene  
Concen: 18.975 ug/l  
RT: 13.963 min Scan# 2112  
Delta R.T. 0.000 min  
Lab File: VX026922.D  
Acq: 10 Feb 2022 15:59

Tgt Ion:180 Resp: 24116  
Ion Ratio Lower Upper  
180 100  
182 94.7 47.2 141.6  
145 35.2 17.5 52.5

