

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021022\
 Data File : VX026948.D
 Acq On : 11 Feb 2022 03:08
 Operator : JC/MD
 Sample : N1467-07
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 42 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-1D

Quant Time: Feb 11 04:09:05 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020822W.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 08 07:07:46 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) Pentafluorobenzene	5.556	168	123282	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	206351	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	191624	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	84932	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	78364	52.953	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	105.900%
35) Dibromofluoromethane	5.391	113	64105	52.144	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	104.280%
50) Toluene-d8	8.653	98	241263	53.462	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	106.920%
62) 4-Bromofluorobenzene	11.079	95	92254	58.065	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	116.140%
Target Compounds						
						Qvalue
4) Vinyl Chloride	1.380	62	18730	14.348	ug/l	99
12) 1,1-Dichloroethene	2.325	96	913	0.724	ug/l	82
16) Acetone	2.380	43	1209	1.860	ug/l	91
21) trans-1,2-Dichloroethene	3.099	96	3462	2.608	ug/l	95
27) cis-1,2-Dichloroethene	4.495	96	679604	445.825	ug/l	88
44) Trichloroethene	7.129	130	141324	83.269	ug/l	97
64) Tetrachloroethene	9.275	164	126602	65.781	ug/l	97
68) m/p-Xylenes	10.299	106	1657	0.614	ug/l	99
69) o-Xylene	10.640	106	1122	0.424	ug/l	96
80) 1,3,5-Trimethylbenzene	11.451	105	4283	0.778	ug/l	92
84) 1,2,4-Trimethylbenzene	11.756	105	14527	2.616	ug/l	98
95) Naphthalene	13.780	128	2603	0.419	ug/l	95


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021022\
Data File : VX026948.D
Acq On : 11 Feb 2022 03:08
Operator : JC/MD
Sample : N1467-07
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 42 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-1D

Quant Time: Feb 11 04:09:05 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020822W.M
Quant Title : SW846 8260
QLast Update : Tue Feb 08 07:07:46 2022
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.556 min Scan# 71
Delta R.T. -0.006 min
Lab File: VX026948.D
Acq: 11 Feb 2022 03:08




Instrument :
MSVOA_X
ClientSampleId :
MW-1D


Tgt Ion:168 Resp: 123282
Ion Ratio Lower Upper
168 100
99 59.6 50.3 75.5



#4
Vinyl Chloride
Concen: 14.348 ug/l
RT: 1.380 min Scan# 48
Delta R.T. 0.000 min
Lab File: VX026948.D
Acq: 11 Feb 2022 03:08

Tgt Ion: 62 Resp: 18730
Ion Ratio Lower Upper
62 100
64 32.3 26.2 39.2





#12

1,1-Dichloroethene

Concen: 0.724 ug/l

RT: 2.325 min Scan# 203

Delta R.T. 0.000 min

Lab File: VX026948.D

Acq: 11 Feb 2022 03:08

Instrument :

MSVOA_X

ClientSampleId :

MW-1D





Tgt Ion: 96 Resp: 913

Ion Ratio Lower Upper

96 100

61 140.9 138.5 207.7

98 61.5 51.8 77.6



#16

Acetone

Concen: 1.860 ug/l

RT: 2.380 min Scan# 212

Delta R.T. -0.006 min

Lab File: VX026948.D




Acq: 11 Feb 2022 03:08


Tgt Ion: 43 Resp: 1209

Ion Ratio Lower Upper

43 100

58 33.5 23.0 34.6





#21

trans-1,2-Dichloroethene

Concen: 2.608 ug/l

RT: 3.099 min Scan# 31

Delta R.T. 0.000 min

Lab File: VX026948.D

Acq: 11 Feb 2022 03:08

Instrument :

MSVOA_X

ClientSampleId :

MW-1D





Tgt Ion: 96 Resp: 3462

Ion Ratio Lower Upper

96 100

61 133.4 112.9 169.3

98 60.2 49.5 74.3



#27

cis-1,2-Dichloroethene

Concen: 445.825 ug/l

RT: 4.495 min Scan# 559

Delta R.T. 0.000 min

Lab File: VX026948.D

Acq: 11 Feb 2022 03:08




Tgt Ion: 96 Resp: 679604


Ion Ratio Lower Upper

96 100

61 126.8 0.0 296.6

98 64.4 0.0 128.8









#33
 1,2-Dichloroethane-d4
 Concen: 52.953 ug/l
 RT: 5.958 min Scan# 799
 Delta R.T. -0.006 min
 Lab File: VX026948.D
 Acq: 11 Feb 2022 03:08




Instrument :
 MSVOA_X
 ClientSampleId :
 MW-1D


Tgt Ion: 65 Resp: 78364
 Ion Ratio Lower Upper
 65 100
 67 52.0 0.0 100.8



#34
 1,4-Difluorobenzene
 Concen: 50.000 ug/l
 RT: 6.763 min Scan# 931
 Delta R.T. -0.006 min
 Lab File: VX026948.D
 Acq: 11 Feb 2022 03:08

Tgt Ion: 114 Resp: 206351
 Ion Ratio Lower Upper
 114 100
 63 20.6 0.0 42.8
 88 16.6 0.0 33.0









#35
Dibromofluoromethane
Concen: 52.144 ug/l
RT: 5.391 min Scan# 706
Delta R.T. 0.000 min
Lab File: VX026948.D
Acq: 11 Feb 2022 03:08




Instrument :
MSVOA_X
ClientSampleId :
MW-1D


Tgt Ion:113 Resp: 64105
Ion Ratio Lower Upper
113 100
111 103.4 81.2 121.8
192 19.9 15.0 22.4



#44
Trichloroethene
Concen: 83.269 ug/l
RT: 7.129 min Scan# 991
Delta R.T. 0.000 min
Lab File: VX026948.D
Acq: 11 Feb 2022 03:08

Tgt Ion:130 Resp: 141324
Ion Ratio Lower Upper
130 100
95 98.4 0.0 202.6









#50
Toluene-d8
Concen: 53.462 ug/l
RT: 8.653 min Scan# 1241
Delta R.T. 0.000 min
Lab File: VX026948.D
Acq: 11 Feb 2022 03:08




Instrument :
MSVOA_X
ClientSampleId :
MW-1D


Tgt Ion: 98 Resp: 241263
Ion Ratio Lower Upper
98 100
100 66.8 53.2 79.8



#62
4-Bromofluorobenzene
Concen: 58.065 ug/l
RT: 11.079 min Scan# 1639
Delta R.T. 0.000 min
Lab File: VX026948.D
Acq: 11 Feb 2022 03:08

Tgt Ion: 95 Resp: 92254
Ion Ratio Lower Upper
95 100
174 75.5 0.0 148.6
176 73.9 0.0 141.4









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.055 min Scan# 1471
Delta R.T. 0.000 min
Lab File: VX026948.D
Acq: 11 Feb 2022 03:08




Instrument :
MSVOA_X
ClientSampleId :
MW-1D


Tgt Ion:117 Resp: 191624
Ion Ratio Lower Upper
117 100
82 55.6 45.2 67.8
119 31.6 26.2 39.2



#64
Tetrachloroethene
Concen: 65.781 ug/l
RT: 9.275 min Scan# 1343
Delta R.T. 0.000 min
Lab File: VX026948.D
Acq: 11 Feb 2022 03:08

Tgt Ion:164 Resp: 126602
Ion Ratio Lower Upper
164 100
166 124.6 101.8 152.6
129 98.9 83.3 124.9
131 94.5 78.0 117.0









#68
m/p-Xylenes
Concen: 0.614 ug/l
RT: 10.299 min Scan# 1511
Delta R.T. -0.006 min
Lab File: VX026948.D
Acq: 11 Feb 2022 03:08




Instrument :
MSVOA_X
ClientSampleId :
MW-1D


Tgt Ion:106 Resp: 1657
Ion Ratio Lower Upper
106 100
91 208.4 165.3 247.9



#69
o-Xylene
Concen: 0.424 ug/l
RT: 10.640 min Scan# 1567
Delta R.T. 0.000 min
Lab File: VX026948.D
Acq: 11 Feb 2022 03:08

Tgt Ion:106 Resp: 1122
Ion Ratio Lower Upper
106 100
91 225.6 109.7 329.0





#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 12.024 min Scan# 1

Delta R.T. 0.000 min

Lab File: VX026948.D

Acq: 11 Feb 2022 03:08

Instrument :

MSVOA_X

ClientSampleId :

MW-1D





Tgt Ion:152 Resp: 84932

Ion Ratio Lower Upper

152 100

115 61.4 44.2 132.6

150 157.7 0.0 348.2



#80

1,3,5-Trimethylbenzene

Concen: 0.778 ug/l

RT: 11.451 min Scan# 1700

Delta R.T. 0.000 min

Lab File: VX026948.D




Acq: 11 Feb 2022 03:08


Tgt Ion:105 Resp: 4283

Ion Ratio Lower Upper

105 100

120 53.2 24.0 72.0









#84
 1,2,4-Trimethylbenzene
 Concen: 2.616 ug/l
 RT: 11.756 min Scan# 11
 Delta R.T. 0.006 min
 Lab File: VX026948.D
 Acq: 11 Feb 2022 03:08

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-1D

Tgt Ion:105 Resp: 14527
 Ion Ratio Lower Upper
 105 100
 120 44.9 21.9 65.5



#95
 Naphthalene
 Concen: 0.419 ug/l
 RT: 13.780 min Scan# 2082
 Delta R.T. 0.000 min
 Lab File: VX026948.D
 Acq: 11 Feb 2022 03:08

Tgt Ion:128 Resp: 2603
 Ion Ratio Lower Upper
 128 100
 127 10.7 10.3 15.5
 129 12.5 8.7 13.1

