



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021022\
Data File : VX026938.D
Acq On : 10 Feb 2022 22:09
Operator : JC/MD
Sample : BFB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 29 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020822W.M
Title : SW846 8260
Last Update : Tue Feb 08 07:07:46 2022



AutoFind: Scans 1638, 1639, 1640; Background Corrected with Scan 1632

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	18.4	4709	PASS
75	95	30	60	50.1	12809	PASS
95	95	100	100	100.0	25555	PASS
96	95	5	9	6.8	1726	PASS
173	174	0.00	2	1.2	205	PASS
174	95	50	100	69.4	17729	PASS
175	174	5	9	8.1	1440	PASS
176	174	95	101	100.7	17848	PASS
177	176	5	9	6.8	1221	PASS