

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021023\
 Data File : VX034078.D
 Acq On : 08 Feb 2023 12:23
 Operator : JC/MD
 Sample : PB150737TB
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB150737TB

Quant Time: Feb 08 13:20:54 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020423W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 03 23:25:45 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	130849	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	184798	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	164715	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	80803	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	63670	50.474	ug/l	0.00
Spiked Amount	50.000	Range 78 - 117	Recovery	= 100.940%		
35) Dibromofluoromethane	5.385	113	58894	50.075	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 100.160%		
50) Toluene-d8	8.647	98	211115	49.821	ug/l	0.00
Spiked Amount	50.000	Range 92 - 112	Recovery	= 99.640%		
62) 4-Bromofluorobenzene	11.079	95	74907	49.495	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	= 99.000%		
Target Compounds						
					Qvalue	
16) Acetone	2.374	43	23742	44.562	ug/l	94
20) Methylene Chloride	2.794	84	4798	3.936	ug/l #	81
25) 2-Butanone	4.562	43	3146	4.055	ug/l	94
37) Ethyl Acetate	4.709	43	15767	10.496	ug/l	98
43) Isopropyl Acetate	6.336	43	68388	29.192	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021023\
Data File : VX034078.D
Acq On : 08 Feb 2023 12:23
Operator : JC/MD
Sample : PB150737TB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB150737TB

Quant Time: Feb 08 13:20:54 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020423W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 03 23:25:45 2023
Response via : Initial Calibration

