

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021023\
 Data File : VX034084.D
 Acq On : 08 Feb 2023 14:43
 Operator : JC/MD
 Sample : 01463-16
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TCLP-6

Quant Time: Feb 09 04:51:14 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020423W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 03 23:25:45 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	126542	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	176410	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	156974	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	80456	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	60421	49.528	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	99.060%
35) Dibromofluoromethane	5.385	113	55600	49.522	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	99.040%
50) Toluene-d8	8.647	98	199940	49.427	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	98.860%
62) 4-Bromofluorobenzene	11.079	95	72256	50.014	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	100.020%
Target Compounds						
					Qvalue	
16) Acetone	2.374	43	22696	44.049	ug/l #	86
20) Methylene Chloride	2.794	84	4287	3.637	ug/l #	92
37) Ethyl Acetate	4.708	43	14418	10.055	ug/l	96
43) Isopropyl Acetate	6.336	43	63875	28.562	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021023\
Data File : VX034084.D
Acq On : 08 Feb 2023 14:43
Operator : JC/MD
Sample : 01463-16
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TCPLP-6

Quant Time: Feb 09 04:51:14 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020423W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 03 23:25:45 2023
Response via : Initial Calibration

