

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021025\
 Data File : VX044870.D
 Acq On : 10 Feb 2025 11:11
 Operator : JC/MD
 Sample : VSTDIC020
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDIC020

Manual Integrations
 APPROVED

Reviewed By : John Carlone 02/11/2025
 Supervised By : Mahesh Dadoda 02/11/2025

Quant Time: Feb 11 02:00:10 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X021025W.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 11 01:47:50 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	129532	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	234445	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	207667	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	92093	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	37195	19.616	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	39.240%#		
35) Dibromofluoromethane	5.379	113	30170	19.793	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	39.580%#		
50) Toluene-d8	8.647	98	117127	20.321	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	40.640%#		
62) 4-Bromofluorobenzene	11.079	95	38475	19.800	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery =	39.600%#		
Target Compounds						
					Qvalue	
2) Dichlorodifluoromethane	1.167	85	37123	20.434	ug/l	98
3) Chloromethane	1.301	50	44666	20.200	ug/l	95
4) Vinyl Chloride	1.374	62	43901	20.494	ug/l	99
5) Bromomethane	1.593	94	13057	20.369	ug/l	99
6) Chloroethane	1.660	64	14523	16.522	ug/l	94
7) Trichlorofluoromethane	1.868	101	56805	20.969	ug/l	98
8) Diethyl Ether	2.130	74	20699	21.069	ug/l	99
9) 1,1,2-Trichlorotrifluo...	2.319	101	34593	21.135	ug/l	99
10) Methyl Iodide	2.441	142	44749	21.096	ug/l	99
11) Tert butyl alcohol	2.989	59	35834	101.870	ug/l	100
12) 1,1-Dichloroethene	2.307	96	34270	20.526	ug/l	99
13) Acrolein	2.233	56	33914	93.445	ug/l	99
14) Allyl chloride	2.654	41	61485	20.377	ug/l	98
15) Acrylonitrile	3.063	53	98518	104.896	ug/l	98
16) Acetone	2.380	43	77078	101.119	ug/l	98
17) Carbon Disulfide	2.502	76	92552	20.216	ug/l	99
18) Methyl Acetate	2.703	43	47970	19.943	ug/l	99
19) Methyl tert-butyl Ether	3.111	73	110377	20.779	ug/l	98
20) Methylene Chloride	2.782	84	38390	20.562	ug/l	99
21) trans-1,2-Dichloroethene	3.081	96	34064	20.730	ug/l	92
22) Diisopropyl ether	3.757	45	120300	21.074	ug/l	93
23) Vinyl Acetate	3.715	43	490239	105.341	ug/l	99
24) 1,1-Dichloroethane	3.599	63	66942	20.958	ug/l	99
25) 2-Butanone	4.556	43	130481	105.417	ug/l	96
26) 2,2-Dichloropropane	4.471	77	51803	20.451	ug/l	100
27) cis-1,2-Dichloroethene	4.483	96	42096	21.328	ug/l	97
28) Bromochloromethane	4.885	49	31474	20.504	ug/l	99
29) Tetrahydrofuran	5.001	42	86875	105.700	ug/l	99
30) Chloroform	5.087	83	65721	21.217	ug/l	98
31) Cyclohexane	5.458	56	60814	20.653	ug/l	97
32) 1,1,1-Trichloroethane	5.373	97	54450	20.724	ug/l	100
36) 1,1-Dichloropropene	5.684	75	45991	20.913	ug/l	99
37) Ethyl Acetate	4.715	43	51242	20.949	ug/l	99
38) Carbon Tetrachloride	5.666	117	44802	20.790	ug/l	99
39) Methylcyclohexane	7.373	83	62505	21.985	ug/l	98
40) Benzene	6.031	78	147912	21.540	ug/l	100

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021025\
 Data File : VX044870.D
 Acq On : 10 Feb 2025 11:11
 Operator : JC/MD
 Sample : VSTDICC020
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDICC020

Manual Integrations
 APPROVED

Reviewed By : John Carlone 02/11/2025
 Supervised By : Mahesh Dadoda 02/11/2025

Quant Time: Feb 11 02:00:10 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X021025W.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 11 01:47:50 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	4.916	41	27347	20.887	ug/l	94
42) 1,2-Dichloroethane	6.086	62	47095	21.519	ug/l	100
43) Isopropyl Acetate	6.336	43	84306	21.343	ug/l	99
44) Trichloroethene	7.123	130	34390	21.910	ug/l	96
45) 1,2-Dichloropropane	7.428	63	36510	21.371	ug/l	99
46) Dibromomethane	7.574	93	25300	21.354	ug/l	99
47) Bromodichloromethane	7.818	83	48202	21.012	ug/l	97
48) Methyl methacrylate	7.690	41	39747	21.341	ug/l	100
49) 1,4-Dioxane	7.665	88	17153	429.262	ug/l	99
51) 4-Methyl-2-Pentanone	8.574	43	263394	109.684	ug/l	99
52) Toluene	8.714	92	89726	21.940	ug/l	99
53) t-1,3-Dichloropropene	8.976	75	48613	20.940	ug/l	98
54) cis-1,3-Dichloropropene	8.366	75	55026	21.259	ug/l	97
55) 1,1,2-Trichloroethane	9.147	97	33930	21.582	ug/l	98
56) Ethyl methacrylate	9.116	69	54674	21.533	ug/l	99
57) 1,3-Dichloropropane	9.305	76	60079	21.841	ug/l	99
58) 2-Chloroethyl Vinyl ether	8.238	63	133072	106.736	ug/l	100
59) 2-Hexanone	9.427	43	190449	110.072	ug/l	98
60) Dibromochloromethane	9.519	129	35706	21.237	ug/l	99
61) 1,2-Dibromoethane	9.604	107	34463	21.628	ug/l	100
64) Tetrachloroethene	9.269	164	28519	21.776	ug/l	97
65) Chlorobenzene	10.073	112	94695	21.230	ug/l	99
66) 1,1,1,2-Tetrachloroethane	10.159	131	30344	21.136	ug/l	99
67) Ethyl Benzene	10.189	91	167865	21.327	ug/l	100
68) m/p-Xylenes	10.299	106	125271	43.144	ug/l	99
69) o-Xylene	10.640	106	62087	21.310	ug/l	97
70) Styrene	10.653	104	103740	22.101	ug/l	98
71) Bromoform	10.799	173	22592	21.086	ug/l #	99
73) Isopropylbenzene	10.957	105	160147	21.597	ug/l	99
74) N-amyl acetate	10.842	43	70542	21.023	ug/l	98
75) 1,1,2,2-Tetrachloroethane	11.213	83	52967	20.787	ug/l	99
76) 1,2,3-Trichloropropane	11.238	75	42040m	20.589	ug/l	
77) Bromobenzene	11.195	156	36021	21.426	ug/l	97
78) n-propylbenzene	11.299	91	183212	21.406	ug/l	99
79) 2-Chlorotoluene	11.360	91	111782	21.250	ug/l	100
80) 1,3,5-Trimethylbenzene	11.451	105	130526	21.691	ug/l	100
81) trans-1,4-Dichloro-2-b...	11.018	75	15375	19.370	ug/l	96
82) 4-Chlorotoluene	11.451	91	122303	21.016	ug/l	100
83) tert-Butylbenzene	11.713	119	130358	21.414	ug/l	100
84) 1,2,4-Trimethylbenzene	11.750	105	130211	21.924	ug/l	100
85) sec-Butylbenzene	11.890	105	162875	21.607	ug/l	100
86) p-Isopropyltoluene	12.006	119	130643	21.551	ug/l	99
87) 1,3-Dichlorobenzene	11.969	146	64141	20.747	ug/l	100
88) 1,4-Dichlorobenzene	12.037	146	64889	20.758	ug/l	98
89) n-Butylbenzene	12.329	91	113705	21.202	ug/l	100
90) Hexachloroethane	12.536	117	23153	20.398	ug/l	97
91) 1,2-Dichlorobenzene	12.335	146	64940	21.373	ug/l	99
92) 1,2-Dibromo-3-Chloropr...	12.939	75	9873	21.230	ug/l	97
93) 1,2,4-Trichlorobenzene	13.585	180	37308	20.292	ug/l	99
94) Hexachlorobutadiene	13.725	225	15312	20.979	ug/l	96
95) Naphthalene	13.774	128	142902	21.502	ug/l	99
96) 1,2,3-Trichlorobenzene	13.957	180	38389	20.722	ug/l	98

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021025\
Data File : VX044870.D
Acq On : 10 Feb 2025 11:11
Operator : JC/MD
Sample : VSTDICC020
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC020

Manual Integrations
APPROVED

Reviewed By :John Carlone 02/11/2025
Supervised By :Mahesh Dadoda 02/11/2025

Quant Time: Feb 11 02:00:10 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X021025W.M
Quant Title : SW846 8260
QLast Update : Tue Feb 11 01:47:50 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021025\
 Data File : VX044870.D
 Acq On : 10 Feb 2025 11:11
 Operator : JC/MD
 Sample : VSTDIC020
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDIC020

Quant Time: Feb 11 02:00:10 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X021025W.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 11 01:47:50 2025
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By : John Carlone 02/11/2025
 Supervised By : Mahesh Dadoda 02/11/2025

