

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021025\
 Data File : VX044881.D
 Acq On : 10 Feb 2025 15:41
 Operator : JC/MD
 Sample : Q1250-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-VPB-192-TB-20250127

Quant Time: Feb 11 06:06:36 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X021025W.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 11 03:41:08 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	92731	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	189692	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	170630	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	69234	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	71967	53.018	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	106.040%
35) Dibromofluoromethane	5.379	113	61025	49.481	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.960%
50) Toluene-d8	8.647	98	234221	50.223	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	100.440%
62) 4-Bromofluorobenzene	11.079	95	79575	50.613	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	101.220%
Target Compounds						
16) Acetone	2.386	43	1601	2.934	ug/l	Qvalue 94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021025\
Data File : VX044881.D
Acq On : 10 Feb 2025 15:41
Operator : JC/MD
Sample : Q1250-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-VPB-192-TB-20250127

Quant Time: Feb 11 06:06:36 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X021025W.M
Quant Title : SW846 8260
QLast Update : Tue Feb 11 03:41:08 2025
Response via : Initial Calibration

