

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021025\
 Data File : VX044883.D
 Acq On : 10 Feb 2025 16:27
 Operator : JC/MD
 Sample : Q1250-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-VPB-192-GW-320-322

Quant Time: Feb 11 06:07:22 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X021025W.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 11 03:41:08 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	96132	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	194678	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	175107	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	76744	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.946	65	73570	52.281	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 104.560%	
35) Dibromofluoromethane	5.379	113	62879	49.678	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 99.360%	
50) Toluene-d8	8.647	98	241572	50.472	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 100.940%	
62) 4-Bromofluorobenzene	11.079	95	83784	51.925	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	= 103.860%	
Target Compounds						Qvalue
16) Acetone	2.385	43	4882	8.630	ug/l	94
17) Carbon Disulfide	2.501	76	1717	0.505	ug/l #	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021025\
Data File : VX044883.D
Acq On : 10 Feb 2025 16:27
Operator : JC/MD
Sample : Q1250-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-VPB-192-GW-320-322

Quant Time: Feb 11 06:07:22 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X021025W.M
Quant Title : SW846 8260
QLast Update : Tue Feb 11 03:41:08 2025
Response via : Initial Calibration

