

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021025\
 Data File : VX044886.D
 Acq On : 10 Feb 2025 17:37
 Operator : JC/MD
 Sample : Q1250-11
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-VPB-192-GW-400-402

Quant Time: Feb 11 06:08:34 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X021025W.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 11 03:41:08 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.549	168	94362	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.756	114	191816	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	178953	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	76716	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	74380	53.848	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 107.700%	
35) Dibromofluoromethane	5.379	113	63934	51.265	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 102.540%	
50) Toluene-d8	8.646	98	240408	50.978	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 101.960%	
62) 4-Bromofluorobenzene	11.079	95	85054	53.499	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	= 107.000%	
Target Compounds						Qvalue
16) Acetone	2.379	43	7097	12.781	ug/l	96
17) Carbon Disulfide	2.507	76	1145	0.343	ug/l #	76

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021025\
Data File : VX044886.D
Acq On : 10 Feb 2025 17:37
Operator : JC/MD
Sample : Q1250-11
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-VPB-192-GW-400-402

Quant Time: Feb 11 06:08:34 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X021025W.M
Quant Title : SW846 8260
QLast Update : Tue Feb 11 03:41:08 2025
Response via : Initial Calibration

