

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021025\
 Data File : VX044888.D
 Acq On : 10 Feb 2025 18:23
 Operator : JC/MD
 Sample : Q1250-06
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-VPB-192-GW-360-362

Quant Time: Feb 11 06:09:25 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X021025W.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 11 03:41:08 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.543	168	93911	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	187948	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	173877	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	73228	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.946	65	72312	52.602	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 105.200%			
35) Dibromofluoromethane	5.379	113	62542	51.181	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 102.360%			
50) Toluene-d8	8.646	98	235663	51.001	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 102.000%			
62) 4-Bromofluorobenzene	11.079	95	82802	53.154	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery = 106.300%			
Target Compounds						
16) Acetone	2.379	43	2643	4.783	ug/l	97
17) Carbon Disulfide	2.501	76	953	0.287	ug/l #	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021025\
Data File : VX044888.D
Acq On : 10 Feb 2025 18:23
Operator : JC/MD
Sample : Q1250-06
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-VPB-192-GW-360-362

Quant Time: Feb 11 06:09:25 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X021025W.M
Quant Title : SW846 8260
QLast Update : Tue Feb 11 03:41:08 2025
Response via : Initial Calibration

