

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021122\
 Data File : VX026960.D
 Acq On : 11 Feb 2022 12:31
 Operator : JC/MD
 Sample : VX0211MBL01
 Misc : 5.0g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0211MBL01

Quant Time: Feb 12 04:26:42 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020822W.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 08 07:07:46 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	119880	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	198835	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	181118	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	77008	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	75647	52.568	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 105.140%	
35) Dibromofluoromethane	5.385	113	62038	52.370	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 104.740%	
50) Toluene-d8	8.647	98	232899	53.560	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 107.120%	
62) 4-Bromofluorobenzene	11.079	95	85105	55.590	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 111.180%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021122\
Data File : VX026960.D
Acq On : 11 Feb 2022 12:31
Operator : JC/MD
Sample : VX0211MBL01
Misc : 5.0g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0211MBL01

Quant Time: Feb 12 04:26:42 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020822W.M
Quant Title : SW846 8260
QLast Update : Tue Feb 08 07:07:46 2022
Response via : Initial Calibration

