

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021122\
 Data File : VX026961.D
 Acq On : 11 Feb 2022 12:54
 Operator : JC/MD
 Sample : VX0211WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0211WBL01

Quant Time: Feb 12 04:27:02 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020822W.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 08 07:07:46 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	120975	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	202826	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	182354	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	76188	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	77086	53.083	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 106.160%	
35) Dibromofluoromethane	5.391	113	64083	53.032	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 106.060%	
50) Toluene-d8	8.653	98	236003	53.206	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 106.420%	
62) 4-Bromofluorobenzene	11.079	95	84731	54.257	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 108.520%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021122\
Data File : VX026961.D
Acq On : 11 Feb 2022 12:54
Operator : JC/MD
Sample : VX0211WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0211WBL01

Quant Time: Feb 12 04:27:02 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020822W.M
Quant Title : SW846 8260
QLast Update : Tue Feb 08 07:07:46 2022
Response via : Initial Calibration

