

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021122\
 Data File : VX026964.D
 Acq On : 11 Feb 2022 14:09
 Operator : JC/MD
 Sample : N1481-01DL 10X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RW7B-GW-20220209DL

Quant Time: Feb 12 04:28:43 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020822W.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 08 07:07:46 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	120284	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	201695	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	173769	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	71986	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	76345	52.875	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 105.740%	
35) Dibromofluoromethane	5.391	113	62937	52.375	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 104.760%	
50) Toluene-d8	8.653	98	229913	52.123	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 104.240%	
62) 4-Bromofluorobenzene	11.079	95	79577	51.243	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 102.480%	
Target Compounds						
						Qvalue
9) 1,1,2-Trichlorotrifluo...	2.337	101	3371	2.744	ug/l	93
12) 1,1-Dichloroethene	2.325	96	2409	1.957	ug/l	97
27) cis-1,2-Dichloroethene	4.501	96	416	0.280	ug/l	85
44) Trichloroethene	7.129	130	58117	35.033	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021122\
Data File : VX026964.D
Acq On : 11 Feb 2022 14:09
Operator : JC/MD
Sample : N1481-01DL 10X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RW7B-GW-20220209DL

Quant Time: Feb 12 04:28:43 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020822W.M
Quant Title : SW846 8260
QLast Update : Tue Feb 08 07:07:46 2022
Response via : Initial Calibration

