

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021122\
 Data File : VX026966.D
 Acq On : 11 Feb 2022 14:56
 Operator : JC/MD
 Sample : N1497-01DL 20X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-7SDL

Quant Time: Feb 12 04:29:29 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020822W.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 08 07:07:46 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	116934	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	196264	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	167043	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	70578	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	75917	54.084	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	108.160%
35) Dibromofluoromethane	5.391	113	61399	52.509	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	105.020%
50) Toluene-d8	8.653	98	223770	52.134	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	104.260%
62) 4-Bromofluorobenzene	11.079	95	76894	50.885	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	101.780%
Target Compounds						
					Qvalue	
4) Vinyl Chloride	1.374	62	4048	3.269	ug/l	96
12) 1,1-Dichloroethene	2.331	96	341	0.285	ug/l #	77
27) cis-1,2-Dichloroethene	4.495	96	57545	39.799	ug/l	88
44) Trichloroethene	7.129	130	13962	8.649	ug/l	92
64) Tetrachloroethene	9.275	164	11587	6.906	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021122\
Data File : VX026966.D
Acq On : 11 Feb 2022 14:56
Operator : JC/MD
Sample : N1497-01DL 20X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-7SDL

Quant Time: Feb 12 04:29:29 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020822W.M
Quant Title : SW846 8260
QLast Update : Tue Feb 08 07:07:46 2022
Response via : Initial Calibration

