

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021122\
 Data File : VX026970.D
 Acq On : 11 Feb 2022 16:28
 Operator : JC/MD
 Sample : N1467-11DL 20X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-23DDL

Quant Time: Feb 12 04:31:04 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020822W.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 08 07:07:46 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	113411	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	191781	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	163203	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	69023	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	73604	54.066	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 108.140%	
35) Dibromofluoromethane	5.391	113	60678	53.106	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 106.220%	
50) Toluene-d8	8.653	98	218133	52.009	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 104.020%	
62) 4-Bromofluorobenzene	11.079	95	75444	51.093	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 102.180%	
Target Compounds						
					Qvalue	
4) Vinyl Chloride	1.380	62	458	0.381	ug/l	98
21) trans-1,2-Dichloroethene	3.099	96	571	0.468	ug/l	98
27) cis-1,2-Dichloroethene	4.495	96	72483	51.688	ug/l	88
44) Trichloroethene	7.129	130	40081	25.410	ug/l	96
64) Tetrachloroethene	9.275	164	40421	24.660	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021122\
Data File : VX026970.D
Acq On : 11 Feb 2022 16:28
Operator : JC/MD
Sample : N1467-11DL 20X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-23DDL

Quant Time: Feb 12 04:31:04 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020822W.M
Quant Title : SW846 8260
QLast Update : Tue Feb 08 07:07:46 2022
Response via : Initial Calibration

