

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021122\
 Data File : VX026971.D
 Acq On : 11 Feb 2022 16:51
 Operator : JC/MD
 Sample : N1467-15DL 20X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-1DL

Quant Time: Feb 12 04:31:24 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020822W.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 08 07:07:46 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	103850	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	175630	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	153496	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	65866	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	68543	54.983	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 109.960%	
35) Dibromofluoromethane	5.391	113	56040	53.557	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 107.120%	
50) Toluene-d8	8.653	98	201201	52.383	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 104.760%	
62) 4-Bromofluorobenzene	11.079	95	71385	52.789	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 105.580%	
Target Compounds						
					Qvalue	
4) Vinyl Chloride	1.374	62	444	0.404	ug/l	90
27) cis-1,2-Dichloroethene	4.495	96	63243	49.251	ug/l	88
44) Trichloroethene	7.129	130	15387	10.652	ug/l	84
64) Tetrachloroethene	9.275	164	31926	20.709	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021122\
Data File : VX026971.D
Acq On : 11 Feb 2022 16:51
Operator : JC/MD
Sample : N1467-15DL 20X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-1DL

Quant Time: Feb 12 04:31:24 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020822W.M
Quant Title : SW846 8260
QLast Update : Tue Feb 08 07:07:46 2022
Response via : Initial Calibration

