

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021122\
 Data File : VX026974.D
 Acq On : 11 Feb 2022 18:01
 Operator : JC/MD
 Sample : N1467-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-30S

Quant Time: Feb 12 06:50:42 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020822W.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 08 07:07:46 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	106363	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	178174	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	166580	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	70993	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	68869	53.940	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 107.880%	
35) Dibromofluoromethane	5.391	113	56352	53.086	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 106.180%	
50) Toluene-d8	8.653	98	210561	54.038	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 108.080%	
62) 4-Bromofluorobenzene	11.079	95	79014	57.597	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 115.200%	
Target Compounds						
					Qvalue	
4) Vinyl Chloride	1.374	62	1874	1.664	ug/l	97
21) trans-1,2-Dichloroethene	3.093	96	462	0.403	ug/l	91
27) cis-1,2-Dichloroethene	4.495	96	104787	79.676	ug/l	88
29) Tetrahydrofuran	5.013	42	6883	11.206	ug/l	98
44) Trichloroethene	7.129	130	47344	32.307	ug/l	96
64) Tetrachloroethene	9.275	164	19587	11.707	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021122\
Data File : VX026974.D
Acq On : 11 Feb 2022 18:01
Operator : JC/MD
Sample : N1467-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-30S

Quant Time: Feb 12 06:50:42 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020822W.M
Quant Title : SW846 8260
QLast Update : Tue Feb 08 07:07:46 2022
Response via : Initial Calibration

