

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021122\
 Data File : VX026976.D
 Acq On : 11 Feb 2022 18:47
 Operator : JC/MD
 Sample : N1519-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 STORAGE-BLANK-SOIL-REF-6

Quant Time: Feb 12 04:33:25 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020822W.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 08 07:07:46 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	105968	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	178648	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	161499	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	67113	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	70006	55.034	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 110.060%	
35) Dibromofluoromethane	5.391	113	56449	53.036	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 106.080%	
50) Toluene-d8	8.653	98	210080	53.771	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 107.540%	
62) 4-Bromofluorobenzene	11.079	95	75689	55.027	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 110.060%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021122\
Data File : VX026976.D
Acq On : 11 Feb 2022 18:47
Operator : JC/MD
Sample : N1519-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-SOIL-REF-6

Quant Time: Feb 12 04:33:25 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020822W.M
Quant Title : SW846 8260
QLast Update : Tue Feb 08 07:07:46 2022
Response via : Initial Calibration

