

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021122\
 Data File : VX026978.D
 Acq On : 11 Feb 2022 19:34
 Operator : JC/MD
 Sample : N1519-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 STORAGE-BLANK-WATER-REF-4

Quant Time: Feb 12 04:34:08 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020822W.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 08 07:07:46 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	107945	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	180668	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	164136	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	68216	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	70347	54.290	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 108.580%	
35) Dibromofluoromethane	5.391	113	57176	53.119	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 106.240%	
50) Toluene-d8	8.653	98	213911	54.140	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 108.280%	
62) 4-Bromofluorobenzene	11.079	95	75782	54.478	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 108.960%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021122\
Data File : VX026978.D
Acq On : 11 Feb 2022 19:34
Operator : JC/MD
Sample : N1519-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-WATER-REF-4

Quant Time: Feb 12 04:34:08 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020822W.M
Quant Title : SW846 8260
QLast Update : Tue Feb 08 07:07:46 2022
Response via : Initial Calibration

