

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021122\
 Data File : VX026991.D
 Acq On : 12 Feb 2022 00:57
 Operator : JC/MD
 Sample : PB142590TB
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 35 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB142590TB

Quant Time: Feb 12 04:41:39 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020822W.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 08 07:07:46 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	114828	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	193938	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	179618	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	78905	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	74176	53.813	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 107.620%	
35) Dibromofluoromethane	5.391	113	60737	52.566	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 105.140%	
50) Toluene-d8	8.653	98	228901	53.969	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 107.940%	
62) 4-Bromofluorobenzene	11.079	95	85511	57.266	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 114.540%	
Target Compounds						
						Qvalue
16) Acetone	2.380	43	4425	7.309	ug/l	93
20) Methylene Chloride	2.794	84	1468	1.119	ug/l	96
43) Isopropyl Acetate	6.342	43	9668	3.033	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021122\
Data File : VX026991.D
Acq On : 12 Feb 2022 00:57
Operator : JC/MD
Sample : PB142590TB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 35 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB142590TB

Quant Time: Feb 12 04:41:39 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020822W.M
Quant Title : SW846 8260
QLast Update : Tue Feb 08 07:07:46 2022
Response via : Initial Calibration

