

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021122\
 Data File : VX026992.D
 Acq On : 12 Feb 2022 01:20
 Operator : JC/MD
 Sample : N1517-05 50X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 36 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 20994

Quant Time: Feb 12 04:42:01 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020822W.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 08 07:07:46 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	116997	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	195112	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	176143	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	73793	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	74833	53.283	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 106.560%	
35) Dibromofluoromethane	5.391	113	61245	52.687	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 105.380%	
50) Toluene-d8	8.653	98	229279	53.733	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 107.460%	
62) 4-Bromofluorobenzene	11.079	95	81436	54.209	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 108.420%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021122\
Data File : VX026992.D
Acq On : 12 Feb 2022 01:20
Operator : JC/MD
Sample : N1517-05 50X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 36 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
20994

Quant Time: Feb 12 04:42:01 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020822W.M
Quant Title : SW846 8260
QLast Update : Tue Feb 08 07:07:46 2022
Response via : Initial Calibration

