

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021122\
 Data File : VX027000.D
 Acq On : 12 Feb 2022 04:25
 Operator : JC/MD
 Sample : N1465-05 20X
 Misc : 6.91g/5mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 44 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 A5-5-7.5

Quant Time: Feb 14 00:28:59 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020822W.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 08 07:07:46 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	115514	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	193273	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	172935	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	72390	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	74329	53.604	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 107.200%	
35) Dibromofluoromethane	5.391	113	59912	52.030	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 104.060%	
50) Toluene-d8	8.653	98	226564	53.602	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 107.200%	
62) 4-Bromofluorobenzene	11.079	95	79843	53.654	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 107.300%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021122\
Data File : VX027000.D
Acq On : 12 Feb 2022 04:25
Operator : JC/MD
Sample : N1465-05 20X
Misc : 6.91g/5mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 44 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
A5-5-7.5

Quant Time: Feb 14 00:28:59 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020822W.M
Quant Title : SW846 8260
QLast Update : Tue Feb 08 07:07:46 2022
Response via : Initial Calibration

